



**Bihar Medical Services & Infrastructure Corporation
Limited**

**2nd & 3rd Floor, Swasthya Bhawan, Behind IGIMS,
Sheikhpura, Adjacent to State Health Society, Bihar,
Patna- 800014**

Phone/Fax: +919006405255, + 919471009193

CORRIGENDUM-II

**Tender for rate contract and supply of Medical Devices/Consumables for different
healthcare facilities of state of Bihar**

Notice Inviting Tender Ref No.: -BMSIC/MEDICAL DEVICES/CONSUMABLES/24-02

Dated: - 19-06-2024

(Only through E- Tender on website: -www.eproc2.bihar.gov.in)

Bihar Medical Services and Infrastructure Corporation Limited (BMSICL) invites E-Bids from the interested parties for “**Tender for rate contract and supply of Medical Devices/Consumables for different healthcare facilities of state of Bihar**”, vide Notice Inviting Tender No.- **BMSIC/MEDICAL DEVICES/CONSUMABLES/24-02**. Detailed tender document containing eligibility criteria, selection mechanism, other terms and conditions are available on the website www.eproc2.bihar.gov.in.

After considering the suggestion/queries received from the prospective bidders & after due deliberation on all aspects, certain amendments has been made in packing unit in Annexure I of this Corrigendum-II. Rest of the terms and condition along with tender schedule shall remain unchanged.

**Sd/-
GM (Procurement)
BMSICL**

ANNEXURE-I

| S.N. | Name of item | Specification / Strength | Packing unit | Estimated Quantity |
|------|--------------------------------|---|--------------|--------------------|
| 1. | Surgical Gloves Sterile 6 | Disposable Sterile Surgical Rubber Gloves Size 6 Inches Made of natural rubber Latex, powdered, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non-allergenic should Conform to IS 13422 ISI marked / CE certified / FDA | Each Pair | 15,00,000 |
| 2. | Surgical Gloves Sterile 6.5 | (Disposable Sterile Surgical Rubber Gloves Size 6 ½ Inches Made of natural rubber Latex, powdered, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non-allergenic should Conform to IS 13422 ISI marked / CE certified / FDA approved) | Each Pair | 1,50,00,000 |
| 3. | Surgical Gloves Sterile 7 | (Disposable Sterile Surgical Rubber Gloves Size 7 Inches Made of natural rubber Latex, powdered, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non-allergenic should Conform to IS 13422 ISI marked / CE certified / FDA approved) | Each Pair | 1,50,00,000 |
| 4. | Surgical Gloves Sterile 7.5 | (Disposable Sterile Surgical Rubber Gloves Size 7 ½ Inches Made of natural rubber Latex, powdered, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non-allergenic should Conform to IS 13422 ISI marked / CE certified / FDA approved) | Each Pair | 30,00,000 |
| 5. | Surgical Gloves Sterile 8 | (Disposable Sterile Surgical Rubber Gloves Size 8 Inches Made of natural rubber Latex, powdered, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non-allergenic should Conform to IS 13422 ISI marked / CE certified / FDA approved) | Each Pair | 3,00,000 |
| 6. | Surgical Gloves Non- Sterile 6 | Disposable Non-Sterile Surgical Rubber Gloves Size 6 Inches Made of natural rubber Latex, powdered free, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 should Conform to IS 4148 : 1989 ISI marked / CE certified / FDA approved | Each Pair | 12,00,000 |

| | | | | |
|-----|------------------------------------|--|------------------|-----------|
| 7. | Surgical Gloves Non-Sterile 6.5 | Disposable Non-Sterile Surgical Rubber Gloves Size 6 ½ Inches Made of natural rubber Latex, powdered free, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 should Conform to IS 4148 : 1989 ISI marked / CE certified / FDA approved | Each Pair | 40,00,000 |
| 8. | Surgical Gloves Non-Sterile 7 | Disposable Non-Sterile Surgical Rubber Gloves Size 7 Inches Made of natural rubber Latex, powdered free, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 should Conform to IS 4148 : 1989 ISI marked / CE certified / FDA approved | Each Pair | 30,00,000 |
| 9. | Surgical Gloves Non-Sterile 7.5 | Disposable Non-Sterile Surgical Rubber Gloves Size 7 ½ Inches Made of natural rubber Latex, powdered free, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 should Conform to IS 4148 : 1989 ISI marked / CE certified / FDA approved | Each Pair | 30,00,000 |
| 10. | Surgical Gloves Non-Sterile 8 | Disposable Non-Sterile Surgical Rubber Gloves Size 8 Inches Made of natural rubber Latex, powdered free, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 should Conform to IS 4148 : 1989 ISI marked / CE certified / FDA approved | Each Pair | 2,50,000 |

Note- I :-

Please refer to the Annexure-I of this corrigendum for Revised price bid before submission of the bid. Those bidders who have already submitted their bids shall also have to Re-Submit their bids accordingly.

Note- II :-

Tendered quantity is just a tentative quantity and shouldn't be treated as final. It may increase or decrease, as per need of the State. Keeping this in mind, the supplier isn't allowed to seek termination of the Rate Contract after finishing supply of tendered quantity.

**Sd/-
GM (Procurement)
BMSICL**