



**Bihar Medical Services & Infrastructure Corporation
Limited 2nd & 3rd Floor, Swasthya Bhawan, Behind
IGIMS, Sheikhpura, Adjacent to State Health Society,
Patna- 800014, Bihar, Phone/Fax: +91612 2283287, 91612
2283288**

Corrigendum-II

Bihar Medical Services and Infrastructure Corporation Limited (BMSICL) had invited E-Bids from the interested parties for Procurement, Rate contract, Supply, Installation of Medical Equipment vide Tender No.-BMSICL/2024-25/ME-401. During Pre-bid meeting various suggestions were received from different prospective bidders regarding amendment in technical specification of equipment which were discussed and deliberated on by the experts. On the basis of their recommendations certain amendments in the technical specification of the equipment have been made which are annexed as **Annexure-I** of this corrigendum.

SD/-

GM (Procurement)

Annexure-I

Name of Equipment :- Oxygen Concentrator (5 Litre)		
Sl. No	Technical Specification as per tender	Technical Specification after Amendment
1	Oxygen concentrator to provide oxygen from ambient air	No Change
2	Oxygen concentration measured at the flow meter by oxygen sensing device (OSD)	No Change
3	Sound level <50 dB	No Change
4	Superior grade of molecular sieve	No Change
5	Maintenance free valve.	No Change
6	Oxygen purity, approx.: 90%. (+/-3)	No Change
7	Oxygen output, approx.: 0 - 5 LPM	No Change
8	Pressure, approx.: 8 psi	No Change
9	Double/ Single outlet for oxygen Delivery	No Change
10	Oxygen tube of 2 m length must be provided with	No Change
11	Facility for nebulization with tube & mask	No Change
12	With one humidifier bottles and one cabinet filters	No Change
13	Unit should function with 200-240Vac, 50/60 Hz input power supply	No Change
14	Should have USFDA/European CE (Issued by notified body)/BIS	No Change

Name of Equipment :- Oxygen Concentrator (10 Litre)		
Sl. No	Technical Specification as per tender	Technical Specification after Amendment
1	Oxygen concentrator to provide oxygen from ambient air	No Change
2	Oxygen concentration measured at the flow meter by oxygen sensing device (OSD)	No Change
3	Sound level < 50 dB	No Change
4	Superior grade of molecular sieve	No Change
5	Maintenance free valve.	No Change
6	Oxygen purity, approx.: 95% (+/-3)	No Change
7	Oxygen output, approx.: 0 - 10 LPM	No Change

8	Pressure, approx.: 8 psi	No Change
9	Double/ Single outlet for oxygen Delivery	No Change
10	Oxygen tube of 2 m length must be provided with	No Change
11	Facility for nebulization with tube & mask	No Change
12	With one humidifier bottles and one cabinet filters	No Change
13	Unit should function with 200-240Vac, 50/60 Hz input power supply	No Change
14	Should have USFDA/European CE (Issued by notified body) /BIS	No Change

Name of Equipment:- Ultrasonic Nebulizer		
Sl. No	Technical Specification as per tender	Technical Specification after Amendment
a.	Should be light weight, portable, Compact and easy to use.	No Change
b.	Frequency of ultrasonic generator should be greater than 2.5 MHz	No Change
c.	Should have 3 speed nebulization rate control (minimum, medium, maximum)	No Change
d.	Should have a nebulization capacity of 0.3 ml/min.	No Change
e.	Transducer element should have life of at least 5000 hours	No Change
f.	Medication cup capacity should have capacity of maximum 8ml.	No Change
g.	Should uses water as ultrasonic conduction medium, no gel is required	No Change
h.	Should provide silent operation.	No Change
i	Should have a built-in timer.	No Change
j	Should works on 200-240 VAC / 50 Hz.	No Change
k	Should be provided with a complete nebulization kit of 10 Nos. including adult and child mask and medication cup – 5 Nos.	No Change
l	Should be US FDA/European CE (Issued by Notified body)/BIS approved model	No Change

Note- The specification of rest of the equipment remain unchanged.