



Bihar Medical Services & Infrastructure Corporation Limited, 2nd & 3rd Floor, Swasthya Bhawan, Behind IGIMS, Sheikhpura, Adjacent to State Health Society, Patna 800023, Phone/Fax: +91612 2283287,+ 91612 2283288

Corrigendum-I

Bihar Medical Services and Infrastructure Corporation Limited (BMSICL) had invited E-Bids from the interested parties for the procurement, rate contract and the supply of medical equipment for different Govt. Institutions of Bihar vide Notice Inviting Re-Tender No.- BMSICL/2024-25/ME-412. During and after Pre-bid meeting various suggestions were received from different prospective bidders regarding amendment in technical specification of equipment which were discussed and deliberated on by the experts, who after due deliberation recommended certain amendments in the technical specification of the equipment, which are annexed as Annexure-I of this corrigendum. Rest of the terms & conditions of the NIT & Corrigendum-I shall remain unchanged:-

Tender Reference No.	BMSICL/2024-25/ME-412
Last date and time of submission of online bids	08th May 2025 till 17:00 Hrs.
Last date and time of submission of original documents of EMD, Tender Fee and Document	09th May 2025 till 14:00 Hrs.
Date, Time and Place of opening of Technical Bid	09th May 2025 (at 15:00 Hrs.) on the website of https://eproc2.bihar.gov.in in the office of BMSICL
Date and time of opening of financial Bids	To be announced later on https://eproc2.bihar.gov.in

Note:-

- 1. Bidders are advised to refer to the Annexure-I of this corrigendum before submission of bid.**
- 2. Those who have submitted their bids are requested to re-submit their bids in accordance with this corrigendum.**

Annexed:- as above

**Sd/-
GM (Procurement)
BMSICL**

Annexure-I		
Name of Equipment :- Automated Perimeter		
Sl. No	Technical Specification as per tender	Final Amendment
	Specification of Perimeter (Goldman Type) should have following: - High quality Goldman standard Imported automated full field perimeter with bowl size 30cm.Computer monitor should be inbuilt with the perimeter.	No Change
1	Maximum intensity 10,000Asb, Bowl illumination 31.5Asb	No Change
2	Internal hard disk drive, USB with future up gradation to MOD	No Change
3	Stimulation duration 200ms, wavelength Broad band visible light	No Change
4	Stimulus/Background colour White on White	No Change
5	Maximum temporal range 90Deg.& above, Suitable for central 30 as well as full field testing	No Change
6	Central field test patterns 30-2,24-2,10-2,Macula	No Change
7	Threshold test strategies full threshold, FastPac, SITA, SITAFast, SITA Standard	Threshold test strategies full threshold, FastPac, SITA/Similar Threshold strategy, SITAFast/ Similar Threshold strategy, SITA Standard/ Similar Threshold strategy. Similar Threshold strategy should be authenticated with National/International glaucoma study (Journal should be indexed)
8	Screening field test P-60,FF-80,FF-120,FF-240,Nasal Step for periphery	Screening field test P-60,FF-80,FF-120,FF-220 or above,Nasal Step for periphery
9	Screening test strategies Two zone, Three Zone and Quantify Defects	No Change
10	Stimulus Size I-V as per Goldman standards	No Change
11	Glaucoma hemifield test ,Heijl –Krakau blind spot monitor	No Change
12	Video eye monitoring , Trial Lens Holder,	No Change
13	Touch screen monitor inbuilt or computer based systems with Keyboard & Mouse	No Change
14	Motorized chinrest, Motorized table with Laser Jet Printer	No Change
15	Glaucoma progression analysis software	No Change
16	Power supply to be 220-240VAC, 50Hz fitted with Indian plug.	No Change
17	Suitable UPS with maintenance free batteries &Back up time 30 minutes.	No Change
18	US FDA/ European CE (Issued by a notified body) approved Model should be offered	No Change