



Bihar Medical Services & Infrastructure Corporation Limited, 2nd & 3rd Floor, Swasthya Bhawan, Behind IGIMS, Sheikhpura, Adjacent to State Health Society, Patna 800023, Phone/Fax: +91612 2283287,+ 91612 2283288

Corrigendum-I

Bihar Medical Services and Infrastructure Corporation Limited (BMSICL) had invited E-Bids from the interested parties for the procurement, rate contract and the supply of medical equipment for different Govt. Institutions of Bihar vide Notice Inviting Tender No.- BMSICL/2025-26/ME-419. During and after Pre-bid meeting various suggestions were received from different prospective bidders regarding amendment in technical specification of equipment which were discussed and deliberated on by the experts, who after due deliberation recommended certain amendments in the technical specification of the equipment, which are annexed as Annexure-I of this corrigendum. Rest of the terms & conditions of the NIT & Corrigendum-I shall remain unchanged:-

Tender Reference No.	BMSICL/2025-26/ME-419
Last date and time of submission of online bids	26th June 2025 till 17:00 Hrs.
Last date and time of submission of original documents of EMD, Tender Fee and Document	27th June 2025 till 14:00 Hrs.
Date, Time and Place of opening of Technical Bid	27th June 2025 (at 15:00 Hrs.) on the website of https://eproc2.bihar.gov.in in the office of BMSICL
Date and time of opening of financial Bids	To be announced later on https://eproc2.bihar.gov.in

Note:-

- 1. Bidders are advised to refer to the Annexure-I of this corrigendum before submission of bid.**
- 2. Those who have submitted their bids are requested to re-submit their bids in accordance with this corrigendum.**

Annexed:- as above

**Sd/-
GM (Procurement)
BMSICL**

Annexure-I		
Name of Equipment:- Colposcope		
SI. No	Technical Specification as per tender	Final Amendment
1	It Should be Full HD Technology with CMOS Sensor	No Change
2	It Should be HDMI Video transfer on screen	No Change
3	It Should be 1080p Resolution Sensor (1920x1080)/Good quality	No Change
4	It Should be Aspect Ratio 16:9	No Change
5	It Should be 21.2 Mega Pixels	No Change
6	It Should be No need of Dongle & Capture Card	No Change
7	It Should be 5 Grades of green Filter	It Should be 4 or more Grades of green Filter
8	It Should be USB3.0 port for transfer HD video speed of 4.8 gbps to the PC or Laptop.	No Change
9	It Should be Recommended Lamp life 50000 hours or max.	No Change
10	It Should be Highest quality 1-52x Magnification.	No Change
11	It Should be Progressive Scanning Lines.	No Change
12	It Should be Latest Digital Matrix Processor	No Change
13	It Should be Fine Advance & fast Auto/Manual focus Fine Technology	No Change
14	It Should be E Flips & Mirror image Facility.	No Change
15	It Should be Display of genuine tissue color and tone.	No Change
16	It Should be Corner to Corner Uniform Brightness	No Change
17	It Should be Real time display of magnification.	No Change
18	It Should be Acetic acid Test Timer	No Change
19	It Should be Electronic Green Filter with no light loss.	No Change
20	It Should be Black Morphological filter	No Change
21	It Should be Minimum illumination 0.1 Lux.	No Change
22	It Should be Remote control.	No Change
23	It Should be Foot Switch.	No Change
24	(a) USFDA/European CE (issued by notified body)/BIS Certified.	No Change
	(b) 9001:2008/ISO 13485 (from NABCB accredited CBs.)	9001:2008 /ISO 13485 (from IAF/NABCB accredited CBs.)
	Multi- function Image processing software	No Change
1	It Should be integrated Management of View Examination, image capture Freeze, Recording, Observation, Processing.	No Change
2	It Should be Can be integrated to LAN and HIS and HIS.	No Change
3	It Should be Reference pictures library to help user.	No Change
4	It Should be Statistics mode with automatic flow chart facility is provided.	No Change
5	It Should be Printing of Multi Format photo report with Hammond graph Report	No Change
6	It Should be Powerful Comparison to compare.	No Change
7	It Should be Colposcopy REID & SWEDE Evaluation.	No Change
8	It Should be Report can be e-mailed as well.	No Change

9	Digital video colposcope and Image processing software shall be from same manufacturer	No Change
	Computer with printer, ups, trolley.	No Change
1	Should Have 21” color TFT/LDC/LCD flat panel Monitor.	No Change
2	Should Have Microsoft Windows 10.	No Change
3	Should Have 500GB Hard Disk	No Change
4	Should Have Intel Core i7 Processor or better.	No Change
5	Should Have 8GB Ram.	No Change
6	Inkjet Colored Printer.	No Change
7	Trolley with good quality.	No Change