



Bihar Medical Services & Infrastructure Corporation Limited, 2nd & 3rd Floor, Swasthya Bhawan, Behind IGIMS, Sheikhpura, Adjacent to State Health Society, Patna 800023, Phone/Fax: +91612 2283287,+ 91612 2283288

Corrigendum-II

Bihar Medical Services and Infrastructure Corporation Limited (BMSICL) had invited E-Bids from the interested parties for Procurement, Rate contract, Supply, Installation of Medical Equipment vide Tender No.-BMSICL/2025-26/ME-453. During Pre-bid meeting various suggestions were received from different prospective bidders regarding amendment in technical specification of equipment which were discussed and deliberated on by the experts. On the basis of their recommendations certain amendments in the technical specification of the equipment have been made which are annexed as **Annexure-I** of this corrigendum.

Note:-

- 1. Bidders are advised to refer to the Annexure-I of this corrigendum before submission of bid.**
- 2. Those who have submitted their bids are requested to re-submit their bids in accordance with this corrigendum.**

Annexed:- as above

**GM (Procurement)
BMSICL**

Annexure-I		
Name of Equipment :- Anesthesia Trolley		
SI. No	Technical Specification as per tender	Final Amendment
1	Emergency cart constructed of steel/aluminum and high density resin.	No Change
2	Defibrillator shelf with monitor straps, glove dispenser, sharp container, oxygen cylinder cradle, IV pole, cardiac chest board, writing surface.	No Change
3	Clear plastic / ABS plastic overlay for top cap.	No Change
4	Push handle built in to the end panel for smooth and stable movement.	No Change
5	Pullout writing surface top.	No Change
6	Carts should be light, sturdy and scratch resistant.	No Change
7	All drawers should be lockable individually or centrally.	No Change
8	Should have minimum five drawers with adjustable/fixed divides.	No Change
9	Should have side bin discarding syringes and gloves.	No Change
10	Castro should not be less than 5” diameter to facilitate quite and easy maneuverability, dust prevention, flexible transportation.	No Change
11	Size should be: Height: 100 – 110 cm Base should not be less than 60 – 70 cm	No Change
12	USFDA/ CE approved (with four digit notified body number)/ BIS /CDSCO certified	USFDA/ CE approved (with four digit notified body number)/ BIS /CDSCO certified/ISO 13485: 2016
Name of Equipment :- Chemical Balance		
SI. No	Technical Specification as per tender	Final Amendment
1	Maximum Capacity – 200 gms or above.	No Change
2	Sensitivity – 10 mg.	No Change
3	Calibration certificates should be submitted for NABH accreditation every year in warranty period	No Change
4	Pan size should be minimum 100mm x 100mm or equivalent.	No Change
5	Stabilization time should be less than 3 seconds.	No Change
Name of Equipment :- Vacuum with Vacuum Cup		
SI. No	Technical Specification as per tender	Final Amendment
1	Available in manual and electric modes.	No Change
	Manual:	No Change
	a) Silicon cups (40mm, 50mm and 60 mm without valves).	No Change
	b) Autoclavable silicone pressure tube.	No Change
	c) Traction handle.	No Change
	d) Suction pump.	No Change

	e) Fitted Fiber jar (600 ml).	Fitted Fiber jar/Polycarbonate jar (500-600 ml).
	Electric:	No Change
	a) Housing ABS plastic molded.	No Change
	b) Capacity :- 600 mm hg + 10.45 Lt/min	No Change
	c) Pump type: diaphragm with Flat spring Technology.	No Change
	d) 1 x 2 Liters Autoclavable polycarbonate jars having Autoclavable Lids with overflow safety.	No Change
	e) Filter: Bacterial filter autoclavable / reusable	No Change
	f) Quality certificate should be ISO 13485 NABCB Accredited	No Change
Name of Equipment: -Interval Timer		
Sl. No	Technical Specification as per tender	Final Amendment
1	It should be Programmable.	No Change
2	It has maximum count up time & Down time of 23 hours 59 min 59 sec.	No Change
3	Should have large LCD display.	No Change
4	Should have easy setting.	No Change
5	Clock feature with 12/24-hour format.	No Change
6	Manufacturer should have ISO (NABCB accredited) certificate.	No Change