

	Bihar Medical Services & Infrastructure Corporation Limited, 2nd & 3rd Floor, Swasthya Bhawan, Behind IGIMS, Sheikhpura, Adjacent to State Health Society, Patna 800023, Phone/Fax: +91612 2283287,+ 91612 2283288
---	---

Corrigendum-II

Bihar Medical Services and Infrastructure Corporation Limited (BMSICL) had invited E-Bids from the interested parties for the procurement, rate contract and the supply of medical equipment for different Govt. Institutions of Bihar vide Notice Inviting tender No.- BMSICL/2025-26/ME-457. During and after Pre-bid meeting various suggestions were received from different prospective bidders regarding amendment in technical specification of equipment which were discussed and deliberated on by the experts, who after due deliberation recommended certain amendments in the technical specification of the equipment, which are annexed as Annexure-I of this corrigendum. Rest of the terms & conditions of the NIT & Corrigendum-II shall remain unchanged:-

Tender Reference No.	BMSICL/2025-26/ME-457
Last date and time of submission of online bids	30th October 2025 till 17:00 Hrs.
Last date and time of submission of original documents of EMD, Tender Fee and Document	31st October 2025 till 14:00 Hrs.
Date, Time and Place of opening of Technical Bid	31st October 2025 (at 15:00 Hrs.) on the website of https://eproc2.bihar.gov.in in the office of BMSICL
Date and time of opening of financial Bids	To be announced later on https://eproc2.bihar.gov.in

Note:-

- 1. Bidders are advised to refer to the Annexure-I of this corrigendum before submission of bid.**
- 2. Those who have submitted their bids are requested to re-submit their bids in accordance With this corrigendum.**

Annexed:- as above

Sd/-
GM (Procurement)
BMSICL

Annexure-I		
Name of Equipment :- Mili Water Purification System		
SI. No	Technical Specification as per tender	Final Amendment
1	Water purification system should be capable of producing 18.2 megaohm resistivity with pre-treatment cartridge, Reverse Osmosis, and mixed bed ion exchange resins and activated carbon in a pack, with feed water acceptance of up to 2000 micro Siemens conductivity, Fouling Index (SDI)<12, Total Chlorine < 1 ppm.	No Change
2	The system should be able to produce type 1 water and Type 2 directly from Tap water, Type 2 water should have a recirculation motion to avoid contamination.	No Change
3	All required consumables, Spares, and accessories shall be replaced free of cost in the warranty and CMC period. Adequate prefilter shall be provided depending upon the water quality.	No Change
4	Detailed indication of all the components, Accessories & Consumables to be mentioned separately along with the technical bid.	No Change
5	Type 2 water flow rate should be 3 liters/ hour.	No Change
6	Water Specification.	No Change
	a) Conductivity : <0.055 uS/cm	No Change
	b) TOC : <10 ppb	No Change
	c) Bacteria : <0.1 cfu/ml	No Change
	d) Pyrogen : NIL	No Change
	e) RNase : <0.01 ng/ml	No Change
	f) Flow Rate : 0.5 L/min	No Change
	g) Particulates>0.22 u : <1 p/ml.	No Change
7	USFDA/European CE (Issued by Notified body) / BIS/ISO 13485 Approved model should be offered.	No Change