

 BMSICL	Bihar Medical Services & Infrastructure Corporation Limited, 2nd & 3rd Floor, Swasthya Bhawan, Behind IGIMS, Sheikhpura, Adjacent to State Health Society, Patna 800023, Phone/Fax: +91612 2283287,+ 91612 2283288
---	---

Corrigendum-I

Bihar Medical Services and Infrastructure Corporation Limited (BMSICL) had invited E-Bids from the interested parties for the procurement, rate contract and the supply of medical equipment for different Govt. Institutions of Bihar vide Notice Inviting Tender No.- BMSICL/2025-26/ME-460. During and after Pre-bid meeting various suggestions were received from different prospective bidders regarding amendment in technical specification of equipment which were discussed and deliberated on by the experts, who after due deliberation recommended certain amendments in the technical specification of the equipment, which are annexed as Annexure-I of this corrigendum. In order to facilitate maximum participation of bidders the tender schedule is being revised as follows:-

Tender Reference No.	BMSICL/2025-26/ME-460
Last date and time of submission of online bids	18th November 2025 till 17:00 Hrs.
Last date and time of submission of original documents of EMD, Tender Fee and Document	19th November 2025 till 14:00 Hrs.
Date, Time and Place of opening of Technical Bid	19th November 2025 (at 15:00 Hrs.) on the website of https://eproc2.bihar.gov.in in the office of BMSICL
Date and time of opening of financial Bids	To be announced later on https://eproc2.bihar.gov.in

Note:-

- 1. Bidders are advised to refer to the Annexure-I of this corrigendum before submission of bid.**
- 2. Those who have submitted their bids are requested to re-submit their bids in accordance with this corrigendum.**

Annexed:- as above

Sd/-
GM (Procurement)
BMSICL

Annexure-I		
Name of Equipment :- Automated Perimeter		
SI. No	Technical Specification as per tender	Final Amendment
1	Specification of Perimeter (Goldman Type) should have following: - High quality Goldman standard Imported automated full field perimeter with bowl size 30cm.Computer monitor should be inbuilt with the perimeter.	No Change
2	Maximum intensity 10,000Asb, Bowl illumination 31.5Asb.	No Change
3	Internal hard disk drive, USB with future up gradation to MOD.	No Change
4	Stimulation duration 200ms, wavelength Broad band visible light.	No Change
5	Stimulus/Background colour White on White.	No Change
6	Maximum temporal range 90Deg.& above, Suitable for central 30 as well as full field testing.	No Change
7	Central field test patterns 30-2,24-2,10-2,Macula.	No Change
8	Threshold test strategies full threshold, FastPac, SITA/Similar Threshold strategy, SITAFast/ Similar Threshold strategy, SITA Standard/ Similar Threshold strategy. Similar Threshold strategy should be authenticated with National/International glaucoma study (Journal should be indexed).	No Change
9	Screening field test P-60,FF- 80,FF-120,FF-220 or above,Nasal Step for periphery.	No Change
10	Screening test strategies Two zone, Three Zone and Quantify Defects.	No Change
11	Stimulus Size I-V as per Goldman standards.	No Change
12	Glaucoma hemifield test ,Heijl –Krakau blind spot monitor.	No Change
13	Video eye monitoring , Trial Lens Holder.	No Change
14	Touch screen monitor inbuilt or computer based systems with Keyboard & Mouse.	No Change
15	Motorized chinrest, Motorized table with Laser Jet Printer.	No Change
16	Glaucoma progression analysis software.	No Change
17	Power supply to be 220-240VAC, 50Hz fitted with Indian plug.	No Change
18	Suitable UPS with maintenance free batteries &Back up time 30 minutes.	No Change
19	US FDA/ European CE (Issued by a notified body) approved Model should be offered.	No Change