

 BMSICL	Bihar Medical Services & Infrastructure Corporation Limited, 2nd & 3rd Floor, Swasthya Bhawan, Behind IGIMS, Sheikhpura, Adjacent to State Health Society, Patna 800023, Phone/Fax: +91612 2283287,+ 91612 2283288
---	---

Corrigendum-I

Bihar Medical Services and Infrastructure Corporation Limited (BMSICL) had invited E-Bids from the interested parties for the procurement, rate contract and the supply of medical equipment for different Govt. Institutions of Bihar vide Notice Inviting Tender No.- BMSICL/2025-26/ME-473. During and after Pre-bid meeting various suggestions were received from different prospective bidders regarding amendment in technical specification of equipment which were discussed and deliberated on by the experts, who after due deliberation recommended certain amendments in the technical specification of the equipment, which are annexed as Annexure-I of this corrigendum. In order to facilitate maximum participation of bidders the tender schedule is being revised as follows:-

Tender Reference No.	BMSICL/2025-26/ME-473
Last date and time of submission of online bids	24th February 2026 till 17:00 Hrs.
Last date and time of submission of original documents of EMD, Tender Fee and Document	25th February 2026 till 14:00 Hrs.
Date, Time and Place of opening of Technical Bid	25th February 2026 (at 15:00 Hrs.) on the website of https://eproc2.bihar.gov.in in the office of BMSICL
Date and time of opening of financial Bids	To be announced later on https://eproc2.bihar.gov.in

Note:-

- 1. Bidders are advised to refer to the Annexure-I of this corrigendum before submission of bid.**
- 2. Those who have submitted their bids are requested to re-submit their bids in accordance With this corrigendum.**

Annexed: - as above

**GM (Procurement)
BMSICL**

Annexure-I		
Name of Equipment: - Foetal Doppler		
SI. No	Technical Specification as per tender	Final Amendment
1	Should be handy, work on Rechargeable battery.	No Change
2	Should have an ON/OFF switch	No Change
3	Should have an LCD display for FHR.	No Change
4	Should have Auto counting.	No Change
5	Should have probe storage facility to avoid accidental falling.	No Change
6	Should have facility to control the volume of FHR sound.	No Change
7	Should provide rechargeable battery along with recharging unit	No Change
8	Should have Automatic power off facility.	No Change
9	Should have battery level indication.	No Change
10	Should be supplied with a carry pouch.	No Change
11	Should work 220VAC +/- 10 %, 50Hz	No Change
12	US FDA and / European CE/BIS approved model should be offered.	No Change
Name of Equipment:- Phototherapy		
SI. No	Technical Specification as per tender	Final Amendment
1	Phototherapy should be based on advanced LED technology.	No Change
2	In LED technology the irradiance should cover the entire treatment area.	No Change
3	Lamp source should be continuous til table to 90° angle to cover the entire treatment area.	No Change
4	System should be height adjustable with built-in non resettable timer.	No Change
5	It should be supplied with a compatible trolley.	No Change
6	Should work with input 200 to 240Vac 50 Hz supply.	No Change
7	US FDA/ European CE (Issued by a notified body) approved Model should be offered.	US FDA/ European CE (Issued by a notified body) /BIS/ISO 13485 (NABCB accredited) approved model should be offered.