



Bihar Medical Services & Infrastructure Corporation Limited, 3rd Floor, Swasthya Bhawan, Behind IGIMS, Sheikhpura, Adjacent to State Health Society, Patna 800023, Phone/Fax: +91612 2283287,+ 91612 2283288

Corrigendum-I

Bihar Medical Services and Infrastructure Corporation Limited (BMSICL) had invited E-Bids from the interested parties for Procurement, Rate contract, Supply, Installation of Medical Equipment vide Tender No.-BMSICL/2026-27/ME-509. During Pre-bid meeting various suggestions were received from different prospective bidders regarding amendment in technical specification of equipment which were discussed and deliberated on by the experts. On the basis of their recommendations final amendment in technical specification of the equipment have been made which is annexed as **Annexure-I** of this corrigendum. In order to facilitate maximum participation of bidders the tender schedule is being revised as follows:-

Tender Reference No.	BMSICL/2026-27/ME-509
Last date and time of submission of online bids	01st September 2026 till 17:00 Hrs.
Last date and time of submission of original documents of EMD, Tender Fee and Document	02nd September 2026 till 14:00 Hrs.
Date, Time and Place of opening of Technical Bid	02nd September 2026 (at 15:00 Hrs.) on the website of https://eproc2.bihar.gov.in in the office of BMSICL
Date and time of opening of financial Bids	To be announced later on https://eproc2.bihar.gov.in

Annexed:- As above

Sd/-
GM (Procurement)

Annexure-I		
Name of Equipment:- OT Table (Minor)		
Sl. No	Technical Specification as per tender	Final Amendment
1	O T Table with following specs Electro Mechanical Adjustments Height 760 mm to 960 mm.	O T Table with following specs Electro Mechanical Adjustments Height 600 mm to 900 mm.
2	Trendelenburg/Rev Trendelberg, 25/25 deg	No Change
3	Lateral Tilt.-15 deg	No Change
4	Manual Adjustments Backrest with gas spring up/down-65/25 deg	No Change
5	Headrest Detachable Leg rest Detachable Standard.	No Change
6	Four Section Mattress – 1 Set	No Change
7	Accessories	No Change
a	Arm Board 02	No Change
b	Anaesthesia Screen 01.	No Change
c	Body Strap 01	No Change
d	Radial setting clamp 03	No Change
e	Goepel Knee crutches 02	No Change
f	a) Perineal cutout filter 01	No Change
8	Manufacturer should have ISO 13485 (NABCB Accredited) certified	No Change