

File No.- BMSIC/40030/311-2024/7319

दिनांक:-20/11/2024

बिहार चिकित्सा सेवाएं एवं आधारभूत संरचना निगम लि०, बिहार पटना
BMSICL की निविदा संख्या-BMSIC/MEDICAL DEVICES/CONSUMABLES/24-06 से संबंधित प्राप्त ई-निविदाओं के तकनीकी मूल्यांकन विवरणी पर साक्ष्य सहित दावा/आपत्ति की माँग संबंधी सूचना

बिहार चिकित्सा सेवाएं एवं आधारभूत संरचना निगम लि०, द्वारा Medical Devices/Consumables की अधिप्राप्ति हेतु ई-निविदा (निविदा संख्या-BMSIC/MEDICAL DEVICES/CONSUMABLES/24-06) आमंत्रित की गई थी, जिसके आलोक में निर्धारित समय सीमा के अन्तर्गत कुल 51 (इक्यावन) ई-निविदा प्राप्त हुई। उक्त 51 (इक्यावन) ई-निविदा के दस्तावेजों के प्राथमिक तकनीकी मूल्यांकन के उपरान्त तैयार की गई तकनीकी मूल्यांकन विवरणी के दावा/आपत्ति के आमंत्रण हेतु BMSICL के website (www.bmsicl.gov.in) एवं e-proc 2.0 के website (www.eproc2.bihar.gov.in) पर अपलोड की गई है।

निविदा में भाग लेने वाले निर्माता संस्थान अथवा अन्य आपूर्तिकर्ता अपलोड किये गए उपर्युक्त तकनीकी मूल्यांकन प्रपत्र में किसी प्रकार की विसंगति पाये जाने की स्थिति में, साक्ष्य सहित अपना दावा/आपत्ति दिनांक 27/11/2024 के अपराह्न 05:00 बजे तक केवल e-proc 2.0 के website (www.eproc2.bihar.gov.in) के माध्यम से कर सकते हैं।

e-proc 2.0 के के website (www.eproc2.bihar.gov.in) के अलावा अन्य किसी माध्यम से प्राप्त किसी प्रकार के दावा/आपत्ति पर कोई विचार नहीं किया जायेगा।

अनुलग्नक- यथोक्त


(हरेन्द्र राम)
महाप्रबंधक (अधिप्राप्ति)

ज्ञापांक:- BMSIC/40030/311-2024/7319

दिनांक:-20/11/2024

प्रतिलिपि :- प्रबंधक (सिस्टम) एवं प्रबंधक (औषधि) को सूचनार्थ प्रेषित। निदेश है कि सूचना का अनुलग्नक सहित अविलंब BMSICL के website (www.bmsicl.gov.in) एवं e-proc2 (www.eproc2.bihar.gov.in) पर अपलोड करना सुनिश्चित करें।


महाप्रबंधक (अधिप्राप्ति)



TENDER NO. BMSIC/MEDICAL DEVICES/CONSUMABLE/24-06

Check list - I: Evaluation sheet to be used for preliminary evaluation of technical bids

Company Name: Romsons Group Pvt. Ltd. Total Number of Pages Submitted in bid documents: 1 to 348

Sl. No	SBD clause serial No	Description of the Clause/ Technical Eligibility Criteria as per SBD as mentioned in the Check List	Document/ fees submission status (Yes/ No)	Page number	Do the provided documents meet the eligibility criteria ? (Yes/ No)	Remarks
1	3. (a)	Tender Fee (Non-Refundable) of Rs 10,000/- + GST @18%, Total Rs 11,800 shall be submitted online (On E-Proc-2). This fee is payable only once for one tender irrespective of items contained therein.	Yes	161	Yes	
2	3.(b)	EMD shall be submitted ONLY in form of Bank Guarantee (Offline mode), drawn in favor of Managing Director, Bihar Medical Services and Infrastructure Corporation Limited from any Scheduled/Nationalized bank payable at Patna, as mentioned in Clause 2, as per following table:- S.N. No. of MEDICAL DEVICES/CONSUMABLE quoted EMD Amount 1. Upto 5 MEDICAL DEVICES/CONSUMABLE Rs 50,000/- (Fifty thousand only) 2. 6 to 10 MEDICAL DEVICES/CONSUMABLE Rs 1,00,000 (One Lakhs only) 3. 11 to 20 MEDICAL DEVICES/CONSUMABLE Rs 1,50,000 (One Lakhs fifty thousand only) 4. More than 20 MEDICAL DEVICES/CONSUMABLE Rs 2,50,000 (Two Lakhs fifty thousand only)	Yes	124-130	Yes	As per Bid Document clause 3 (b) BG Submitted for EMD must be Scheduled /Nationalized Bank Patna at Patna But firm has not submitted the same
3	3.(c)	Documentary evidence of the constitution of the company/firm/Proprietorship such as Memorandum and Articles of Association, Partnership Deed etc. should be submitted with details of the Name, Address, Telephone Number, Fax Number, e-mail address of the firm and of the Managing Director / Partners / Proprietor should be submitted.	Yes	1	Yes	
4	3.(k)	Copies of the Audited Balance Sheet and Profit and Loss statement showing details of their annual average turnover not less than 5 Crores for any three of the last four consecutive financial years (Auditor/CA certificate of turnover will not be accepted). Self-attested copies are to be submitted.	Yes	180-186	Yes	
5	3.(l)	Copy of Income Tax Return for any three of last four consecutive Assessment years should be submitted (Self Attested).	Yes	177-179	Yes	
6	3.(o)	Copy of PAN Card of the bidder company should be submitted (self-attested).	Yes	176	Yes	
7	3.(p)	Copy of certificate of valid GST registration of the bidder company should be submitted (self-attested).	Yes	163-165	Yes	

Chandra

Seedor

TENDER NO. BMSIC/MEDICAL DEVICES/CONSUMABLE/24-06

Check list - II: Evaluation sheet to be used for preliminary evaluation of technical bids

Company Name: - M/s Romsons Group Pvt. Ltd. Total Number of Pages Submitted in bid documents:- 348

Sl. No	SBD Clause serial No	Description of the Clause/ Technical Eligibility Criteria as per SBD, as mentioned in Check List	Document/ fees submission status (Yes/ No)	Page Number	Do the provided documents meet the eligibility criteria ? (Yes/ No)	Remarks
1	3.(d)	The details of Bidder Name, Address, Telephone Number, Fax Number, e-mail address of the bidder and of the Managing Director / Partners / Proprietor should be submitted in Annexure-V.	Yes	191	Yes	
2	3.(e)	Power of Attorney or Resolution of the Board by which the authorized signatory has been authorized by the bidder firm to sign the documents should be submitted.	Yes	2-3	Yes	
3	3.(f)	(I) Minimum three years old valid Manufacturing License of the quoted product.	Yes	40-63	No	(i) Manufacturing license no. Mfg/MD/2022/000283, MFG/MD/2023/000229 & MD/2022/000125 less than three year old. (ii) Manufacturing license no. MD/2018/000104 valid upto 09.01.2024 and retention not submitted. (iii) Product approval for quoted product NIT sl. no. 51 size 26 is not mention. (iv) In product approval of quoted product NIT sl. no. 52 IS12655 (part- 4 of 2003) & Needle size- 21G (Adult use) is not mention. Pollution certificate of Mfg. unit having license no. MFG/MD/2021/000151, MFG/MD/2023/000229 and MFG/MD/2022/000283 is not submitted.
4		In case of manufacturer, the bidder firm must have a valid manufacturing license or duly acknowledged renewal application with old license issued by the state licensing authority/Central licensing approving authority under Medical Devices Rules 2017 where ever applicable.	Yes	40-63	No	
5		Approved product list as per the license issued for quoted MEDICAL DEVICES/CONSUMABLE as Per Medical Devices Rules 2017.	Yes	40-63	No	
6		Manufacturing License along with approved product list must be valid till the last date of the submission of tender.	Yes	40-63	No	
7		Valid Pollution Control Clearance Certificate in accordance with Water [Prevention and control of Pollution] Act, 1974 & Air [Prevention and control of Pollution] Act, 1981 and Hazardous Wastes (Management, Handling & Trans Boundary Movement) Rules 2008 (Self Attested Copy of Certificate to be enclosed) (as applicable).	Yes	64-68	No	
8		Bidders shall submit self-attested copies of required manufacturing license and approved product list in support of above-mentioned condition and they are required to specify the quoted product in their approved product list by highlighting it.	Yes	40-63	Yes	

BMM
28/10/24AMM
29/10/24Rajesh
01/11/24

9		(II) In case of Importer, the bidder (importer) firm must have a valid import License with product registration certificate issued by the Drugs controller General of India as per Medical Device Rule 2017. All Quoted products should be accompanied by their invoices, statement and import License showing that the quoted product are being imported and sold in India by the bidder (Importer) firm minimum for Last three years. Import license must be valid on the last date of submission of tender.	NA	NA	NA	
10		(III) In case of Non-drug item(s) where neither the D & C Act 1945 nor the Medical Device Rule 2017 is applicable the bidder must have a manufacturing license/import export certificate (IEC) with an undertaking/Self declaration in his letter pas as per Annexure-VII that the item(s) quoted by the bidder is/are non-drug item(s) i.e. neither covered under D & C Act nor Under Medical Device Rule 2017.	NA	NA	NA	
11	3.(g)	Bidder must have Market Standing Certificate of minimum three years issued by the concerned Licensing Authority from Drugs Control Department/Concerned Government Department for the quoted product. Self-attested copies are to be submitted.	Yes	131-160	No	(i) Market standing certificate is less than three year of all quoted products. (ii) MSC of quoted product NIT sl. No. 51 is not submitted. (iii) MSC is not submitted for quoted product NIT sl. No. 33.
12	3.(h)	Non-Conviction Certificate (NCC) issued by the concerned Licensing Authority from Drugs Control Administration of the state for last three years should be submitted (wherever applicable). It should be not more than one year old. Self-attested copies are to be submitted.	Yes	21-39	No	Non-conviction certificate not submitted for Manufacturing license no. MFG/MD/2022/000283 & MFG/MD/2023/000229
13	3.(i)	WHO-GMP/GMP (Good Manufacturing Practice) as per revised Schedule- 'M'/COPP Certificate of the manufacturing unit issued by the Licensing Authority/Drugs Control Department, (wherever applicable). The GMP certificate must not be older than one year from the date of publication of tender. Self-attested copies are to be submitted. Explanation- Generally the GMP Certificate is issued for one-year validity. Hence the provision that it should not be older than one year from the date of publication of tender implies mutatis mutandis that the GMP certificate should remain valid till the date of submission of tender.	Yes	78-85	Yes	QMS & ISO 13485:2016 is submitted.
14	3.(j)	Maximum Production Capacity Certificate (Section wise) issued by concerned Licensing Authority from Drugs Control Department highlighting the quoted product section (wherever applicable). Self-attested copies are to be submitted. In case of Importer, an affidavit (with Stamp) sworn before first class magistrate/ Notary stating the batch production capacity of the firm and also that said production (Importing) capacity shall be adequate for requirement laid in NIT. Importer will have also to submit Invoices/evidence of import in items of said product with quantity details.	Yes	115-123	Yes	

BALM
28/10/24

Handwritten signature
29/10/24

B Patel
04/11/24

15	3.(m)	The tenderer should give an affidavit sworn before first class magistrate / Notary stating that the firm & its quoted product is not black listed/Debarred currently (as on the date of submission of the tender) by Central Government / Central Government agencies/any state government or any of the state government agencies / or any Drug procurement agencies or by BMSICL as per Annexure-II .	Yes	12-19	Yes	
16	3.(n)	List of items quoted in prescribed format as per Annexure-III duly signed.	Yes	4-11	Yes	
17	5.(k)	Affidavit declaration regarding acceptance of tender conditions to be submitted by the bidding firm as per Annexure-IV .	Yes	187-190	Yes	
18	5.(l)	Filled check list as per given Annexure-VI to be submitted at the time of uploading the bid.	Yes	343-344	Yes	
19	7.(c)	Power of Attorney or Resolution of the Board by which the authorized signatory has been authorized by the bidder firm should sign the documents in cases where person other than the Managing Director/Managing Partner or sole Proprietor signs the document.	Yes	2-3	Yes	
20	10.(c)	If a particular document/Certificate to be uploaded as specified in bid, is not applicable for a bidder, the bidder shall attach a scanned copy of declaration in the letter head stating that the specific document is not applicable/exempted for the bidder in connection to this tender.	NA	NA	NA	
21	2 (d)	The price quoted by the bidders must not exceed the ceiling price as fixed by NPPA (National Pharmaceutical Pricing Authority) as per the provision of "Drugs Price Control Order (wherever applicable)" and the quoted rate should be at least 20% less than its MRP. In extraordinary cases the Managing Director, BMSICL has discretion to take decision). Explanation- In order to ensure procurement of the tendered drugs at the most appropriate and competitive rate, the bidders are directed to quote their lowest price as compared to the Rate provided to their respective Distributors/ Dealers/Wholesalers/ Carrying and Forwarding Agents/Authorized depot sales point in the State of Bihar. A Notarized affidavit to this effect on a Rs 100/- Non-judicial Stamp Paper should be submitted with the Bid in the prescribed format as mentioned in Annexure-X .	Yes	192-193	Yes	
22	—	Performance Statement- Annexure- IX	Yes	194-342	Yes	
23	—	AFFIDAVIT FOR NON DRUG ITEM(S) ANNEXURE-VII	NA	NA	NA	

BAN
28/10/24

AKP
29/10/24

Prasad
04/11/24

TENDER NO. BMSIC/MEDICAL DEVICES/CONSUMABLE/24-06

Check list - III: Evaluation sheet to be used for preliminary evaluation of technical bids

Company Name: - M/s Romsons Group Pvt. Ltd. Total Number of Pages Submitted in bid documents:- 348

Sheet to be used for verification of product approval and market standing

Sl. No	NIT Sl. No	Name of the Quoted MEDICAL DEVICES/CONSUMABLE		Specification		Pack Size		Product category	Approval Details			Last 3yrs market standing certificate/ New Drug
		As per NIT	As per Approval	As per NIT	As per Approval	As per NIT	As per Approval	(Medical Device/Non Drug)	Product Approval	Approved Upto	Approved in Brand /Generic Name	
1	5	ADK Drain Sizes 20"	Abdominal Drainage tube	Specially designed for post-operative abdominal drainage. Soft and smooth catheter with large atraumatic eyes for efficient drainage. Radio-opaque line throughout its length help x-ray visualization. Specially designed foldable handle suitable for bed side hanging. Colour coded for instant size identification. Sterile, individually packed in a peelable pouch. Soft drainage catheter 50 cm long. Collection bag of 2000 ml capacity.	Sterilized, Dimension- FG- 16, 20, 24, 28 & 32 Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
2	6	ADK Drain Sizes 24"	Abdominal Drainage tube	Specially designed for post-operative abdominal drainage. Soft and smooth catheter with large atraumatic eyes for efficient drainage. Radio-opaque line throughout its length help x-ray visualization. Specially designed foldable handle suitable for bed side hanging. Colour coded for instant size identification. Sterile, individually packed in a peelable pouch. Soft drainage catheter 50 cm long. Collection bag of 2000 ml capacity.	Sterilized, Dimension- FG- 16, 20, 24, 28 & 32 Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
3	7	ADK Drain Sizes 28"	Abdominal Drainage tube	Specially designed for post-operative abdominal drainage. Soft and smooth catheter with large atraumatic eyes for efficient drainage. Radio-opaque line throughout its length help x-ray visualization. Specially designed foldable handle suitable for bed side hanging. Colour coded for instant size identification. Sterile, individually packed in a peelable pouch. Soft drainage catheter 50 cm long. Collection bag of 2000 ml capacity	Sterilized, Dimension- FG- 16, 20, 24, 28 & 32 Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
4	8	ADK Drain Sizes 32"	Abdominal Drainage tube	Specially designed for post-operative abdominal drainage. Soft and smooth catheter with large atraumatic eyes for efficient drainage. Radio-opaque line throughout its length help x-ray visualization. Specially designed foldable handle suitable for bed side hanging. Colour coded for instant size identification. Sterile, individually packed in a peelable pouch. Soft drainage catheter 50 cm long. Collection bag of 2000 ml capacity	Sterilized, Dimension- FG- 16, 20, 24, 28 & 32 Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.

BAM
28/10/24HWP
29/10/24Raj
04/11/24

5	20	Blood Transfusion Set with Filter 170 Micron	Transfusion Sets for single use	(The soft Kink resistance, translucent tubing is prepared from medical grade PVC material, double drip clearly visible chamber facilitate visual access and rapid adjustment of fluid level. Specially design roller controller offer accurate regulation of infusion rate with self sealing latex bulb to avoid any contamination and Easy flushing. Length 150 cm.)	Filter specification 170 micron not mention in product approval	Each Piece	Each Piece	MD	01.06.2021	31.05.2026	Both	Less than 3 years.
6	23	Central Venous Catheter Central Line- Monoluminal, Adult use	Central Venous Cathetar Kit	Central Venous Catheter, monoluminal, Adult use, Radio-opaque polyurethane catheter, with soft tip, length markings, junction hub for stitch fixation and fully transparent extension lines, with colorcoded hub and white slide clamps. Components: Kink-proof guide wire with "J" tip and straight tip on dispenser, movable fixation wing with clamp, introducer needle, scalpel, dilator, valve, 5 ml luer lock syringe (depending on set configuration).	Sterilized, Dimension- Single lumen 5 Fr x 20cm	Each Piece	Not mentioned	MD	10.01.2019	09.01.2024 (Current valid license not submitted)	Both	Less than 3 years.
7	24	Central Venous Catheter Central Line-Biluminal, Adult use	Central Venous Cathetar Kit	Central Venous Catheter biluminal, Adult use, Radio-opaque polyurethane catheter, with soft tip, length markings, junction hub for stitch fixation and fully transparent extension lines, with color-coded hub and white slide clamps. Components: Kink-proof guide wire with "J" tip and straight tip on dispenser, movable fixation wing with clamp, introducer needle, scalpel, dilator, valve, 5 ml luer lock syringe (depending on set configuration)	Sterilized, Dimension- Double lumen 5 Fr x 8cm, 7 Fr x 16cm, 7 Fr x 20cm	Each Piece	Not mentioned	MD	10.01.2019	09.01.2024 (Current valid license not submitted)	Both	Less than 3 years.
8	25	Central Venous Catheter Central Line-Triluminal, Adult Use	Central Venous Cathetar Kit	Central Venous Catheter Triluminal, Adult Use, Radio-opaque polyurethane catheter, with soft tip, length markings, junction hub for stitch fixation and fully transparent extension lines, with color-coded hub and white slide clamps. Components: Kink-proof guide wire with "J" tip and straight tip on dispenser, movable fixation wing with clamp, introducer needle, scalpel, dilator, valve, 5 ml luer lock syringe (depending on set configuration)	Sterilized, Dimension- Single lumen 5 Fr x 20cm, Double lumen 5 Fr x 8cm, 7 Fr x 16cm, 7 Fr x 20cm, Triple lumen 7 Fr x 16cm, 7 Fr x 20cm Shelflife- 5 years	Each Piece	Not mentioned	MD	10.01.2019	09.01.2024 (Current valid license not submitted)	Both	Less than 3 years.
9	26	Chest Tube No.- 24	Thoracic Drainage Catheter	(Made up of Biocompatible, non-Toxic PVC, proximal end fitted with tapered tongue connector and open, distal end, smoothly finished large eyes for efficient drainage, Graduated at 2 cm intervals from the beginning of eyes length 45 cm.)	Sterilized, Dimension- Straigh & Right Angle- FG- 10, 12, 16, 20, 24, 28, 32, 36 & 40, Made up of PVC, MB, PP Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.

B.M.
28/10/24

M.H.D.
29/10/24

Patel
04/11/24

10	27	Chest Tube No.- 28	Thoracic Drainage Catheter	(Made up of Biocompatible, non-Toxic PVC, proximal end fitted with tapered tongue connector and open, distal end, smoothly finished large eyes for efficient drainage, Graduated at 2 cm intervals from the beginning of eyes length 45 cm.)	Sterilized, Dimension- Straigh & Right Angle- FG- 10, 12, 16, 20, 24, 28, 32, 36 & 40, Made up of PVC, MB, PP Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
11	28	Chest Tube No.- 32	Thoracic Drainage Catheter	(Made up of Biocompatible, non-Toxic PVC, proximal end fitted with tapered tongue connector and open, distal end smoothly finished large eyes for efficient drainage, Graduated at 2 cm intervals from the beginning of eyes length 45 cm.)	Sterilized, Dimension- Straigh & Right Angle- FG- 10, 12, 16, 20, 24, 28, 32, 36 & 40, Made up of PVC, MB, PP Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
12	29	Cord Clamp	Umbilical Occlusion Device	EO Sterile, Safe Locking of cord Clamp, Individual pack and for Single Use	Polypropylene (PP) Dimension- Standard Sterilized Shelflife- 5 years	Per Piece	Not mentioned	MD	17.06.2022 (Less than 3 year)	16.06.2027	Both	Less than 3 years.
13	33	Disposable Mask	Public Respiratory 3ply face mask	Fabric & Construction: - It must have a 3- ply type construction consisting of: a. First layer – 30 GSM – SS b. Second layer – 25 GSM – Melt blown c. Third layer – 25 GSM– SMS UV – C Sanitized Sealing: - The mask should be heat sealed and no stitching to keep three layers together. • Size: - Not less than 17cm x 09cm of mask and having comfortable elastic ear loop. The length of elastic loop should be minimum of 24cm in normal condition i.e., without stretching. • Nose Clip: - Rust proof adjustable nose clip. • Ear loop:- Must be with ultrasonic welding (No manual welding or stitching is permitted). • Packaging: - 25 nos. in individual poly packaging with adequate labelling/noting • Must be ISI marked or European CE/ USFDA Approved. • Breathability: - Differential pressure (Delta P) As per IS 16289 Standard. • Splash Resistance: - As per ASTM F1862 Standard. • Particulate Filtration efficiency at 0.3-microns: - As per ASTM F2299. • Bacterial Filtration Efficiency: - As per ASTM F2101. • Flammability: - As per 16 CFR. • All batches to be supplied need to be accompanied with certificate of analysis from laboratories designated for this purpose by the CDSO/Ministry of Textiles/Health, Government of India	Non-Sterilized, Dimension- Standard Made up of PVC, MB, PP Polypropylene Spunbond non woven fabric, Polypropylene melt blown fabric Non woven elastic, metal wire with polypropylene coating, IS 16289.	Each Piece	Not mentioned	MD	21.03.2023 (Less than 3 year)	20.03.2028	Both	Less than 3 years.

BAM
28/10/24

29/10/24

04/11/24

14	35	Epidural Set with catheter & fitter	Epidural Catheter kit	Epidural set with catheter & fitter made of non-toxic and non-irritant material. Size of Epidural Catheter 16 G, length 100 cm and outer diameter 2 mm. Size of Epidural Needle 16 G and length of needle 80 mm. Plastic LOR syringe 10 ml provided for clear identification. Provided with detachable connector, Separate twist lock connector is provided for safe and secure attachment to the catheter. Sterile and single used. All components are packed in Transparent individual tray packed kit	Sterilized, Dimension- NG- 16 & 18, Made up of PEBAX, ABS, PC, PE, SE, MB, PP, TPE & PU	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
15	36	Folley's Cathetar - Assorted Sizes 10	Foley Catheter	(Sterile Catheter, Single Use, for Urinary Drainage (Foley Balloon Catheter), 2 Way, Size 10 FG Made of Silicone elastomer bonded with Latex Should have hard plastic valve Smooth distal end with smooth eyes for atraumatic intubation Symmetrical foley balloon capacity 3- 5 ml Should conform to IS 11497 Color coding marking to identify size Length, wall thickness and balloon capacity should be mentioned as per IS 11497. Specification for B,C,D,E,F,G should be mentioned as per IS 11497 for particular size)	Sterilized, Dimension- 10 Latex, IS 11497	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
16	37	Folley's Cathetar - Assorted Sizes 12	Foley Catheter	(Sterile Catheter, Single Use, for Urinary Drainage (Foley Balloon Catheter), 2 Way, Size 12 FG Made of Silicone elastomer bonded with Latex Should have hard plastic valve Smooth distal end with smooth eyes for atraumatic intubation Symmetrical foley balloon capacity 30- 50 ml Should conform to IS 11497 Color coding marking to identify size Length, wall thickness and balloon capacity should be mentioned as per IS 11497. Specification for B,C,D,E,F,G should be mentioned as per IS 11497 for particular size)	Sterilized, Dimension- 12 Latex, IS 11497	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
17	38	Folley's Cathetar - Assorted Sizes 14	Foley Catheter	(Sterile Catheter, Single Use, for Urinary Drainage (Foley Balloon Catheter), 2 Way, Size 14 FG Made of Silicone elastomer bonded with Latex Should have hard plastic valve Smooth distal end with smooth eyes for atraumatic intubation Symmetrical foley balloon capacity 30- 50 ml Should conform to IS 11497 Color coding marking to identify size Length, wall thickness and balloon capacity should be mentioned as per IS 11497. Specification for B,C,D,E,F,G should be mentioned as per IS 11497 for particular size)	Sterilized, Dimension- 14 Latex, IS 11497	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.

Bm
28/10/24

29/10/24

04/11/24

18	39	Folley's Cathetar - Assorted Sizes 16	Foley Catheter	(Sterile Catheter, Single Use, for Urinary Drainage (Foley Balloon Catheter), 2 Way, Size 16 FG Made of Silicone elastomer bonded with Latex Should have hard plastic valve Smooth distal end with smooth eyes for atraumatic intubation Symmetrical foley balloon capacity 30 ml Should conform to IS 11497 Color coding marking to identify size Length, wall thickness and balloon capacity should be mentioned as per IS 11497. Specification for B,C,D,E,F,G should be mentioned as per IS 11497 for particular size)	Sterilized, Dimension- 16 Latex, IS 11497	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
19	40	Folley's Cathetar - Assorted Sizes 18	Foley Catheter	(Sterile Catheter, Single Use, for Urinary Drainage (Foley Balloon Catheter), 2 Way, Size 18 FG Made of Silicone elastomer bonded with Latex Should have hard plastic valve Smooth distal end with smooth eyes for atraumatic intubation Symmetrical foley balloon capacity 30 ± 1 ml Should conform to IS 11497 Color coding marking to identify size Length, wall thickness and balloon capacity should be mentioned as per IS 11497. Specification for B,C,D,E,F,G should be mentioned as per IS 11497 for particular size)	Sterilized, Dimension- 18 Latex, IS 11497	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
20	41	Folley's Cathetar - Assorted Sizes 20	Foley Catheter	(Sterile Catheter, Single Use, for Urinary Drainage (Foley Balloon Catheter), 2 Way, Size 20 FG Made of Silicone elastomer bonded with Latex Should have hard plastic valve Smooth distal end with smooth eyes for atraumatic intubation Symmetrical foley balloon capacity 30 ± 1 ml Should conform to IS 11497 Color coding marking to identify size Length, wall thickness and balloon capacity should be mentioned as per IS 11497. Specification for B,C,D,E,F,G should be mentioned as per IS 11497 for particular size)	Sterilized, Dimension- 20 Latex, IS 11497	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
21	42	Folley's Cathetar - Assorted Sizes 6	Foley Catheter	(Sterile Catheter, Single Use, for Urinary Drainage (Foley Balloon Catheter), 2 Way, Size 6 FG Made of Silicone elastomer bonded with Latex Should have hard plastic valve Smooth distal end with smooth eyes for atraumatic intubation Symmetrical foley balloon capacity 3- 5 ml Should conform to IS 11497 Color coding marking to identify size Length, wall thickness and balloon capacity should be mentioned as per IS 11497. Specification for B,C,D,E,F,G should be mentioned as per IS 11497 for particular size)	Sterilized, Dimension- 6 Latex, IS 11497	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
22	43	Folley's Cathetar - Assorted Sizes 8	Foley Catheter	(Sterile Catheter, Single Use, for Urinary Drainage (Foley Balloon Catheter), 2 Way, Size 8 FG Made of Silicone elastomer bonded with Latex Should have hard plastic valve Smooth distal end with smooth eyes for atraumatic intubation Symmetrical foley balloon capacity 3- 5 ml Should conform to IS 11497 Color coding marking to identify size Length, wall thickness and balloon capacity should be mentioned as per IS 11497. Specification for B,C,D,E,F,G should be mentioned as per IS 11497 for particular size)	Sterilized, Dimension- 8 Latex, IS 11497	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
23	44	Infant Feeding Tube Assorted Sizes 5	Feeding Tube	Sterile, Pyrogen free, non-Toxic with Radio Opaque Line, Soft, Frosted, Kink resistant PVC tubing. Minimum length 50 cm.	Sterilized, Dimension- FG- 5, 6, 7, 8, 9, 10, PVC Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.

BAN
28/10/24

Adh
29/10/24

Rany
04/11/24

24	45	Infant feeding Tube Assorted Sizes 6	Feeding Tube	Sterile, Pyrogen free, non-Toxic with RadioOpaque Line, Soft, Frosted, Kink resistant PVC tubing. Minimum length 50 cm	Sterilized, Dimension- FG- 5, 6, 7, 8, 9, 10, PVC Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
25	46	Infant feeding Tube Assorted Sizes 8	Feeding Tube	Sterile, Pyrogen free, non-Toxic with RadioOpaque Line, Soft, Frosted, Kink resistant PVC tubing. Minimum length 50 cm	Sterilized, Dimension- FG- 5, 6, 7, 8, 9, 10, PVC Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
26	47	IV Canula No. 18	Intravenous Cannula	(Sterile Disposable (single Use) Teflon / PTFE IV Cannula with injection valve/port with wings. Size 18 . Should be pack in Transparent, single blister pack. Should conform to IS 10555 Standard)	Sterilized, Dimension- G- 14, 16, 18, 20, 22, 24, Made up of PEP, SS, MB, PP, LDPE, Silicon, ISO- 10555- 1 & 5	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
27	48	IV Canula No. 20	Intravenous Cannula	(Sterile Disposable (single Use) Teflon / PTFE IV Cannula with injection valve/port with wings. Size 20. Should be pack in Transparent, single blister pack. Should conform to IS 10555 Standard)	Sterilized, Dimension- G- 14, 16, 18, 20, 22, 24, Made up of PEP, SS, MB, PP, LDPE, Silicon, ISO- 10555- 1 & 5	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
28	49	IV Canula No. 22	Intravenous Cannula	(Sterile Disposable (single Use) Teflon / PTFE IV Cannula with injection valve/port with wings. Size 22. Should be pack in Transparent, single blister pack. Should conform to IS 10555 Standard)	Sterilized, Dimension- G- 14, 16, 18, 20, 22, 24, Made up of PEP, SS, MB, PP, LDPE, Silicon, ISO- 10555- 1 & 5	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
29	50	IV Canula No. 24	Intravenous Cannula	(Sterile Disposable (single Use) Teflon / PTFE IV Cannula with injection valve/port with wings. Size 24. Should be pack in Transparent, single blister pack. Should conform to IS 10555 Standard)	Sterilized, Dimension- G- 14, 16, 18, 20, 22, 24, Made up of PEP, SS, MB, PP, LDPE, Silicon, ISO- 10555- 1 & 5	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
30	51	IV Canula No. 26	Intravenous Cannula	(Sterile Disposable (single Use) Teflon / PTFE IV Cannula with injection valve/port with wings. Size 26. Should be pack in Transparent, single blister pack. Should conform to IS 10555 Standard)	Sterilized, Dimension- G- 14, 16, 18, 20, 22, 24, Made up of PEP, SS, MB, PP, LDPE, Silicon, ISO- 10555- 1 & 5 (Size- 26 is not mention in product approval)	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	MSC is not submitted.

B.M.
28/10/24

A.K.P.
29/10/24

E.P.
04/11/24

31	52	IV set - Adult	Infusion sets for single use	(Sterile, Disposable, Nontoxic, Non-Pyrogenic, Sterilized by ETO, Perfusion Set (Infusion set) with Built Airway moulded chamber and Needle- 21G (Adult Use). Drop size of approx. 15-20 drops per ml. Injection port. Soft and kink resistant PVC tubing. Roller controller for flow control Tube, length not less than 150 cm. Should conform to IS No.12655 (part-4 of 2003) As per Drugs and Cosmetics Act 1940)	Dimension- STD, Vented or NonVented, with or without Needle Sterilized, Made up of LDPE+MB, ABS+MB, ABST, PVC (DEHP Fr), Polyamid, latex Isoprene, PP+MB, SS (IS 12655 (part- 4 of 2003) & needle- 21G (Adult use) is not mention)	Each Piece	Not mentioned	MD	01.06.2021 (Less than 3 year)	31.05.2026	Both	Less than 3 years.
32	55	PM Line	Extension sets	Pressure Monitoring Line/High Pressure Extension Line. Suitable for high pressure Monitoring and for connection between syringe infusion pump and patient. Male luer lock at one end and female luer lock at other end should fit all standard equipment. Luer lock connectors should provide secure fitting, Pressure upto 800psi, Length 200cm, Sterile	Dimension- 10, 20, 25, 30, 50, 100, 150, 200, 250, 300, 350, 400 cm. Type, Plain, spiral, stop cock, (Red/Blue/Yellow) Sterilized, Made up of PC, PVC, ABS, MB Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
33	63	Rhyles Tube - Assorted Sizes 10	Rhyles Tube	Rhyles Tube / Nasogastric Tube (P.V.C.) with ratio Opaque Lining Size 10. Soft kink, resistance PVC tubing for atraumatic intubation. Closed distal end should be coned with radio opaque Material for accurate intubation. Four lateral eyes for greater efficiency. Radio Opaque Line. Marking at 50, 60,70 cm from trip. Colour coded funnel. With luer connector / closure. Length 105 cm.	Dimension- FG- 8, 10, 12, 14, 16, 18, 20 Sterilized, PVC Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
34	64	Rhyles Tube - Assorted Sizes 12	Rhyles Tube	Rhyles Tube / Nasogastric Tube (P.V.C.) with ratio Opaque Lining Size 12. Soft kink, resistance PVC tubing for atraumatic intubation. Closed distal end should be coned with radio opaque Material for accurate intubation. Four lateral eyes for greater efficiency. Radio Opaque Line. Marking at 50, 60,70 cm from trip. Colour coded funnel. With luer connector / closure. Length 105 cm.	Dimension- FG- 8, 10, 12, 14, 16, 18, 20 Sterilized, PVC Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
35	65	Rhyles Tube - Assorted Sizes 14	Rhyles Tube	Rhyles Tube / Nasogastric Tube (P.V.C.) with ratio Opaque Lining Size 14. Soft kink, resistance PVC tubing for atraumatic intubation. Closed distal end should be coned with radio opaque Material for accurate intubation. Four lateral eyes for greater efficiency. Radio Opaque Line. Marking at 50, 60,70 cm from trip. Colour coded funnel. With luer connector / closure. Length 105 cm.	Dimension- FG- 8, 10, 12, 14, 16, 18, 20 Sterilized, PVC Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
36	66	Rhyles Tube - Assorted Sizes 16	Rhyles Tube	Rhyles Tube / Nasogastric Tube (P.V.C.) with ratio Opaque Lining Size 16. Soft kink, resistance PVC tubing for atraumatic intubation. Closed distal end should be coned with radio opaque Material for accurate intubation. Four lateral eyes for greater efficiency. Radio Opaque Line. Marking at 50, 60,70 cm from trip. Colour coded funnel. With luer connector / closure. Length 105 cm.	Dimension- FG- 8, 10, 12, 14, 16, 18, 20 Sterilized, PVC Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.
37	67	Rhyles Tube - Assorted Sizes 18	Rhyles Tube	(Rhyles Tube / Nasogastric Tube (P.V.C.) with ratio Opaque Lining Size 18. Soft kink, resistance PVC tubing for atraumatic intubation. Closed distal end should be coned with radio opaque Material for accurate intubation. Four lateral eyes for greater efficiency. Radio Opaque Line. Marking at 50, 60,70 cm from trip. Colour coded funnel. With luer connector / closure. Length 105 cm.)	Dimension- FG- 8, 10, 12, 14, 16, 18, 20 Sterilized, PVC Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year)	13.03.2027	Both	Less than 3 years.

BMM
28/10/24

APL
29/10/24

Spared
04/11/24

38	69	Rhyles Tube - Assorted Sizes 8	Rhyles Tube	Rhyles Tube / Nasogastric Tube (P.V.C.) with ratio Opaque Lining Size 8. Soft kink, resistance PVC tubing for atraumatic intubation. Closed distal end should be coned with radio opaque Material for accurate intubation. Four lateral eyes for greater efficiency. Radio Opaque Line. Marking at 50, 60,70 cm from trip. Colour coded funnel. With luer connector / closure. Length 105 cm.	Dimension- FG- 8, 10, 12, 14, 16, 18, 20 Sterilized, PVC Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year	13.03.2027	Both	Less than 3 years.
39	70	Spinal Needle - Assorted Sizes 18	Spinal Needle	Specially designed to administer lumbar/ subarachnoid anaesthesia. Transparent hub provides rapid detection of Cerebro-spinal fluid (CSF) flashback for confirming accurate placement. Sharp bevel design for low puncture force ensures minimal puncture trauma. Fine gauze needle design greatly reduces the risk of PDPH (Post Dura Puncture Headache). Color coded style enables rapid size identification.) Sterile, individually packed in hard blister pack. /Peelable Pouche. Gauze: 1.2 mm and colour Pink	Dimension- NG- 18, 20, 21, 22, 23, 24, 25, 26, 27, (Quincke Type Bevel Point, Pencil Point with Introducer guide Needle). Sterilized Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year	13.03.2027	Both	Less than 3 years.
40	71	Spinal Needle - Assorted Sizes 20	Spinal Needle	Specially designed to administer lumbar/ subarachnoid anaesthesia. Transparent hub provides rapid detection of Cerebro-spinal fluid (CSF) flashback for confirming accurate placement. Sharp bevel design for low puncture force ensures minimal puncture trauma. Fine gauze needle design greatly reduces the risk of PDPH (Post Dura Puncture Headache). Color coded style enables rapid size identification.) Sterile, individually packed in hard blister pack. /Peelable Pouche. Gauze: 0.9 mm and colour Yellow	Dimension- NG- 18, 20, 21, 22, 23, 24, 25, 26, 27, (Quincke Type Bevel Point, Pencil Point with Introducer guide Needle). Sterilized Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year	13.03.2027	Both	Less than 3 years.
41	72	Spinal Needle - Assorted Sizes 22	Spinal Needle	Specially designed to administer lumbar/ subarachnoid anaesthesia. Transparent hub provides rapid detection of Cerebro-spinal fluid (CSF) flashback for confirming accurate placement. Sharp bevel design for low puncture force ensures minimal puncture trauma. Fine gauze needle design greatly reduces the risk of PDPH (Post Dura Puncture Headache). Color coded style enables rapid size identification.) Sterile, individually packed in hard blister pack. /Peelable Pouche. Gauze: 0.7 mm and colour Black	Dimension- NG- 18, 20, 21, 22, 23, 24, 25, 26, 27, (Quincke Type Bevel Point, Pencil Point with Introducer guide Needle). Sterilized Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year	13.03.2027	Both	Less than 3 years.
42	73	Spinal Needle - Assorted Sizes 23	Spinal Needle	Specially designed to administer lumbar/ subarachnoid anaesthesia. Transparent hub provides rapid detection of Cerebro-spinal fluid (CSF) flashback for confirming accurate placement. Sharp bevel design for low puncture force ensures minimal puncture trauma. Fine gauze needle design greatly reduces the risk of PDPH (Post Dura Puncture Headache). Color coded style enables rapid size identification.) Sterile, individually packed in hard blister pack. /Peelable Pouche. Gauze: 0.6 mm and colour Deep blue	Dimension- NG- 18, 20, 21, 22, 23, 24, 25, 26, 27, (Quincke Type Bevel Point, Pencil Point with Introducer guide Needle). Sterilized Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year	13.03.2027	Both	Less than 3 years.

BM
28/10/24

Handwritten signature
29/10/24

Handwritten signature
04/11/24

43	74	Spinal Needle - Assorted Sizes 25	Spinal Needle	Specially designed to administer lumbar/ subarachnoid anaesthesia. Transparent hub provides rapid detection of Cerebro-spinal fluid (CSF) flashback for confirming accurate placement. Sharp bevel design for low puncture force ensures minimal puncture trauma. Fine gauze needle design greatly reduces the risk of PDPH (Post Dura Puncture Headache). Color coded style enables rapid size identification.) Sterile, individually packed in hard blister pack. /Peelable Pouch. Gauze: 0.5 mm and colour Orange	Dimension- NG- 18, 20, 21, 22, 23, 24, 25, 26, 27, (Quincke Type Bevel Point, Pencil Point with Introducer guide Needle). Sterilized Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year	13.03.2027	Both	Less than 3 years.
44	75	Spinal Needle - Assorted Sizes 26	Spinal Needle	(Specially designed to administer lumbar/ subarachnoid anaesthesia. Transparent hub provides rapid detection of Cerebro-spinal fluid (CSF) flashback for confirming accurate placement. Sharp bevel design for low puncture force ensures minimal puncture trauma. Fine gauze needle design greatly reduces the risk of PDPH (Post Dura Puncture Headache). Color coded style enables rapid size identification. Sterile, individually packed in hard blister pack/peelable pouch. Gauze: 0.45 mm and colour Brown)	Dimension- NG- 18, 20, 21, 22, 23, 24, 25, 26, 27, (Quincke Type Bevel Point, Pencil Point with Introducer guide Needle). Sterilized Shelflife- 5 years	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year	13.03.2027	Both	Less than 3 years.
45	77	Water Seal Bag	Under Water seal Drainage System	Water Seal bag of 1 litre Capacity	Bag capacity 1000 ml Shelflife- 5 years Sterilized, Made up of SS Polyethylene (PE) + EVA.Q	Each Piece	Not mentioned	MD	14.03.2022 (Less than 3 year	13.03.2027	Both	Less than 3 years.

BAM
28/10/24

29/10/24

01/11/24

TENDER NO. BMSIC/MEDICAL DEVICES/CONSUMABLE/24-06

Check list - IV: Evaluation sheet to be used for preliminary evaluation of technical bids

Company Name: - M/s Romsons Group Pvt. Ltd. Total Number of Pages Submitted in bid documents:- 348

Sheet for verification of licence details

Sl. No	NIT Sl. No	Name of the MEDICAL DEVICES/CONSUMABLE as per NIT	Manufacturer		Importer		Validity		GMP/ WHO GMP/ COPP (Import)/BIS/ISI/QMS	
			Forms Number	Manufacturing license Number	Forms Number	Import license Number	From	To	From	To
1	5	ADK Drain Sizes 20"	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
2	6	ADK Drain Sizes 24"	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
3	7	ADK Drain Sizes 28"	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
4	8	ADK Drain Sizes 32"	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
5	20	Blood Transfusion Set with Filter 170 Micron	MD- 5	MFG/MD/2021/000151	NA	NA	01.06.2021	31.05.2026	QMS 14.06.2024	25.04.2025
6	23	Central Venous Cathetar Central Line- Monoluminal, Adult use	MD- 9	MFG/MD/2018/000104	NA	NA	10.01.2019	09.01.2024	ISO 13485:2016 11.05.2023	27.11.2025

B/M
28/10/24H/P
29/10/24E/S
04/11/24

7	24	Central Venous Cathetar Central Line-Biluminal, Adult use	MD- 9	MFG/MD/2018/000104	NA	NA	10.01.2019	09.01.2024	ISO 13485:2016 11.05.2023	27.11.2025
8	25	Central Venous Cathetar Central Line-Triluminal, Adult Use	MD- 9	MFG/MD/2018/000104	NA	NA	10.01.2019	09.01.2024	ISO 13485:2016 11.05.2023	27.11.2025
9	26	Chest Tube No.- 24	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	14.06.2024	13.06.2025
10	27	Chest Tube No.- 28	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	14.06.2024	13.06.2025
11	28	Chest Tube No.- 32	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	14.06.2024	13.06.2025
12	29	Cord Clamp	MD- 5	MFG/MD/2022/000283	NA	NA	17.06.2022	16.06.2027	ISO 13485:2016 11.05.2023	27.11.2025
13	33	Disposable Mask	MD- 5	MFG/MD/2023/000229	NA	NA	21.03.2023	20.03.2028	ISO 13485:2016 11.05.2023	27.11.2025
14	35	Epidural Set with catheter & fitter	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
15	36	Folley's Cathetar - Assorted Sizes 10	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025

Signature
28/10/24

Signature
29/10/24

Signature
04/11/24

16	37	Folley's Cathetar - Assorted Sizes 12	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
17	38	Folley's Cathetar - Assorted Sizes 14	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
18	39	Folley's Cathetar - Assorted Sizes 16	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
19	40	Folley's Cathetar - Assorted Sizes 18	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
20	41	Folley's Cathetar - Assorted Sizes 20	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
21	42	Folley's Cathetar - Assorted Sizes 6	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
22	43	Folley's Cathetar - Assorted Sizes 8	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
23	44	Infant Feeding Tube Assorted Sizes 5	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
24	45	Infant feeding Tube Assorted Sizes 6	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
25	46	Infant feeding Tube Assorted Sizes 8	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025

DM
28/10/24

Alch
29/10/24

Spares
04/11/24

26	47	IV Canula No. 18	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
27	48	IV Canula No. 20	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
28	49	IV Canula No. 22	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
29	50	IV Canula No. 24	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
30	51	IV Canula No. 26	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
31	52	IV set - Adult	MD- 5	MFG/MD/221/000151	NA	NA	01.06.2021	31.03.2026	QMS 26.04.2024	25.04.2025
32	55	PM Line	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
33	63	Rhyles Tube - Assorted Sizes 10	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
34	64	Rhyles Tube - Assorted Sizes 12	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
35	65	Rhyles Tube - Assorted Sizes 14	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025

QAM
28/10/24

Alch
29/10/24

Exane
04/11/24

36	66	Rhyles Tube - Assorted Sizes 16	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
37	67	Rhyles Tube - Assorted Sizes 18	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
38	69	Rhyles Tube - Assorted Sizes 8	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
39	70	Spinal Needle - Assorted Sizes 18	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
40	71	Spinal Needle - Assorted Sizes 20	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
41	72	Spinal Needle - Assorted Sizes 22	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
42	73	Spinal Needle - Assorted Sizes 23	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
43	74	Spinal Needle - Assorted Sizes 25	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
44	75	Spinal Needle - Assorted Sizes 26	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025
45	77	Water Seal Bag	MD- 5	MFG/MD/2022/000125	NA	NA	14.03.2022	13.03.2027	QMS 14.06.2024	13.06.2025

Note:- Assisted in technical evaluation in reference to letter no. BMSIC/40030/311-2024/4013 dt 08.08.2024 on the basis of documents provided by BMSICL as check list II, III & IV. Provided checklist compiled with due diligence and care. In spite, some inadvertent discrepancies could have been crept in. Humble request to all concerned to bring to the notice in due time if any discrepancies are observed for rectification.

[Signature]
28/10/24

[Signature]
29/10/24

[Signature]
31/11/24