

**BIDDING DOCUMENTS  
FOR  
Procurement and Rate Contracting of Medical  
Equipments for Government Medical Colleges in Bihar**

**Bid Reference: BMSICL/2013-14 / MC-005**

**Bihar Medical Services And Infrastructure Corporation Limited  
5<sup>th</sup> Floor, Biscomaun Bhavan,  
Gandhi Maidan, Patna (Bihar) India**

**Bihar Medical Services and Infrastructure Corporation, Limited, Patna.  
5<sup>th</sup> Floor Biscomaun Bhavan  
Gandhi Maidan, Patna (Bihar) India**

Telephones: 0612-2219634

Fax: —————

e-mail:

**INVITATION FOR BIDS  
FOR  
PROCUREMENT OF MEDICAL EQUIPMENTS  
*[Modify***

*as appropriate to indicate general description of items under procurement]*

BID REFERENCE	: BMSICL/2013-14/MC-005
DATE OF COMMENCEMENT OF SALE OF BIDDING DOCUMENT	: 10 <sup>th</sup> Oct 2013 from 10:00 AM
LAST DATE FOR SALE OF BIDDING DOCUMENT	: 31 <sup>st</sup> Oct 2013 till 5:00 PM
LAST DATE AND TIME FOR RECEIPT OF BIDS	: 01 <sup>st</sup> Nov 2013 till 2:00 PM
TIME AND DATE OF OPENING	: 01 <sup>st</sup> Nov 2013 at 3:00 PM
PLACE OF OPENING OF BIDS	: Bihar Medical Services & Infrastructure Corporation Limited, 5 <sup>th</sup> Floor, Bisomaun Bhavan, Gandhi Maidan, Patna 800001. Bihar
ADDRESS FOR COMMUNICATION	: Bihar Medical Services & Infrastructure Corporation Limited, 5 <sup>th</sup> Floor, Bisomaun Bhavan, Gandhi Maidan, Patna 800001. Bihar

# CONTENTS OF BIDDING DOCUMENT

<b>Invitation for Bids .....</b>	<b>4</b>
<b>Section I. Instructions to Bidders .....</b>	<b>7</b>
Table of Clauses.....	8
<b>Section II. General Conditions of Contract.....</b>	<b>23</b>
Table of Clauses.....	24
<b>Section III. Special Conditions of Contract.....</b>	<b>37</b>
<b>Section IV. Schedule of Requirements .....</b>	<b>39</b>
Consignee List .....	41
<b>Section V. Technical Specifications .....</b>	<b>42</b>
<b>Section VI. Sample Forms .....</b>	<b>68</b>
Notes to Bidders on the Preparation of Sample Forms.....	69
Bid Forms .....	70
<b>Section VII. Check List.....</b>	<b>83</b>

**INVITATION FOR BIDS**  
**(IFB)**

**INVITATION FOR BIDS (IFB)**  
**FOR**  
**SUPPLY, TESTING, DEMONSTRATION, INSTALLATION & COMMISSIONING**  
**OF MEDICAL EQUIPMENT AT GOVT. MEDICAL COLLEGES IN BIHAR**

**Managing Director,**  
**Bihar Medical Services And Infrastructure Corporation Limited**  
**5<sup>th</sup> Floor, Biscomaun Bhavan, Gandhi Maidan,**  
**Patna-800001 (Bihar)**

Bid Reference No.: BMSICL/2013-14/MC-005

Date: 10<sup>th</sup> Oct 2013

1. The Bihar Medical Services and Infrastructure Corporation Limited, Patna (name of purchaser) on behalf of Governor of Bihar, invites sealed bids from manufacturers or their authorized dealer / distributor / sole selling agent (having authorization in the format (Form-6) given in the bidding document) for Supply, testing, Demonstration, Installation and Commissioning of Medical Equipment and related services as listed below:-

<b>Schedule No.</b>	<b>Brief Description of Goods and Services</b>	<b>Qty./No.</b>	<b>Delivery Schedule (Days)</b>	<b>Earnest Money Deposit (EMD) in Indian Rupees</b>
1	Pediatric Ventilator	10	30	3,00,000/-
2	Neonatal Ventilator	8	30	2,40,000/-
3	Portable X Ray	2	30	20,000/-
4	Digital Mammography	2	30	8,00,000/-
5	C Arm Image Intensifier	2	30	60,000/-
6	Portable Ultrasound	1	30	20,000/-
11	Portable USG with color Doppler	1	30	40,000/-
8	Echocardiography with pediatric and neonatal probe	2	30	72,000/-
9	Mortuary Chamber	2	30	24,000/-
10	EEG	1	30	10,000/-
11	EMG	2	30	20,000/-
12	Evoked Potential	1	30	10,000/-

2. Interested bidders may obtain further information from and inspect the bidding documents at the office of the Managing Director, Bihar Medical Services and Infrastructure Corporation Limited, 5<sup>th</sup> Floor, Biscomaun Bhavan, Gandhi Maidan, Patna-800001, Bihar.
3. The Bidding Document may be purchased from the office of the Managing Director, Bihar Medical Services and Infrastructure Corporation Limited, 5<sup>th</sup> Floor, Biscomaun Bhavan, Gandhi Maidan, Patna-800001, Bihar (Name and address of Purchaser), from 10<sup>th</sup> Oct 2013 to 31<sup>st</sup> Oct 2013 during office hours, from 10:00 hrs to 17:00 hrs on all working days either in person or by post.
4. A complete set of bidding documents may be purchased by interested bidders upon submission of a written application to the address given in para 2 and upon payment of a nonrefundable fee of Rs. 10,000/- in the form of a cash or Demand Draft or banker's Cheque in favor of Managing Director, Bihar Medical Services and Infrastructure Corporation Limited. The tender document can also be

downloaded from the website: [www.bmsicl.gov.in](http://www.bmsicl.gov.in) Such bidders are required to submit non-refundable tender document cost in the form of Demand draft in favour of Managing Director, Bihar Medical Services and Infrastructure Corporation Limited.

5. The bidding documents requested by mail will be dispatched by registered post / speed post / courier service on payment of an extra amount of Rs. 500/-. The Purchaser will not be responsible for postal delay, if any, in the delivery of the bidding documents or of the non-receipt of the same
6. Bidders are free to quote for any or all of the items listed in the schedule of requirements and the evaluation of bids will be conducted on per – item basis. The bidder must quote at least for the full quantity of one schedule.
7. The bids must be submitted/delivered at the address given in para 2 on or before 11.00 hrs. on 01<sup>st</sup> Nov 2013. All bids must be accompanied by an Earnest Money Deposit (EMD) as specified in the bidding document. Late bids will be rejected.
8. The Pre-bid meeting shall be organized at the purchaser's office given at para 2 on 21<sup>st</sup> October 2013 at 01:00 PM for **Schedule no 1 to 12**. In the Pre-bid meeting, the prospective bidders may clarify any issues related to the terms, conditions and technical specifications given in the bidding documents.
9. Bids will be opened in the presence of bidder's representatives who chose to attend at Bihar Medical Services & Infrastructure Corporation Ltd., 5<sup>th</sup> Floor Biscomaun Bhavan on 01<sup>st</sup> November 2013 at 03:00 PM.
10. The Purchaser reserves the right to cancel / annul the bidding process without assigning any reason thereof.
11. In the event of the date specified for the bid receipt and opening being declared as a closed holiday for purchaser's office, the due date for submission of bids and opening of bids will be the following working day at the appointed time.

**(Managing Director)**  
**Bihar Medical Services and Infrastructure Corporation**  
\*\*\*

**INSTRUCTION TO BIDDERS (ITB)**

## TABLE OF CLAUSES

<b>A. Introduction</b> .....	<b>9</b>
1. Scope of Bid.....	9
2. Fraud and Corruption.....	9
3. Eligible Bidders.....	9
4. One Bid per Bidder.....	9
5. Cost of Bidding.....	10
<b>B. The Bidding Documents</b> .....	<b>10</b>
6. Content of Bidding Documents.....	10
7. Clarification of Bidding Documents.....	10
8. Pre-bid Meeting.....	10
9. Amendment of Bidding Documents.....	11
<b>C. Preparation of Bids</b> .....	<b>11</b>
10. Language of Bid.....	11
11. Documents Constituting the Bid.....	11
12. Bid Form.....	11
13. Bid Prices.....	11
14. Documents to be submitted by the Bidder.....	12
15. Documents establishing Bidder's Qualification.....	13
16. Documents establishing Goods Conformity to Bidding Documents.....	14
17. Earnest Money Deposit (EMD).....	14
18. Period of Validity of Bids.....	15
19. Format and Signing of Bid.....	16
<b>D. Submission of Bids</b> .....	<b>16</b>
20. Sealing and Marking of Bids.....	16
21. Deadline for Submission of Bids.....	17
22. Late Bids.....	17
23. Modification and Withdrawal of Bids.....	17
<b>E. Bid Opening and Evaluation</b> .....	<b>17</b>
24. Opening of Bids by Purchaser.....	18
25. Clarification of Bids.....	18
26. Preliminary Evaluation.....	19
27. Evaluation and Comparison of Substantially Responsive Bids.....	20
28. Contacting the Purchaser.....	21
<b>F. Award of Contract</b> .....	<b>21</b>
29. Post-qualification.....	21
30. Award Criteria.....	21
31. Purchaser's Right to vary Quantities.....	21
32. Purchaser's Right to Accept Any Bid and to Reject Any or All Bids.....	21
33. Issue of Notification of Award.....	21
34. Signing of Contract.....	22
35. Performance Security.....	22

## INSTRUCTIONS TO BIDDERS

### A INTRODUCTION

---

#### 1. SCOPE OF BID

*Bihar Medical Services and Infrastructure Corporation Limited [name of purchaser]* on behalf of Governor of Bihar (hereinafter referred to as ‘Purchaser’), invites bids for the supply/testing/installation /commissioning of Medical Equipments as specified in the Schedule of Requirements.

#### 2. FRAUD AND CORRUPTION

2.1 It is required that the Purchasers as well as Bidders/Suppliers/Contractors observe the highest standard of ethics during the procurement and execution of Contracts. In pursuance of this policy, the Purchaser:

(a) defines, for the purposes of this provision, the terms set forth below as follows:

(i) “corrupt practice” means the offering, giving, receiving, or soliciting of any thing of value to influence the action of a public official in the procurement process or in Contract execution; and

(ii) “fraudulent practice” means a misrepresentation of facts and / or concealment of fact in order to influence a procurement process or the execution of a Contract to the detriment of the Purchaser; it includes collusive practices among bidders (prior to or after bid submission) designed to establish bid prices at artificial, noncompetitive levels and to deprive the Purchaser of the benefits of free and open competition.

(b) will declare a firm ineligible and debar the firm, either indefinitely or for a stated period of time, to be awarded a Contract if it at any time determines that the firm has engaged in corrupt or fraudulent practices in competing for, or in executing, a contract. In such cases, appropriate legal action as per court of law shall be initiated for which the concerned bidder shall be solely responsible.

2.2 Furthermore, bidders shall be aware of the provision stated in Sub-Clauses 19.4 and 22.1 d. of the General Conditions of Contract

#### 3. ELIGIBLE BIDDERS

3.1 The eligible bidder should be registered with appropriate authorities in India to manufacture / supply the tendered item, against Technical Specifications given in the bid document and should have successfully executed orders of similar nature in past. In case of imported goods, the Indian agent / bidder should be duly authorized by the manufacturer of Goods in the format given in the bidding document.

3.2 A firm declared ineligible by the Purchaser in accordance with ITB Sub-Clause 2.1 (b) and GCC Sub-Clause 19.4 shall be ineligible to bid for a contract during the period of time determined by the Purchaser.

3.3 Pursuant to ITB Sub-Clause 11, the Bidder shall furnish, as part of its bid, documents establishing, to the Purchaser’s satisfaction, the Bidder’s eligibility to bid.

**4. ONE BID PER BIDDER**

A firm shall submit only one bid either individually or as a partner of a joint venture. A firm that submits either individually or, as a member of a joint venture, more than one bid will cause all the proposals with the firm's participation to be disqualified.

**5. COST OF BIDDING**

The bidder shall bear all costs associated with the preparation and submission of the bid. The Purchaser will, in no case, be responsible or liable for these costs, regardless of the conduct or outcome of the bidding process.

**B. THE BIDDING DOCUMENTS**

---

**6. CONTENTS OF BIDDING DOCUMENTS**

6.1 The goods required to be supplied; bidding procedures and contract terms and conditions are prescribed in the Bidding Documents. The Bidding Document include, the following :

- Section I Instructions to Bidders (ITB)
- Section II General Conditions of Contract (GCC)
- Section III Special Conditions of Contract (SCC)
- Section IV Schedule of Requirements (SOR)
- Section V Technical Specifications
- Section VI Sample Forms

6.2 The "Invitation for Bids" does not form part of the Bidding Documents and is included as a reference only. In case of discrepancies between the Invitation for Bid and the Bidding Documents listed in 6.1 above, said Bidding Documents will take precedence.

6.3 The Bidder is expected to examine all instructions, forms, terms and specifications in the Bid Documents. Failure to furnish all information required as per the Bid Documents or submission of the bids not substantially responsive to the Bid Documents in every respect will be at the bidder's risk and may result in rejection of the bid.

**7. CLARIFICATION OF BID DOCUMENTS**

7.1 A prospective bidder, requiring any clarification on the Bid Documents shall notify the Purchaser in writing or by FAX/e-mail at the Purchaser's mailing address indicated in the invitation of Bid. The Purchaser shall respond in writing to any request for the clarification of the Bid Documents, which it receives not later than 10 days prior to the date of opening of the Tenders. Copies of the query (without identifying the source) and clarifications by the Purchaser shall be sent to all the prospective bidders who have received the bid documents.

7.2 Any clarification issued by the Purchaser in response to query raised by prospective bidders shall form an integral part of bid documents and it may amount to an amendment of relevant clauses of the bid documents.

**8. Pre-bid Meeting**

8.1 The bidder or his representative is invited to attend a pre-bid meeting, which will take place in the office on 21<sup>st</sup> October 2013 at 13.00 hrs. for **Schedule no 1 to 12**.

8.2 The purpose of the meeting will be to clarify issues and to answer questions on any matter that may be raised at that stage.

- 8.3 The bidder may submit any question in writing or by FAX/ e-mail to reach the purchaser not later than one week before the pre-bid meeting.
- 8.4 The Minutes of the pre-bid meeting, including the text of the questions raised and the responses given will be transmitted without delay to all purchasers of the bidding documents. Any modification of the bidding document listed in ITB Clause 6.1 which may become necessary as a result of the pre-bid meeting shall be made exclusively through the issue of an Addendum pursuant to ITB Clause 9 and not through the minutes of the pre-bid meeting.
- 8.5 Non-attendance at the pre-bid meeting will not be a cause for disqualification of a bidder.

## **9. AMENDMENT OF BIDDING DOCUMENTS**

- 9.1 At any time, prior to the date of submission of Bids, the Purchaser may, for any reason, whether at its own initiative or in response to a clarification requested by a prospective bidder, or pursuant to ITB Clause 8, modify bid documents by amendments.
- 9.2 The amendments shall be notified in writing or by FAX to all prospective bidders on the address intimated at the time of purchase of the bid document from the purchaser and these amendments will be binding on them.
- 9.3 In order to afford prospective bidders a reasonable time to take the amendment into account in preparing their bids, the purchaser may, at its discretion, extend the deadline for the submission of bids suitably.

## **C. PREPARATION OF BIDS**

### **10. LANGUAGE OF BID**

The bid, as well as all correspondence and documents relating to the bid exchanged by the Bidder and the Purchaser, shall be written in English language. However, the purchaser as well as bidder may correspond in Hindi language also.

### **11. DOCUMENTS CONSTITUTING THE BID**

The bid prepared by the bidder shall comprise the following components:

- (a) a Bid Form and a Price Schedule completed in accordance with ITB Clauses 12 and 13;
- (b) documentary evidence established in accordance with ITB Clause 14 and 15 that the Bidder is eligible and qualified to perform the contract if its bid is accepted;
- (c) documentary evidence established in accordance with ITB Clause 16 that the goods and ancillary services to be supplied by the Bidder conform to the bidding documents; and
- (d) Earnest Money Deposit (EMD) furnished in accordance with ITB Clause 17.
- (e) Tender Document fee in the form of Demand Draft in favour of Managing Director, Bihar Medical services and Infrastructure Corporation Ltd. Payable at Patna or Money receipt of Tender Document cost if purchased by hand.

## **12. BID FORM**

The bidder shall complete the Bid Form and appropriate Price Schedule furnished in the Bidding Documents, indicating the goods to be supplied, brief description of the goods, quantity and prices as per section VI.

## **13. BID PRICES**

The bidder shall give the total composite price inclusive of all Levies & Taxes i.e. Sales / Trade Tax & Excise, packing, forwarding, freight, octroi/entry tax and insurance etc. The basic unit price and all other components of the price need to be individually indicated against the goods it proposes to supply under the contract as per the price schedule given in Section VI. Prices of incidental services should also be quoted. The offer shall be quoted in Indian Rupees. No Foreign exchange will be made available by the purchaser.

13.2 Break-up of the prices indicated in the Price Schedule shall be entered in the following manner:

- (i) The Basic Unit price (Ex-Factory Price) of the goods, Excise duty, Sales Tax, Freight, octroi/entry tax Forwarding, Packing, Insurance and any other Levies/Charges already paid or payable by the supplier shall be quoted separately item wise.
- (ii) The supplier shall quote as per price schedule given in section VI for all the items given in schedule of requirement.

13.3 The price quoted by the bidder shall remain fixed during the entire period of contract and shall not be subject to variation on any account. A bid submitted with an adjustable price quotation will be treated as non - responsive and rejected.

13.4 The prices quoted by the bidder shall be in sufficient detail to enable the Purchaser to arrive at the price of equipment/system offered.

13.5 "DISCOUNT, if any, offered by the bidders shall not be considered unless specifically indicated in the price schedule. Bidders desiring to offer discount shall therefore modify their offers suitably while quoting and shall quote clearly net price taking all such factors like Discount, free supply, etc, into account".

13.6 The price approved by the Purchaser for procurement will be FOR destination which will be inclusive of all Taxes, Levies, packing, forwarding, freight and insurance as mentioned in Para 13.1 above. Breakup in various heads like excise duty, sales / trade tax, insurance, freight and other taxes paid/payable as per clause 13.2 (i) is for the information of the purchaser and any change in these shall have no effect on price during the scheduled delivery period.

## **14. DOCUMENTS REQUIRED TO BE SUBMITTED**

14.1 The bidder shall furnish, as part of the bid documents, the following documents or whichever is applicable as per terms and conditions of Bidding Documents.

- (i) Certificate of incorporation / registration.
- (ii) Article or Memorandum of Association or partnership deed as the case may be.
- (iii) Registration certificate from State Director of Industries.
- (iv) Registration certificate from central excise and trade/sales tax department.
- (v) Approval from Reserve Bank of India in case of foreign collaboration.

- (vi) In case of bidder, other than manufacturer, the manufacturer's authorization certificate in the format given in the bidding document.
- (vii) Non-conviction certificate / an affidavit duly notarized.
- 14.2 (i) The bidder shall furnish Balance Sheet for last 3 financial years as evidence that he has financial capability to perform the contract.
- (ii) The bidder shall furnish documentary evidence about technical and production capability necessary to perform the contract.
- 14.3 In order to enable the Purchaser to assess the proven ness of the system offered, the bidder shall provide documentary evidence regarding the system being offered by him.
- 14.4 The offered product may be required to be type approved / demonstrated at the Purchaser's office as a part of technical evaluation of bids. For this purpose, the supplier shall submit a sample for type evaluation. The sample would be evaluated for its ability to meet the technical specifications, manufacturability, reliability, testability, ease of installation, maintainability etc. Necessary documents to substantiate these attributes will have to be submitted at the time of application for approval by the supplier for obtaining type approval.

Or

In case, it is not possible to get / accord type approval, the bidder has to make necessary arrangements for inspection at the place where the equipment is installed and functioning or at the manufacturer's premises.

Or

In case goods offered have already been type approved/ validated by the Purchaser, documentary evidence to this effect shall be submitted by the bidder.

## **15. DOCUMENTS ESTABLISHING BIDDER'S QUALIFICATION**

- 15.1 Pursuant to ITB Clause 11, the bidder shall furnish, as part of its bid, documents establishing the Bidder's qualification to perform the Contract if its bid is accepted.
- 15.2 The documentary evidence of the Bidder's qualifications to perform the Contract shall establish to the Purchaser's satisfaction that:
  - a) The bidder should be a manufacturer who must have manufactured, tested and supplied the equipment(s) similar to the type specified in the 'Schedule of Requirements' up to at least 80% of the quantity required in any one of the last 3 years and should be in satisfactory operation for 6 months as on date of bid opening.
  - b) Bids of bidders quoting as authorized representative of a manufacturer, meeting with the above requirement in full, can also be considered provided:
    - (i) The manufacturer furnishes authorization

- (ii) in the prescribed format given at Section VI, assuring full guarantee and warranty obligations as per GCC Clause 14 for the equipment offered; and
  - (iii) The bidder, as authorized agent has supplied/installed/commissioned and provided after sales services satisfactorily at least 80% of the quantity specified in the Schedule of Requirements in any one of the last 3 years which must be in satisfactory operation for at least 6 months on the date of bid opening.
- c) The bidder should furnish the information on past supplies and satisfactory performance for both 15.2 (a) and (b) above, in the proforma given under Section VI, Form No. 7.
  - d) Bidders shall invariably furnish documentary evidence in support of the satisfactory operation of the equipment (issued from the end user) as specified above.
  - e) The bidder should clearly confirm that all the facilities exist in his factory for inspection and testing and these will be made available to the Purchaser or his representative for inspection.
  - f) The Bidder shall furnish data to support that he has the financial and production capacity to perform the contract and complete the supplies within the stipulated delivery period.
  - g) The bidder should furnish profit and loss statement, balance sheets and auditor's report for the past three years, banker's certificates, etc. in support of its financial standing.
- 15.3 If an agent submits bid in behalf of more than one manufacturer unless each such bid is accompanied by a separate bid form for each bid and bid securities, when required for each bid and authorization from the respective Manufacturer, all such bids will be rejected as non-responsive

## **16. DOCUMENTS ESTABLISHING GOODS CONFORMITY TO BIDDING DOCUMENTS**

- 16.1 Pursuant to ITB Clause 11, the Bidder shall furnish, as part of its bid, documents establishing the conformity to the bidding documents of all goods and services which the Bidder proposes to supply under the contract.
- 16.2 The documentary evidence of conformity of the goods and services to the bidding documents may be in the form of literature, drawings and data, and shall consist of :
- (a) a detailed description of the essential technical and performance characteristics of the goods ;
  - (b) a list giving full particulars, including available sources and current prices, of spare parts, special tools, etc., necessary for the proper and continuing functioning of the goods for a period of three years, following commencement of the use of the goods by the Purchaser; and
  - (c) an item-by-item commentary on the Purchaser's Technical Specifications demonstrating substantial responsiveness of the goods and services to those specifications or a statement of deviations and exceptions to the provisions of the Technical Specifications.
- 16.3 For purposes of the commentary to be furnished pursuant to ITB Clause 16.2 (c) above, the Bidder shall note that standards for workmanship, material and equipment, and references to brand names or catalogue numbers designated by the Purchaser in its Technical Specifications are intended to be descriptive only and not restrictive. The Bidder may substitute alternative standards, brand names and/or catalogue numbers in its bid, provided that it demonstrates to the Purchaser's satisfaction that the substitutions ensure substantial equivalence to those designated in the Technical Specifications.

## **17. EARNEST MONEY DEPOSIT (EMD)**

- 17.1 Pursuant to ITB Clause 11, the bidder shall furnish, as part of his bid, the Earnest Money Deposit (EMD) for an amount of mentioned in Section IV – Schedule of Requirements in the form of Demand Draft or Bank Guarantee.
- 17.2 The Earnest Money Deposit (EMD) is required to protect the purchaser against the risk of bidder's conduct, which would warrant the forfeiture of Earnest Money Deposit (EMD) pursuant to ITB Clause 17.7. No exemption from payment of EMD is permitted except those who are registered with the Central Purchase Organistaion/State Purchase Organistaion, National Small Scale Industries Corporation (NSIC) or the concerned Department. The small scale units located in Bihar shall not be liable to deposit earnest money.
- 17.3 The Earnest Money Deposit (EMD) shall be in the form of a Bank Draft drawn in favor of Purchaser or FDR or Bank Guarantee issued by a scheduled bank in favour of the purchaser, valid for a period of 45 days beyond the validity of Bid.
- (i) The bank guarantee for Earnest Money Deposit (EMD) or NSIC certificate for claiming exemption from submission of bank guarantee against Earnest Money Deposit (EMD), as prescribed in ITB Clause 17.1 of Section I of the bid document shall be submitted along with the bids in a separate cover. The bank guarantee so submitted shall be as per the format given in Section VI on prescribed judicial paper with stamps of proper value and should contain full address of the issuing branch of the bank with its telephone number and FAX number. This cover should be superscribed as **“EARNEST MONEY DEPOSIT (EMD) FOR TENDER No BMSICL/2013-14/MC- 005 issued on 10<sup>th</sup> October 2013.**
- (ii) In case where the document of Earnest Money Deposit (EMD) is not submitted in the manner prescribed under clause 2 (i) above, cover containing the commercial, technical and financial offers **SHALL NOT BE OPENED AND THE BID SHALL BE REJECTED AND RETURNED TO THE BIDDER UNOPENED.**
- 17.4 A bid not secured in accordance with para 17.1, and 17.3 shall be rejected by the Purchaser being non-responsive at the bid opening stage and returned to the bidder unopened.
- 17.5 The Earnest Money Deposit (EMD) of the unsuccessful bidder will be discharged/returned as promptly as possible, but not later than 30 days after the expiry of the period of the bid validity prescribed by the purchaser pursuant to ITB Clause 18.
- 17.6 The successful bidder's Earnest Money Deposit (EMD) will be discharged upon the bidder's acceptance of the advance purchase order satisfactorily in accordance with GCC Clause 5 and furnishing the performance security.
- 17.7 The Earnest Money Deposit (EMD) may be forfeited :
- (a) If the bidder withdraws his bid during the period of bid validity specified by the bidder in the Bid form or
- (b) In the case of successful bidder , if the bidder fails :
- (i) to sign the contract in accordance with ITB Clause 29 or
- (ii) to furnish performance security in accordance with ITB Clause 30.

## **18. PERIOD OF VALIDITY OF BIDS**

- 18.1 Bid shall remain valid for **150 days** from the date of opening of bids prescribed by the purchaser pursuant to ITB Clause 24.1. A bid valid for a shorter period shall be rejected by the purchaser being non-responsive.
- 18.2 In exceptional circumstances, the purchaser may request the consent of the bidder for an extension to the period of bid validity. The request and the response thereto shall be made in writing. The Earnest Money Deposit (EMD) provided under ITB Clause 17 shall also be suitably extended. The bidder may refuse the request without forfeiting his Earnest Money Deposit (EMD). A bidder accepting the request and granting extension will not be permitted to modify his bid.

## **19. FORMAT AND SIGNING OF BID**

- 19.1 (i) The bidder shall prepare single stage two part bids, i.e. (a) Technical bid (un-priced) in duplicate and (b) Price Bid in duplicate clearly marking them as 'ORIGINAL' and 'COPY' and in addition shall enclose Earnest Money Deposit (EMD) in a single separate envelope. In the event of any discrepancy between the copy bid, the original shall govern.
- (ii) The copy of quality manual and Article or Memorandum of Association may be provided in the original bid only.
- 19.2 The original and copy of Bid shall be typed or printed and all the pages numbered consecutively and shall be signed by the bidder or a person or persons duly authorized to bind the bidder to the contract. The letter of authorization shall be indicated by written power-of-attorney accompanying the bid. All pages of the original bid, except for un-amended printed literatures, shall be signed by the person or persons signing the bid. The bids submitted shall be sealed properly.
- 19.3 The bid shall contain no interlineations, erasures or overwriting except as necessary to correct errors made by the bidder in which case such corrections shall be signed by the person or persons signing the bid.

## **D. SUBMISSION OF BIDS**

---

### **20. SEALING AND MARKING OF BIDS**

- 20.1 The bidder shall seal the original and copy bids in separate envelopes duly marking the envelopes, separately as

Cover 'A'

- i. Technical Bid (original)
- ii. Technical Bid (copy)
- iii. Earnest Money Deposit (EMD)

Cover 'B'

- i. Price Bid (original)
- ii. Price Bid (copy)

All the envelopes mentioned above should be enclosed in another sealed outer envelope duly marked by the personal seal of the bidder.

- 20.2 (a) The envelopes shall be addressed to the purchaser at the following address :

Bihar Medical Services And Infrastructure Corporation Limited

5<sup>th</sup> Floor Biscomaun Bhavan, Gandhi Maidan,  
Patna- 800001. Bihar.

- (b) The envelope shall bear (the name and address of the Purchaser), the tender number and the words 'DO NOT OPEN BEFORE' (due date & time).
- (c) The inner and outer envelopes shall indicate the name and address of the bidders to enable the bid to be return unopened in case it is declared 'late' or rejected.
- (d) Bids may be sent by registered post or delivered in person on above mentioned address (address is given in Clause 20.2 (a) above). The responsibility for ensuring that the bids are delivered in time would vest with the bidder.
- (e) Bids delivered in person on the day of bid opening shall be delivered up to 27<sup>th</sup> Jun 2012 by 11.00 Hrs to Bihar Medical Services & Infrastructure Corporation Ltd., 5<sup>th</sup> Floor, Biscomaun Bhavan, Gandhi Maidan, Patna. The purchaser shall not be responsible if the bids are delivered elsewhere.
- (f) Venue of bid opening: Bids will be opened at BMSICL, Patna, at 13.000 Hrs. on the due date. If due to administrative reason, the venue of Bid opening is changed, it will be displayed prominently on the notice board of the Purchaser's office.

20.2 If both the envelopes are not sealed and marked as required at ITB Clause 20.1 and 20.2 , the bid shall be rejected.

## **21. DEADLINE FOR SUBMISSION OF BIDS**

- 21.1 Bids must be received by the Purchaser at the address and up to the due date and time specified under ITB Clause 20.2.
- 21.2 The Purchaser may, at its discretion, extend this deadline for the submission of bids by amending the Bid Documents in accordance with clause 6 in which case all rights and obligations of the purchaser and bidders previously subject to the deadline will thereafter be subjected to the deadline as extended.

## **22. LATE BIDS**

Any bid received by the purchaser after the deadline for submission of bids prescribed by the purchaser pursuant to clause 21, shall be rejected and returned unopened to the bidder.

## **23. MODIFICATION AND WITHDRAWAL OF BIDS**

- 23.1 No bid may be modified subsequent to the deadline for submission of bids. The bidder may modify or withdraw its bid after submission, provided that written notice of the modification or withdrawal is received by the purchaser prior to the deadline prescribed for submission of bids along with a written power of attorney authorizing the signatory of the withdrawal.
- 23.2 The bidder's modification or withdrawal notice shall be prepared, sealed, marked and dispatched as required in the case of bid submission in accordance with the provision of ITB Clause 20. A withdrawal notice may also be sent by FAX/ e-mail but followed by a signed confirmation copy by post not later than the deadline for submission of bids.
- 23.3 Bids requested to be withdrawn in accordance with ITB Clause 23.1 above, shall be returned unopened to the Bidders.

- 23.4 No bid may be withdrawn in the interval between the bid submission deadline and the expiration of the bid validity period specified in ITB Clause 18. Withdrawal of a bid during this interval may result in the forfeiture of the Bidder's Earnest Money Deposit (EMD), pursuant to ITB Clause 17.7

## **E. BID OPENING AND EVALUATION**

---

### **24. OPENING OF BIDS BY PURCHASER**

- 24.1 The purchaser shall open the technical bids in the presence of bidders or their authorized representatives who chose to attend, at the due date and time of bid opening. The bidder's representatives, who are present, shall sign in an attendance register. Authority letter to this effect shall be submitted by the bidders before they are allowed to participate in bid opening (A Format is given in Section VI).
- 24.2 A maximum of two representatives of any bidder shall be authorized and permitted to attend the bid opening.
- 24.3 The bidder's names, modifications, bid withdrawals, requisite Earnest Money Deposit (EMD) and such other details as the purchaser, at its discretion, may consider appropriate will be announced at the time of opening. No bid shall be rejected at the time of bid opening, except for late bids which shall be returned unopened to the bidder pursuant to ITB clause 22.
- 24.4 The price bids of bidders whose Technical bids are found technically responsive and comply with the bid documents will only be opened at a later date. The date of opening of financial bids shall be communicated to such bidders, whose Technical bids are found technically responsive. The bidder's representative may be present at the time of opening of price bid at the pre-appointed time, date and venue.
- 24.5 The date fixed for opening of bids, if subsequently declared as holiday by the Government, the revised date of schedule will be notified. However, in absence of such notification, the bids will be opened on next working day, time and venue remaining unaltered.

### **25. CLARIFICATION OF BIDS**

To assist in the examination, evaluation and comparison of bids, the purchaser may, at its discretion ask the bidder for the clarification of its bid. The request for the clarification and the response shall be in writing. Unless the purchaser asks for change in price due to clarifications sought, the bidder is not permitted to alter the price furnished in Price Bid "**Cover B**". The change in price shall be submitted in a separately sealed covers with marking in the cover "**Supplemental Price Bid**" before opening of the "**Original Price Bid**"

### **26. PRELIMINARY EVALUATION**

- 26.1 Purchaser shall evaluate the bids to determine whether they are complete, whether any computational errors have been made, whether required sureties have been furnished, whether the documents have been properly signed and whether the bids are generally in order. Bids from representatives, without proper Authorization from the manufacturer as per Section VI, shall be treated as non-responsive
- 26.2 Arithmetical errors shall be rectified on the following basis. If there is a discrepancy between the unit price and total price that is obtained by multiplying the unit price and quantity, the unit price shall prevail and the total price shall be corrected by the purchaser. If there is a discrepancy

between words and figures, the amount in words shall prevail. If the supplier does not accept the correction of the errors, his bid shall be rejected.

- 26.3 Prior to the detailed evaluation pursuant to ITB Clause 27, the Purchaser will determine the substantial responsiveness of each bid to the Bid Document. For purposes of these clauses, a substantially responsive bid is one which confirms to all the terms and conditions of the Bid Documents without material deviations. Deviations from or objections or reservations to critical provisions such as those concerning Performance Security (GCC clause 5) , Warranty (GCC clause 14), Force Majeure (GCC clause 21), Applicable Law (GCC clause 28) and Taxes and duties (GCC clause 30) along with deviation in Technical Specifications will be deemed as material deviation. The purchaser's determination of bid's responsiveness shall be based on the contents of the bid itself without recourse to extrinsic evidence.
- 26.4 A bid, determined as substantially non-responsive will be rejected by the purchaser and shall not subsequent to the bid opening be made responsive by the bidder by correction of the non-conformity.
- 26.5 The Purchaser may waive any minor infirmity or non-conformity or irregularity in a bid which doesn't constitute a material deviation, provided such waiver doesn't prejudice or affect the relative ranking of any bidder.

## **27. EVALUATION AND COMPARISON OF SUBSTANTIALLY RESPONSIVE BIDS**

- 27.1 The Purchaser shall evaluate in detail and compare the bids previously determined to be substantially responsive pursuant to ITB Clause 26.
- 27.2 The purchasers evaluation of bid will take into account, in addition to the bid price (ex-factory/ex-warehouse/off-the-shelf price of goods offered from India, such price to include all costs as well as duties and taxes paid or payable on components and raw materials incorporated or to be incorporated in the goods, and excise duty on finished goods if payable) and price of incidental services, the following factors, in the manner and to the extent indicated in ITB clause 27.3 and in the Technical Specifications:
- (a) i) cost of inland transportation, insurance and other costs within India incidental to the delivery of goods to their final destination;
    - ii) the Comprehensive Annual Maintenance Charges for a period of 7 years subsequent to free guarantee maintenance period of 3 years
  - (b) delivery schedule offered in the bid;
  - (c) deviations in payment schedule from that specified in the Special Conditions of Contract.
  - (d) The availability in India of spare parts and after sales services for the equipment offered in the bid.
- 27.3 Pursuant to ITB clause 27.2 the following evaluation methods will be applied:
- (a) Inland transportation, ex-factory/ from port-of-entry, insurance and incidentals.
    - (i) Inland transportation, insurance and other incidentals, for delivery of goods to the Project site as stated in ITB clause 13.2. These costs will be added to bid price.
  - (b) Delivery schedule:

The **Purchaser** desires to have delivery of the goods covered under the invitation, at the time specified in the schedule of requirements. The estimated time of the arrival of the goods at the project site should be calculated for each bid after allowing for reasonable transportation time. Treating the bid offering the scheduled time of arrival as the base, a delivery "adjustment" will be calculated for other bids at 2% of the exfactory price for each month of delay beyond the base and this will be added to the bid price for evaluation. No credit will be given to earlier

deliveries and bids offering delivery beyond 2 months of stipulated delivery will be treated as unresponsive.

(c) Deviation in Payment Schedule:

The General Conditions of Contract clause 15 indicate the payment schedule offered by the **Purchaser**. If a bid deviates from the schedule and if such deviation is considered acceptable to the **Purchaser**, the bid will be evaluated by calculating interest earned for any earlier payments involved in the terms outlined in the bid as compared to those stipulated in this invitation at a rate of 12% per annum.

(d) Spare parts and after sales service facilities in India:

The cost of the **Purchaser** of establishing the minimum service facilities and parts inventories, as outlined elsewhere in the bid invitation, if quoted separately, shall be added to the bid price.

(e) Annual Maintenance Contract (AMC):

(i) The Purchaser desires to have **separately** comprehensive maintenance charges for a period of 7 years after the expiry of free maintenance period, clearly indicating year wise comprehensive maintenance charges, which shall be added to the bid price at a discount rate of 8% per annum. **Bids without this charge will be considered as non responsive.**

(ii) Any major repair pointed out by the **Purchaser** shall be rectified by the Supplier from the date of intimation within a period of 3 calendar days and commission the equipment to the satisfaction of the Purchaser, failing which the purchaser has write to levy a penalty on the Supplier a sum of Rs.\_2,500/- per day or part thereof for each equipment until the equipments are repaired and commission to the satisfaction of the Purchaser.

(f) Spares:

(i) The supplier shall be required to provide a list and rates of spare parts recommended for maintenance for three years after the end of Guarantee period of three years. The purchaser may elect to purchase the recommended spares from the supplier at any time including at the end of warranty/ AMC, provided that such purchase shall not relieve the supplier from any warranty/ AMC obligations under the contract.

(ii) The cost of spares shall be discounted @ 15% over warranty/ AMC period (if there is a provision for AMC in the contract) to arrive at the final price of the equipment for the purpose of tender evaluation.

(iii) Over a period of three years starting from the date of final acceptance of the equipment or after the procurement of spares, supplier shall supply at his own cost, spare parts needed which have not been included in the offer. These spares should be supplied within a maximum period of thirty days from the notification by the purchaser of his need, without demur.

(iv) In the event of termination of production of the equipment/ spare parts, the supplier shall notify the purchaser at least two years in advance of the impending termination to enable the purchaser to procure life time spares. The supplier shall also provide at his own cost to the purchaser, the blue print drawings and specifications of spare parts if and when

(g) Repair of faulty equipment and setting up of Repair Facilities:

(i) The supplier shall establish adequate repair facilities for repair of faulty equipment in India within a period six months from the date of purchase order. The number and location of repair facilities should be such as to meet the requirement of repairs and turn around time provided in the special conditions in Section IV. The performance bank guarantee shall not be released until the purchaser is satisfied that sufficient repair facilities have been

established in addition to the fulfillment of other conditions of the contract. The purchaser reserves the right to blacklist a supplier who does not meet the repair obligation as per the conditions of contract.

## **28. CONTACTING THE PURCHASER**

- 28.1 Subject to ITB Clause 25, no bidder shall try to influence the Purchaser on any matter relating to its bid, from the time of the bid opening till the time the contract is awarded.
- 28.2 Any effort by a bidder to modify his bid or influence the purchaser in the purchaser's bid evaluation, bid comparison or contract award decision shall result in the rejection of the bid.

## **F AWARD OF CONTRACT**

---

### **29. POST-QUALIFICATION**

- 29.1 The Purchaser will determine to its satisfaction whether the Bidder that is selected as having submitted the lowest evaluated responsive bid is qualified to perform the Contract satisfactorily, in accordance with the criteria listed in ITB Sub-Clause 15 & 16.
- 29.2 The determination will evaluate the Bidder's financial, technical, and production capabilities. It will be based on an examination of the documentary evidence of the Bidder's qualifications submitted by the Bidder, pursuant to ITB Sub-Clause 15 & 16, and the information submitted by the Bidder in the 'Proforma For Performance Statement' for the period of last 5 years given in Section VI as well as other information the Purchaser deems necessary and appropriate.
- 29.3 An affirmative post-qualification determination will be a prerequisite for award of the contract to the lowest evaluated Bidder. A negative determination will result in rejection of the Bidder's bid, in which event the Purchaser will proceed to the next-lowest evaluated Bidder to make a similar determination of that Bidder's capabilities to perform satisfactorily.

### **30. AWARD CRITERIA**

Subject to ITB Clause 32, the Purchaser shall award the Contract to the Bidder whose bid has been determined to be substantially responsive and has been determined to be the lowest evaluated bid and whose goods have been type approved/validated by the purchaser.

### **31. PURCHASER'S RIGHT TO VARY QUANTITIES**

The Purchaser reserves the right at the time of Contract award or within the stipulated last date of delivery, to increase or decrease, by 25%, the quantity of goods and services beyond that originally specified in the Schedule of Requirements without any change in unit price or other terms and conditions.

### **32. PURCHASER'S RIGHT TO ACCEPT ANY BID AND TO REJECT ANY OR ALL BIDS**

The Purchaser reserves the right to accept or reject any bid, and to annul the bidding process and reject all bids, at any time prior to award of contract without assigning any reason whatsoever and without thereby incurring any liability to the affected bidder or bidders on the grounds of purchaser's action.

**33. ISSUE OF NOTIFICATION OF AWARD**

- 33.1 The issue of Notification of Award shall constitute the intention of the Purchaser to enter into contract with the bidder.
- 33.2 Prior to the expiration of the period of bid validity, the Purchaser will notify the successful Bidder in writing by registered letter or by cable, to be subsequently confirmed in writing by registered letter, that its bid has been accepted
- 33.3 The bidder shall within 7 days of issue of the Notification of Award, give his acceptance along with performance security in conformity with Section VI provided with the bid document.

**34. SIGNING OF CONTRACT**

- 34.1 The issue of Notification of Award shall constitute the award of contract on the bidder.
- 34.2 Promptly after the Purchaser notifies the successful Bidder that its bid has been accepted, the Purchaser will send the Bidder the Contract Form provided in the Bidding Documents, incorporating all agreements between the parties.
- 34.3 Within seven (7) days of receipt of the Contract Form, the successful Bidder shall sign and date the Contract Form and return it to the Purchaser

**35. PERFORMANCE SECURITY**

- 35.1 Within seven (7) days of the receipt of notification of award from the Purchaser, the successful Bidder shall furnish the performance security in accordance with the Conditions of Contract, using the Performance Security Form provided in the Bidding Documents, or in another form acceptable to the Purchaser.
- 35.2 Failure of the successful Bidder to comply with the requirement of ITB Clause 34 and ITB Clause 35.1 shall constitute sufficient grounds for the annulment of the award and forfeiture of the Earnest Money Deposit (EMD), in which event the Purchaser may make the award to the next-lowest evaluated bid submitted by a qualified Bidder or call for new bids.
- 35.3 Payment: 95% payment will be made by BMSICL within 30 days against supply, installation and commissioning of the equipment at the respective Institution and upon submission of satisfactory completion of installation certificate signed by the hospital authorities. The balance 5% payment will be released on submission of Bank Guarantee for the said amount with a validity to cover up the warranty period. In case, Bank Guarantee is not submitted by the “Supplier”, the said amount will be released after expiry of warranty period of three years.

**SECTION II- GENERAL CONDITIONS OF CONTRACT**

## Table of Clauses

1.	Definitions and application.....	25
2.	Standards.....	25
3.	Use of Contract Documents and Information; Inspection and Audit by the Purchaser.....	26
4.	Patent Rights.....	26
5.	Performance Security.....	26
6.	Inspections and Tests.....	26
7.	Packing.....	27
8.	Delivery and Documents.....	28
9.	Training.....	28
10.	Incidental Services.....	29
11.	Spares.....	29
12.	Insurance.....	29
13.	Transportation.....	29
14.	Warranty.....	30
15.	Payment Terms.....	30
16.	Prices.....	31
17.	Change Orders.....	31
18.	Subcontracts.....	31
19.	Delays in the Supplier's Performance.....	31
20.	Liquidated Damages.....	32
21.	Force Majeure.....	32
22.	Termination for Default.....	33
23.	Termination for Insolvency.....	33
24.	Termination for Convenience.....	33
25.	Settlement of Disputes.....	34
26.	Limitation of Liability.....	35
27.	Governing Language.....	35
28.	Applicable Law.....	35
29.	Notices.....	36
30.	Taxes and Duties.....	36

## SECTION III

### GENERAL CONDITIONS OF CONTRACT

#### 1. DEFINITIONS

In this Contract, the following terms shall be interpreted as indicated:

- (a) **“The Purchaser”** means the Bihar Medical Services and Infrastructure Corporation Limited (BMSICL), the organization purchasing the Goods.
- (b) **“The Bidder”** means the individual or firm who participates in the tender and submits its bid.
- (c) **“Days”** means calendar days.
- (d) **“GCC”** means Conditions of Contract.
- (e) **“The Supplier”** means the individual or firm supplying the goods and Services under the contract.
- (f) **“The Goods”** means all equipment, machinery, and/or other materials which the Supplier is required to supply to the Purchaser under the contract.
- (g) **“Services”** means services ancillary to the supply of the Goods, such as transportation and insurance, and any other incidental services, such as installation, commissioning, provision of technical assistance, training and other obligations of the Supplier covered under the Contract.
- (h) **“End User”** means the consignees stated in the Schedule of Requirements.
- (i) **“The Notification of Award”** means the intention of the Purchaser to place the Purchase order on the bidder or to enter in to contract with the bidder.
- (j) **“The Contract”** means the agreement entered into between the Purchaser and the Supplier, as recorded in the Contract Form signed by the parties, including all the attachments and the appendices thereto and all documents incorporated by reference therein.
- (k) **“The Contract Price”** means the price payable to the Supplier under the contract for the full and proper performance of its contractual obligations.
- (l) **“Validation”** is a process of testing the equipment as per the specifications including requirements for use in hospital is carried out in simulated field environment.

1.1 **Application:** The General Conditions shall apply to the extent that they are not superseded by provisions in other parts of the contract.

#### 2. STANDARDS

The goods supplied under this contract shall conform to the standards prescribed in the Technical Specifications mentioned in section VI and when no applicable standard is mentioned, to the authoritative standard appropriate to the Goods Country or origin and such standards shall be latest issued by concerned Institution.

### **3. USE OF CONTRACT DOCUMENTS AND INFORMATION; INSPECTION AND AUDIT BY THE PURCHASER**

- 3.1** The Supplier shall not, without the Purchaser's prior written consent, disclose the Contract, or any provision thereof, or any specification, plan, drawing, pattern, sample, or information furnished by or on behalf of the Purchaser in connection therewith, to any person other than a person employed by the Supplier in the performance of the Contract. Disclosure to any such employed person shall be made in confidence and shall extend only so far as may be necessary for purposes of such performance.
- 3.2** The Supplier shall not, without the Purchaser's prior written consent, make use of any document except for purposes of performing the Contract.
- 3.3** Any document, other than the Contract itself, enumerated in GCC Sub-Clause 3.1 shall remain the property of the Purchaser and shall be returned (all copies) to the Purchaser on completion of the Supplier's performance under the Contract if so required by the Purchaser.
- 3.4** The Supplier shall permit the Purchaser to inspect the Supplier's accounts and records relating to the performance of the Contract and to have them audited by auditors appointed by the Purchaser, if so required.

### **4. PATENT RIGHTS**

The supplier shall indemnify the purchaser against all third-party claims of infringement of patent, trademark or industrial design rights arising from use of the goods or any part thereof in India.

### **5. PERFORMANCE SECURITY**

- 5.1** The supplier shall furnish performance security to the purchaser for an amount equal to 5% of the value of purchase order within 7 **days** from the date of issue of Notification of Award by the Purchaser.
- 5.2** The proceeds of the performance security shall be payable to the Purchaser as compensation for any loss resulting from the supplier's failure to complete its obligations under the contract.
- 5.3** The performance security denominate in Indian Rupees shall be in the form of Bank Guarantee issued by a Scheduled / Nationalized Bank and in the form provided in 'Section VI' of this Bid Document or in the form of cashiers cheque, certified cheque or demand draft.. The performance security should be valid for the period beyond sixty (60) days following the date of completion of the Supplier's performance obligations under the Contract, including any warranty obligations
- 5.4** The performance security will be discharged by the Purchaser and returned to the Supplier not later than thirty (30) days following the date of completion of the Supplier's performance obligations under the Contract, including any warranty obligations.

### **6. INSPECTION AND TESTS**

- 6.1** The Purchaser or his representative shall have the right to inspect and test the goods as per prescribed test schedules for their conformity to the specifications. Where the Purchaser decides to conduct such tests on the premises of the supplier or its subcontractor(s), all reasonable facilities and assistance like Testing instruments and other test gadgets including access to drawings and production data shall be furnished to the inspectors at no charge to the purchaser. The supply will

be accepted only after quality assurance tests are carried out by the Purchaser as per prescribed schedule and material passing the test successfully.

- 6.2 Should any inspected or tested goods fail to conform to the specifications the purchaser may reject them and the supplier shall either replace the rejected goods or make all alterations necessary to meet Specification requirements free of cost to the purchaser.
- 6.3 Notwithstanding the pre-supply tests and inspections prescribed in GCC Clause 6.1 & 6.2 above, the equipment and accessories on receipt in the Purchaser's premises will also be tested during and after installation before "take over" and if any equipment or part thereof is found defective, the same shall be replaced free of all cost to the purchaser as laid down in GCC Clause 6.4 below.
- 6.4 If any equipment or any part thereof, before it is taken over under GCC Clause 6.5, is found defective or fails to fulfill the requirements of the contract, the inspector shall give the Supplier notice setting forth details of such defects or failure and the supplier shall make the defective equipment good, or alter the same to make it comply with the requirements of the contract forthwith and in any case within a period not exceeding three months of the initial report. These replacements shall be made by the supplier free of all charges at site. Should it fail to do so within this time, the purchaser reserves the discretion to reject and replace at the cost of the supplier the whole or any portion of equipment as the case may be, which is defective or fails to fulfill the requirements of the contract? The cost of any such replacement made by the purchaser shall be deducted from the amount payable to the supplier.
- 6.5 When the performance tests called for have been successfully carried out, the inspector / ultimate consignee will forthwith issue a Taking Over Certificate. The inspector /ultimate consignee shall not delay the issue of any "taking Over Certificate" contemplated by this clause on account of minor defects in the equipment which do not materially affect the commercial use thereof provided that the supplier shall undertake to make good the same in a time period not exceeding two months. The Taking Over Certificate shall be issued by the ultimate consignee within six weeks of successful completion of tests. In this case, a Consignee Receipt Certificate issued by the consignee as per the Format given in Section VI shall be equivalent to "Taking Over Certificate", issuance of which shall certify receipt of goods in safe and sound condition. However, they shall not discharge the supplier of their warranty obligation. The Consignee Receipt Certificate in respect of last consignment against the Contract will be equivalent to "Taking Over Certificate".
- 6.6 Nothing in GCC Clause 6 shall in any way release the Supplier from any warranty or other obligations under this contract.

## 7.1 **PACKING**

The Supplier shall provide such packing of the Goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in the Contract. The packing shall be sufficient to withstand, without limitation, rough handling during transit and exposure to extreme temperatures, salt, and precipitation during transit and open storage. Packing case size and weights shall take into consideration, where appropriate, the remoteness of the Goods' final destination and the absence of heavy handling facilities at all points in transit.

- 7.2 The packing, marking and documentation within and outside the packages shall comply strictly with such special requirements as shall be provided for in the Contract including additional requirements, if any, specified in SCC and in any subsequent instructions ordered by the purchaser.
- 7.3 Packing Instruction: The supplier will be required to mark separate packages for each consignee. Each package will be marked on three sides with proper paint/indelible ink, the following:
  - i. Purchaser:
  - ii. Contract No.

- iii. Supplier Name
- iv. Packing List reference Number

## **8. DELIVERY AND DOCUMENTS**

8.1 Upon or before delivery of the Goods, the Supplier shall notify the Purchaser in writing and deliver the following documents to the Purchaser:

- (i) two originals and two copies of the Supplier's invoice, showing Purchaser, the Contract number, Goods' description, quantity, unit price, and total amount. Invoices must be signed in original and stamped or sealed with the company stamp/seal;
- (ii) two copies of delivery note, railway consignment note, road consignment note, truck or air waybill, or multi-modal transport document showing Purchaser as Bihar Medical Services and Infrastructure Corporation Limited [ *enter correct name of Purchaser for excise purposes* ] and delivery through to final destination as stated in the Contract;
- (iii) copy of the Insurance Certificate, showing the Purchaser as the beneficiary;
- (iv) three copies of the packing list identifying contents of each package;
- (v) one original of the manufacturer's or Supplier's Warranty certificate covering all items supplied;
- (vi) original copy of the Certificate of Inspection furnished to Supplier by the nominated inspection agency;
- (vii) other procurement-specific documents required for delivery/payment purposes.

The above documents shall be received by the Purchaser before arrival of the Goods (except where it is handed over to the Consignee with all documents) if not received, the Supplier will be responsible for any consequent expenses.

**Note:** *In the event that the documents presented by the Supplier are not in accordance with the Contract, then payment will be made against issue of the 'Consignee Receipt Certificate', to be issued in accordance with GCC Clause 6 above*

8.2 The delivery of the goods and documents shall be completed within 3 months from the date of issue of Notification of Award. First month is for lead period and evenly distributed supplies are expected in remaining two months. The actual delivery schedule will be given in Notification of Award.

8.3 All Technical assistance for installation, commissioning and monitoring of the equipment shall be provided by the Supplier at no extra cost during laboratory evaluation, validation/ type approval and field trial, if any.

*[Hint: Generally three months delivery time is envisaged. The delivery period will be decided on case-to-case basis considering specific requirement. The delivery period for procurement will be two months for store items where no trial run and installation & commissioning is required.]*

## **9. TRAINING**

9.1 The bidder shall demonstrate and provide training on use and maintenance of the Equipments to the consignee's personnel the purchaser free of cost where required.

9.2 The bidder shall specify in his bid the number of trainees, quantum of proposed training, pre-training qualifications required of the trainees and duration of the proposed training.

9.3 The bidder shall provide all training material and documents.

9.4 Conduct of training of the purchaser's personnel may be at the supplier's plant and/or on-site in assembly start-up operation, maintenance and/or repair of the supplied goods.

## **10. INCIDENTAL SERVICES**

10.1 The supplier may be required to provide any or all of the following services:

- (a) Performance or supervision of on-site assembly and/or start-up of the supplied Goods;
- (b) Furnishing of tools required for assembly and/or maintenance of supplied Goods;
- (c) Performance of supervision or maintenance and/or repair of the supplied Goods, for a period of time agreed by the parties provided that this service shall not relieve the supplier of any warranty obligations under this contract.
- (d) Furnish detailed operations and maintenance manual for each appropriate unit of supplied goods.

## **11. SPARES**

11.1 The supplier shall be required to provide a list of the following material and notifications pertaining to spare parts manufactured or distributed by the supplier of spares including cost and quantity considered for arriving at the price of spares in ITB Clause 9.

- (a) Such spare parts as the purchaser may elect to purchase from the supplier provided that such purchase shall not relieve the supplier of any warranty obligation under the contract.
- (b) In the event of termination of production of the spare parts, the supplier shall :
  - i) give advance notification to the purchaser pending termination (not less than 2 years), in sufficient time to enable the purchaser to procure life time spare; and
  - ii) following such advance intimation of termination, furnish at no cost to the purchaser, the blue prints, drawings and specifications of spare parts, if and when requested.

11.2 Over a period of three years starting from the date of final acceptance, the supplier shall supply, at his own cost, all necessary spares which have not been included in the offer as part of the requirement. These spares should be supplied within a maximum period of 30 days from the notification by the purchaser of his need.

## **12. INSURANCE**

12.1 The Goods supplied under the Contract shall be insured in an amount equal to 110% of the EXW value of the Goods from "warehouse to warehouse" on "all risks" basis including war risks and strikes.

## **13. TRANSPORTATION**

Where the Supplier is required under the Contract to transport the Goods to a specified place of destination, defined in Consignee list, transport to such place of destination, including insurance and storage, as shall be specified in the Contract, shall be arranged by the Supplier, and related costs shall be included in the Contract Price.

## **14. WARRANTY**

- 14.1 The supplier shall warrant that the goods to be supplied shall be new and free from all defects and faults in materials used, workmanship and manufacture and shall be of the highest grade and consistent with the established and generally accepted standards for materials of the type ordered and shall perform in full conformity with the specifications and drawings. The supplier shall be responsible for any defect that may develop under the conditions provided by the contract and under proper use, arising from faulty material, design or workmanship such as corrosion of the equipment, inadequate quantity of material to meet equipment requirements, inadequate contact protection, deficiencies in circuit design and/or otherwise and shall remedy such defects at his own cost when called upon to do so by the Purchaser who shall state in writing in what respect the stores are faulty. This warranty shall survive inspection or payment for / and acceptance of goods, but shall expire (except in respect of complaints notified prior to such date) three years after the goods have been taken over under GCC Clause 6.5 above.
- 14.2 This warranty shall remain valid for three years after the goods or any portion thereof as the case may be, have been delivered to and accepted at the final destination indicated in the contract.
- 14.3 If it becomes necessary for the Supplier to replace or renew any defective portion(s) of the equipment under this clause, the provisions of the GCC Clause 14.1 shall apply to the portion(s) of the equipment so replaced or renewed or until the end of the above mentioned period of three years, whichever may be later. If any defect is not remedied by the supplier within a reasonable time, the Purchaser may proceed to get the defects remedied from other supplier etc., at the supplier's risk and expenses, but without prejudice to any other rights which the purchaser may have against the supplier in respect of such defects.
- 14.4 Replacement under warranty clause shall be made by the supplier free of all charges at site including freight, insurance and other incidental charges.

## **15. PAYMENT TERMS**

- 15.1 The method and conditions of payment to be made to the supplier under the contract shall be specified in the Special Conditions of Contract.
- 15.2 The Supplier's request(s) for payment shall be made to the Purchaser in writing, accompanied by an invoice describing, as appropriate, the Goods delivered and Services performed, and by documents submitted pursuant to GCC Clause 8, and upon fulfillment of other obligations stipulated in the Contract.
- 15.3 Payments shall be made promptly by the Purchaser, but in no case later than sixty (60) days after submission of an invoice or claim by the Supplier.

*[Hint: The actual payment conditions for new products or procurements having installation and AMC services may be decided on case to case basis and incorporated in special conditions of the contract]*

- 15.4 (i) Form C and also a certificate stating that the tendered item (stores) are meant for the use of Govt. Hospital shall be provided by the purchaser on the request of the bidder as and when asked for.
- (ii) No payment will be made for goods rejected at the site on testing.
- 15.5 Payment for goods shall be made in Indian Rupees as follows:
- a) No advance payment is payable.

- b) **Payment:** That the 95% payment will be made by BMSICL within 30 days against supply, installation and commissioning of the equipment at the respective Institution and upon submission of satisfactory completion of installation certificate signed by the hospital authorities. The balance 5% payment will be released on submission of Bank Guarantee for the said amount with a validity to cover up the warranty period. In case, Bank Guarantee is not submitted by the “Supplier”, the said amount will be released after expiry of warranty period of three years.

## **16. PRICES**

- 16.1 (i) (a) Prices charged by the supplier for goods delivered and services performed under the contract shall not be higher than the prices quoted by the Supplier in his Bid.
  - (b) In the case of revision of Statutory Levies/Taxes during the finalization period of tender, the Purchaser reserves the right to ask for reduction in the prices.
- (ii) (a) Prices once fixed will remain valid during the schedule delivery period. Increase and decrease of Taxes and other statutory duties will not affect the price during this period.
  - (b) Any increase in taxes and other statutory duties/levies after the expiry of the delivery date shall be to the supplier’s account. However benefit of any decrease in these taxes/duties shall be passed on to the Purchaser by the supplier.

## **17. CHANGES ORDERS**

- 17.1 The purchaser may, at any time, by a written order given to a supplier, make changes within the general scope of the contract in any one or more of the following:
  - (a) drawings, designs or specifications, where Goods to be supplied under the contract are to be specifically manufactured for the Purchaser;
  - (b) the method of transportation or packing;
  - (c) the place of delivery; or
  - (d) the services to be provided by the supplier.
- 17.2 If any such change causes an increase or decrease in the cost of, or the time required for the execution of the contract an equitable adjustment shall be made in the contract price or delivery schedule, or both, and the contract shall accordingly be amended. Any proposal by the supplier for adjustment under this clause must be made within thirty days from the date of the receipt of the change in order.

## **18. SUBCONTRACTS**

The Supplier shall notify the Purchaser in writing of all subcontracts awarded under this contract if not already specified in his bid. Such notification, in his original bid or later shall not relieve the supplier from any liability or obligation under the Contract.

## **19. DELAYS IN THE SUPPLIER’S PERFORMANCE**

- 19.1 Delivery of the Goods and performance of the services shall be made by the Supplier in accordance with the time schedule specified by the purchaser in its purchase order. In case the supply is not completed in the stipulated delivery period, as indicated in the Purchase Order, purchaser reserves the right either to short close/cancel this purchase order and/or recover liquidated damage charges. The cancellation/short closing of the order shall be at the risk and responsibility of the supplier and purchaser reserves the right to purchase balance unsupplied item at the risk and cost of the defaulting vendors.
- 19.2 Delay by the Supplier in the performance of its delivery obligations shall render the Supplier liable to any or all of the following sanctions: forfeiture of its performance security, imposition of liquidated damages and/or termination of the contract for default.
- 19.3 If at any time during the performance of the contract, the supplier encounters condition impending timely delivery of the goods and performance of service, the Supplier shall promptly notify to the Purchaser in writing the fact of the delay, its likely duration and its cause(s). As soon as practicable after receipt of the supplier's notice, the Purchaser shall evaluate the situation and may at its discretion extend the period for performance of the contract (by not more than 20 weeks) subject to furnishing of additional performance security by the supplier @ 5% of the total value of the Purchase Order.

*[Hint: Each case of delivery extension shall have to be examined a fresh vis-à-vis the prevailing market prices]*

- 19.4 If supplier fails to perform its contractual obligations, pursuant to GCC Clause 19.3 above, the purchaser may consider debaring the firm for the period of 1-5 years for participation in future invitation of bids. The period of debar, as stated above, shall be at the sole discretion of the Purchaser

## **20 LIQUIDATED DAMAGES**

- 20.1 The date of delivery of the goods stipulated in the acceptance of the tender should be deemed to be the essence of the contract and delivery must be completed, not later than the dates specified therein. Extension will not be given except in exceptional circumstances. Should, however, deliveries be made after expiry of the contracted delivery period, without prior concurrence of the purchaser and be accepted by the consignee, such delivery will not deprive the purchaser of his right to recover liquidated damage under GCC Clause 20.2 below.
- 20.2 Should the supplier fails to deliver the store or any consignment thereof within the period prescribed for delivery, the purchaser shall be entitled to recover 0.5 % of the value of the delayed supply for each week of delay or part thereof for a period up to 20 (Twenty) weeks. In the case of package supply where the delayed portion of the supply materially hampers installation and commissioning of the systems, L/D charges shall be levied as above on the total value of the concerned package of the Purchase Order. Quantum of liquidated damages assessed and levied by the purchaser shall be final and not challengeable by the supplier. However, when supply is made within 21 days of QA clearance in the extended delivery period, the consignee may accept the stores and in such cases the LD shall be levied upto the date of QA clearance.

## **21. FORCE MAJEURE**

- 21.1 If, at any time, during the continuance of this contract, the performance in whole or in part by either party of any obligation under this contract is prevented or delayed by reasons of any war or hostility, acts of the public enemy, civil commotion, sabotage, fires, floods, explosions, epidemics, quarantine restrictions, strikes, lockouts or act of God (hereinafter referred to as events) provided notice of happenings of any such eventuality is given by either party to the other within 21 days

from the date of occurrence thereof, neither party shall by reason of such event be entitled to terminate this contract nor shall either party have any claim for damages against other in respect of such non-performance or delay in performance, and deliveries under the contract shall be resumed as soon as practicable after such an event come to an end or cease to exist, and the decision of the Purchaser as to whether the deliveries have been so resumed or not shall be final and conclusive. Further that if the performance in whole or part of any obligation under this contract is prevented or delayed by reasons of any such event for a period exceeding 60 days, either party may, at its option, terminate the contract.

21.2 Provided, also that if the contract is terminated under this clause, the Purchaser shall be at liberty to take over from the Supplier at a price to be fixed by the purchaser, which shall be final, all unused, undamaged and acceptable materials, bought out components and stores in course of manufacture which may be in possession of the Supplier at the time of such termination or such portion thereof as the purchaser may deem fit, except such materials, bought out components and stores as the Supplier may with the concurrence of the purchaser elect to retain.

## **22. TERMINATION FOR DEFAULT**

22.1 The Purchaser may, without prejudice to any other remedy for breach of contract, by written notice of default, sent to the supplier, terminate this contract in whole or in part

- a) if the supplier fails to deliver any or all of the goods within the time period(s) specified in the contract, or any extension thereof granted by the purchaser pursuant to GCC Clause 19;
- b) if the supplier fails to perform any other obligation(s) under the Contract; and
- c) if the supplier, in either of the above circumstances, does not remedy his failure within a period of 15 days (or such longer period as the purchaser may authorize in writing) after receipt of the default notice from the purchaser.
- d) If the Supplier, in the judgment of the Purchaser, has engaged in corrupt and fraudulent practices in competing for executing the Contract, pursuant to ITB Clause 2.

22.2 In the event the purchaser terminates the contract in whole or in part pursuant to GCC Clause 22.1 the purchaser may procure, upon such terms and in such manner as it deems appropriate, goods similar to those undelivered and the supplier shall be liable to the Purchaser for any excess cost for such similar goods. However the supplier shall continue the performance of the contract to the extent not terminated.

22.3 In the event, any sums found due to the Purchaser / Government under or by virtue of the fulfillment of contractual obligations, these shall be recoverable from the Supplier and his / its properties, movable and immovable, under the provisions of the Revenue Recovery Act, for the time being in force as tough as they are arrears of land revenue or in any manner and within such time as the Purchaser / Government may deem fit. Any sum of money due and payable to the Supplier from Government / Purchaser may be adjusted against sum of money due to the Supplier under any other contract.

## **23. TERMINATION FOR INSOLVENCY**

The Purchaser may at any time terminate the Contract by giving written notice to the Supplier, without compensation to the supplier. If the supplier becomes bankrupt or otherwise insolvent as declared by the competent court provided that such termination will not prejudice or affect any right of action or remedy which has accrued or will accrue thereafter to the purchaser.

## **24. TERMINATION FOR CONVENIENCE**

- 24.1 The Purchaser, by written notice sent to the Supplier, may terminate the Contract, in whole or in part, at any time for its convenience. The notice of termination shall specify that termination is for the Purchaser's convenience, the extent to which performance of the Supplier under the Contract is terminated, and the date upon which such termination becomes effective.
- 24.2 The Goods that are complete and ready for shipment within thirty (30) days after the Supplier's receipt of notice of termination shall be accepted by the Purchaser at the Contract terms and prices. For the remaining Goods, the Purchaser may elect:
- (a) to have any portion completed and delivered at the Contract terms and prices; and/or
  - (b) to cancel the remainder and pay to the Supplier an agreed amount for partially completed Goods and Services and for materials and parts previously procured by the Supplier.

## **25. SETTLEMENT OF DISPUTES**

- 25.1 If any dispute or difference of any kind whatsoever shall arise between the Purchaser and the Supplier in connection with or arising out of the Contract, the parties shall make every effort to resolve amicably such dispute or difference by mutual consultation.
- 25.2 If, after thirty (30) days, the parties have failed to resolve their dispute or difference by such mutual consultation, then either the Purchaser or the Supplier may give notice to the other party of its intention to commence arbitration, as hereinafter provided, as to the matter in dispute, and no arbitration in respect of this matter may be commenced unless such notice is given.
- 25.2.1 Any dispute or difference in respect of which a notice of intention to commence arbitration has been given in accordance with this Clause shall be finally settled by arbitration. Arbitration may be commenced prior to or after delivery of the Goods under the Contract.
- 25.2.2 The dispute resolution mechanism to be applied shall be as follows:
- (a) In case of Dispute or difference arising between the Purchaser and a domestic supplier relating to any matter arising out of or connected with this agreement, such disputes or difference shall be settled in accordance with the Arbitration and Conciliation Act, 1996. The arbitral tribunal shall consist of 3 arbitrators one each to be appointed by the Purchaser and the Supplier. The third Arbitrator shall be chosen by the two Arbitrators so appointed by the Parties and shall act as Presiding arbitrator. In case of failure of the two arbitrators appointed by the parties to reach upon a consensus within a period of 30 days from the appointment of the arbitrator appointed subsequently, the Presiding Arbitrator shall be appointed by the Medical Council of India.
  - (b) Where the value of the contract is Rs.1 crore and below, the disputes or differences arising shall be referred to the Sole Arbitrator. The Sole Arbitrator should be appointed by agreement between the parties; failing such agreement, by the Medical Council of India.
  - (c) In case of Dispute with a foreign supplier, the dispute shall be settled in accordance with provision of UNCITRAL (United Nations Commission on International Trade Law) Arbitration Rules. The Arbitral Tribunal shall consist of 3 Arbitrators one each to be appointed by the Purchaser and the Supplier. The third Arbitrator shall be chosen by the two Arbitrators so appointed by the Parties and shall act as presiding arbitrator. In case of failure of the two arbitrators appointed by the parties to reach upon a consensus within a period of 30

days from the appointment of the arbitrator appointed subsequently, the Presiding Arbitrator shall be appointed by the Medical Council of India.

- (d) If one of the parties fails to appoint its arbitrator in pursuance of sub-clause (a) and (c) above, within 30 days after receipt of the notice of the appointment of its arbitrator by the other party, then the Medical Council of India, both in cases of the Foreign supplier as well as Indian supplier, shall appoint the arbitrator. A certified copy of the order of the Medical Council of India making such an appointment shall be furnished to each of the parties.
- (e) The venue of Arbitration shall be the place from where the contract is issued i.e Patna, and the language of the arbitration proceedings and that of all councils and communications between the parties shall be English.
- (f) The decision of the majority of arbitrators shall be final and binding upon parties. The cost and expenses of Arbitration proceedings will be paid as determined by the arbitral tribunal. However, the expenses incurred by each party in connection with the preparation, presentation, etc. of its proceedings as also the fees and expenses paid to the arbitrator appointed by such party or on its behalf shall be borne by each party itself.
- (g) The Arbitration and Conciliation Act of 1996 the rules herewith and any statutory modification or reenactment thereof shall apply to arbitration proceedings.

25.3 Notwithstanding any reference to arbitration herein,

- (a) the parties shall continue to perform their respective obligations under the Contract unless they otherwise agree; and
- (b) the Purchaser shall pay the Supplier any monies due the Supplier.

25.4 The contract shall be governed by and interpreted in accordance with the laws of India from the time being in force. All disputes arising out of this tender will be subject to jurisdiction of courts of law in Patna

## **26. LIMITATION OF LIABILITY**

26.1 Except in cases of criminal negligence or willful misconduct, and in the case of infringement pursuant to GCC Clause 4,

- (a) the Supplier shall not be liable to the Purchaser, whether in contract, or otherwise, for any indirect or consequential loss or damage, loss of use, loss of production, or loss of profits or interest costs, provided that this exclusion shall not apply to any obligation of the Supplier to pay liquidated damages to the Purchaser and
- (b) the aggregate liability of the Supplier to the Purchaser, whether under the Contract, in tort or otherwise, shall not exceed the total Contract Price, provided that this limitation shall not apply to the cost of repairing or replacing defective equipment.

## **27. GOVERNING LANGUAGE**

27.1 The Contract shall be written in English language. All correspondence and other documents pertaining to the Contract that are exchanged by the parties shall be written in the Hindi / English language.

**28. APPLICABLE LAW**

28.1 The Contract shall be interpreted in accordance with the laws of Union of India.

**29. NOTICES**

29.1 Any notice given by one party to the other pursuant to this Contract shall be sent to the other party in writing or by cable, telex, or facsimile and confirmed in writing to the other party's address.

29.2 A notice shall be effective when delivered or on the notice's effective date, whichever is later.

**30. Taxes and Duties**

30.1 The Supplier shall be entirely responsible for all taxes, duties, octroi, road permits, license fees, etc., incurred until delivery of the contracted Goods to the Purchaser.



**SECTION III- SPECIAL CONDITIONS OF CONTRACT**

## SPECIAL CONDITIONS OF CONTRACT

1. The special conditions of contract shall supplement the ‘**Instructions to the Bidders**’ as contained in Section I & “**General Conditions of the Contract**” as contained in Section II and wherever there is a conflict, the provisions herein shall prevail over those in Section I and Section II. .
2. The small scale industries registered with National Small Scale Industries Corporation(NSIC) for the tendered item under single point registration scheme and desirous of claiming concessions available to such units inclusive of Earnest Money Deposit (EMD) should submit their latest NSIC certificates and documents in respect of their monetary limit and financial capability duly certified by NSIC.
3. **Rate Contract:** The tender is also a ‘Rate Contract’. The bidders are expected to quote their best rates for the equipments. The rates quoted by the bidder shall remain valid for one year from the date of signing of contract and the bidder will have the option to extend the period of price firmness for a further period of upto six months, during which BMSICL or any of the user Institutions under the Government of Bihar, may place order for the supply and installation of same equipments procured under this tender. If the tender inviting authority/user institutions choose to place the orders for supply, installation and commissioning, the successful bidder is bound to supply the same make/model of the equipment at the same rate and same terms and conditions of this tender to such agencies/institutions, placing the repeat order. The rate contractors can withdraw at any point of time, after the minimum price firmness period of six months, but not after accepting the Letter of Intent or entering into Agreement with BMSICL or any other user Institution under the Government for the Quantity for which it has entered into Agreement with BMSICL/User Institutions during the minimum price firmness period. BMSICL/User Institutions can also withdraw from rate at any point of time after minimum price firmness periods of six months, but not after entering into Agreement with the rate contractor for the Quantity for which the Contract is already signed by both parties.

**SECTION IV- SCHEDULE OF REQUIREMENTS**

## SCHEDULE OF REQUIREMENTS

Note: Delivery Schedule expressed below is the number of days required to deliver the Equipment at Consignee Location from the date of receipt of Purchase order.

<b>Schedule No.</b>	<b>Brief Description of Goods and Services</b>	<b>Qty./No.</b>	<b>Delivery Schedule</b>	<b>Earnest Money Deposit (EMD) in Indian Rupees</b>
1	Paediatric Ventilator	10	30 days	3,00,000/-
2	Neonatal Ventilator	8	30 days	2,40,000/-
3	Portable X Ray	2	30 days	20,000/-
4	Digital Mammography	2	30 days	8,00,000/-
5	C Arm Image Intensifier	2	30 days	60,000/-
6	Portable Ultrasound	1	30 days	20,000/-
7	Portable USG with color Doppler	1	30 days	40,000/-
8	Echocardiography with Paediatric and Neonatal probes	2	30 days	72,000/-
9	Mortuary Chamber	2	30 days	24,000/-
10	EEG Machine	1	30 days	20,000/-
11	EMG Machine	2	30 days	20,000/-
12	Evoked Potential	1	30 days	10,000/-

## Consignee list

*[Table of consignee list to be inserted in the bidding document by the Purchaser to indicate the quantity of each Goods to be delivered at every consignee location.]*

Sr. no.	Equipment Name	Consignee wise Qty.					Total Qty.
		PMCH Patna	SKMCH Muzaffarpur	JLNMCH Bhagalpur	ANMMCH Gaya	IGIC Patna	
1	Paediatric Ventilator	5	2	3			10
2	Neonatal Ventilator	8					8
3	Portable X-Ray	1			1		2
4	Digital Mammography		1		1		2
5	C Arm Image Intensifier		1		1		2
6	Portable Ultrasound			1			1
7	Portable USG with Color Doppler	1					1
8	Echocardiography with Paediatric and Neonatal Probe	2					2
9	Mortuary Chamber			2			2
10	EEG Machine			1			1
11	EMG Machine			2			2
12	Evoked Potential			1			1

## **SECTION V : TECHNICAL SPECIFICATIONS**

## Paediatric Ventilator with All Connectors

1. Should be a Microprocessor based Pressure Limited.
2. Ventilator for use on Paediatric to Adult Patients.
3. Should have the following modes:
  - a. Controlled Assisted,
  - b. SIMV,
  - c. CPAP and
  - d. Advanced modes – PVC, PC, PSV, SIMV/CPAP, Flow Trigger etc.
4. Should have a REUSABLE external flow sensor which works on pressure difference (Should not be Semi-disposable or disposable)
5. Should be able to control the inspiratory flow depending on patient conditions.
6. Should have Volume Guarantee features during spontaneous breaths.
7. Should have back-up ventilation.
8. Should be able to monitor the following:
  - a. Exhaled Tidal Volume
  - b. Total Minute Volume
  - c. Total Breath Rate
  - d. Spontaneous Rate, I:E Ratio, Insp. Time
  - e. Backup percentage, mean, peak & plateau pressure
  - f. Different types of breaths etc.,
9. Should have all necessary Audio- Visual Alarms.
10. Should have:
  - a. Tidal Volume Range : 0.03 – 2.00 Ltrs
  - b. Rate : 0 – 120 BPM
  - c. Peak Flow : 5 – 150 LPM
  - d. O<sub>2</sub>% : 21 – 100%
  - e. Peep : 0 – 50 Cms H<sub>2</sub>O
11. Should have facility for inspiratory pause and expiratory hold
12. Should have built-in Nebulizer.
13. Should have built-in/External NOISE LESS compressor.
14. Should have Graphic Display to Monitor Flow
15. Volume Plus pressure Curves & Loops.
16. Unit to be supplied with UPS with a battery back up of 2 hours.
17. A low battery alarm should visibly and audibly indicate when the battery voltage falls to a level below which the unit may fail to perform satisfactory.
18. Recharging of battery within 16 hours after depletion.
19. Battery should be easy to recharge / replace.
20. Non – suppressible alarm for oxygen supply failure.
21. Equipment should conform to the equivalent Indian Standard or other Equivalent international or third country standard (CE/FDA certification) for good manufacturing practice and safety.

## NEONATAL VENTILLATOR WITH ALL CONNECTOR

1.The Ventilator should be a microprocessor based for use on premature babies from 300gms onwards .

2.It should be a time cycled, pressure limited ventilator with continuous flow and with the following settings parameters.

- i. Continuous flow range : 1 to 30 LPM
- ii. Inspiratory Time : 0.1 to 3 seconds
- iii. Expiratory Time : 0 to 99.9 seconds
- iv. Ventilation Frequency : 0-150 BPM
- v. Inspiratory Pressure : 0 to 72cm H<sub>2</sub>O
- vi. Trigger Mechanism : Flow Triggering mechanism detected by flow sensor of the circuit.
- vii. PEEP : 0 to 30LPM
- viii. Base Flow : 1 to 30LPM
- ix. Assist Sensitivity : 0.2 to 5.0 LPM

3.It should have the following operating modes:

Assist-control, SIMV/IMV, SIMV/PSV, CPAP, PSV, Manual Breath, PEEP, Apnea backup ventilation with user selectable apnea time interval, etc.

4.The unit should be mobile mounted on trolleys for easy transportation

5.The unit should have provision for Nebulization during ventilation and for nasal CPAP function.

6.The ventilator should have built in digital monitoring facility to display important parameters like breath rate, patient initiated indicator (LED), minute volume, tidal volume (inspired and expired), percentage of tube leak, inspiratory time, expiratory time, I:E ratio, peak inspiratory pressure, mean airway pressure, peep.

7.The Ventilator should have easy control setting, alarm setting and monitoring panels for user friendly operation with electronic knob controls.

8.The unit should have audio visual alarm facility to indicate the following alarm conditions. Low PEEP, high breath rate, low Inspiratory pressure, high pressure limit, low gas supply, low battery patient circuit, prolonged inspiratory pressure, flow sensor fault, Apnea, etc.

9.The unit should have an integral air oxygen blender with proper bleed facility for accurate blending of air and o<sub>2</sub>.

10.The unit should be supplied with electronic humidifier, pole stand and two sets of reusable patient circuit.

11.The unit should have an analogue pressure gauge to measure proximal pressure and peep.

12.The unit should be supplied with the reusable flow sensor which is factory calibrated and should not have onsite calibration.

- 13.The unit should have volume limit feature to avoid volutrauma conditions.
  - 14.The unit should operate on 240 VAC supply with built in rechargeable battery backup for 30 minutes.
  - 15.The unit should have an audio-visual low battery alarm to indicate the user when the battery voltage falls to a level below which the unit may fail to perform satisfactorily.
  - 16.The ventilator shall be adequately protected against and be able to withstand the prevailing climatic conditions in Bihar.
  - 17.The Ventilator should be suitable for the operation with temperatures from 10 degrees to 40 degrees with a relative humidity of 0 – 99%.
  - 18.The Ventilator should be provided with a line power cord of acceptable durability, quality, length and current carrying capacity.
  - 19.Equipment should include power [plugs that are sufficient for maximum voltage and current of the equipment.
  - 20.If fuses are used, a spare fuse should be provided permanent marking near each fuse holder should indicate fuse ratings.
  - 21.Equipment performance should not be affected by electromagnetic interference radiated or conducted through power lines from another device.
  - 22.Equipment should have no sharp edges, should be securely mounted and should provide adequate protection against moving and electrically energized parts. for easy access to serviceable parts.
- 21.Packaging and Storage:**  
Packing of the equipment should be easy to open and well labeled and marked with devices name and sellers name and address.
- 22.Equipments should be able to withstand temperature and humidity extremes likely to be encountered during storage and transportation.
  - 23.Equipment should conform to international quality standards like CE/USFDA

## Portable X Ray Machine

High Frequency mobile X ray machine with output 60mA or more. The mobile X ray equipment is required to perform x ray studies in emergency and trauma center and bedside in wards and ICU. The unit should be compact, lightweight and easily transportable. It should have following specifications.

1. The unit should be operating on mains voltage from single phase 170-260V AC.
2. **Generator:**
  - Power: 2.5 KW or more
  - KVp range: 40-100KVp
  - mAs range: 200mAs or more
  - mA range: 30 mA- 60 mA
  - Exposure time: 3 ms to 4 sec.
3. **X ray Tube:** Stationary Anode tube. Anode speed 300 rpm, thermal capacity 40 KHU or better.
4. The tube stand should be fully counterbalanced with rotation in all directions.
5. The unit should have automatic calibrator. It should have auto shut off facility for lamp.
6. The equipment should have cassette storage Box for minimum of 4 cassettes.
7. The equipment should be light weight not more than 160 KG.
8. The unit should have small foot print.
9. The height of the column stand should not be more than 150 cm for easy transportation in the Lift and areas of small height doors.
10. The unit should have effective braking system for parking.
11. The equipment should be AERB approved.

# **TECHNICAL SPECIFICATION OF FULL FIELD DIGITAL MAMMOGRAPHY SYSTEM ALONG WITH DIGITAL STEREOTACTIC BREAST BIOPSY SYSTEM**

Full field digital mammography system should offer the breast imaging and breast cancer detection through the combination of lowest possible dose of radiation and state of art facility. The system should deliver outstanding performance with high-resolution image quality at the lowest possible dose.

## **Technical Specifications**

### **(1) X- Ray Generator**

- i) X- Ray generator system should be high frequency constant with minimum rating of 3.2Kw with 100 mA at 35KV.
- ii) 22kV to 35kV in 1 kV increment or more
- iii) 3 to 500 mAs or More
- iv) Range of mA should be 100mA at large focal spot and 25at small focal spot or more

### **(2) X – Ray Tube:**

- i) Bi directional rotating anode with speed 8500 RPM or more.
- ii) Heat storage capacity of the x ray tube must be 150 KHU or more.
- iii) Dual focal spot size of 0.3mm (Large) & 0.1mm (Small).
- iv) Dual filtration X-ray tube.
- v) Fully automatic collimation or user selectable. System must have predefined collimation setting.

### **(3) C- Arm assembly:**

- i) Fully motorized vertical and iso centric rotation and vertical movement
- ii) Define Angular rotation of the c-arm
- iii) SID – 65cm or more with removable patient face shield.

### **(4) Compression Mode:**

- i) Precompression, full compression, Dual compression, manual compression, Compression paddle tilt. (Motorized/ User selectable)
- ii) Magnification factor of minimum 1.5 x
- iii) Digital Display of compression should be available.

### **(5) Digital Flat Panel Detector**

- i) Detector should be TFT Based Direct capture technology, with X ray absorption material of Amorphous Selenium/Cesium, size of 24 x 29 cm.
- ii) Pixel size of the detector should be 100 micron or less.
- iii) Spatial resolution 35Lp/mm
- iv) Dynamic range - 13 bit or less image data for output of image.

### **(6) Acquisition Workstation:**

- i) Multi core Intel based CPU with minimum 1 GB RAM, hard drive of 60 GB or more DVD+/- R/W
- ii) 3 mega pixel grey scale medical grade LCD Display.

- iii) Facility for Work list, print, storage query/ retrieve, modality performed procedure step, scheduled workflow, patient information, reconciliation, and mammography image.
  - iv) Image storage capacity of the system should be approximately 9000 screening mammography.
- (7) Reporting Workstation**
- i) High end Dual processor window based 3GB high Speed RAM with high speed Hard disc of minimum 160 GB
  - ii) Dual Medical Grade monitors with Display of 5 Mega pixel each (High definition) 19" or more.
  - iii) Dedicated mammography workflow keypad, Mouse Key Board.
  - iv) High accuracy calibration photometer with automated QC monitors Software.
  - v) Customizable Image layout, orientation of Images for diagnostic, screening or multiple modalities, annotation.
- (8) Mammography Image Management System**
- i) Should provide a unique workflow and archiving requirements for mammography with PACS or without PACS network.
  - ii) Should design to support all women's imaging modalities and unique workflow needs
- (9) CAD Solutions**\_CAD solution should be FDA approved.
- (10) Digital Stereotactic Breast Biopsy along with Vacuum Assisted Breast Biopsy Device:**
- i) System should be patient comfort, efficient accurate testing in upright position with superb image quality with advanced digital spot mammography system,
  - ii) Motorized release of compression.
  - iii) Stereotactic guidance system should be Cartesian coordinate system with smart window, accuracy of  $\pm 1.0\text{mm}$ , Stereotactic angle of  $\pm 15^\circ$
  - iv) Facility for needle core biopsy, fine needle aspiration and wire localization with 10 Nos. of core needle, 10 Nos. of aspiration needle and bard magnum gun for core procedure.
  - v) Vacuum assisted breast biopsy system should standard along with the Stereotactic biopsy system with all reusable standard accessories.
- (11) Out Put Device**
- Dry direct Digital camera of minimum 500 DPI or more with 3 online sizes, and supporting 5 sizes films, one of them should be 11 x 14 inch. should be provided.
- (12) Standard Accessories along with Mammography Unit:**
- i) Screening Compression paddles of 18 x 24 and 24 x 30 cm.
  - ii) Small breast compression paddle.
  - iii) 9 cm or more contact paddle for diagnostic compression
  - iv) 7 cm or more spot contact paddle
  - v) Frame less spot compression paddle
  - vi) 9 cm or more magnification compression paddle
  - vii) 7 cm or more spot magnification paddle.
  - viii) Magnification platform
  - ix) Dual function footswitches

**(13) Standard Accessories along with Biopsy Unit:**

- i) Scout biopsy paddle.
- ii) ultrasound biopsy paddle
- iii) perforated biopsy paddle
- iv) scout and digital stereo aperture
- v) Gel pads, arm slings, filter panel foam cushion.
- vi) Stool with backrest
- vii) Quality assurance needle
- viii) Needle guide
- ix) Biopsy gun holder
- x) Air Phantom
- xi) DSM computer cart
- xii) Needle guide holder
- xiii) Biopsy Gun - 4 Nos.

**(14) DICOM:** The system should be DICOM 3 ready for send and print images.

**(15) Other Requirements**

- i) UPS along with 3 Hrs. back up time
- ii) Storage space for accessories
- iii) Zero Lead aprons 4 in Nos. (light weight), 5 nos. face shields and 5 nos. thyroid shields.
- iv) One set of mammography atlas.

**(16) Training:**

2 weeks application training at the site of installation on mammography and Stereo tactic biopsy device in a good centre for biophysics for technicians.

**(17) Approvals--**Complete system should be FDA/CE and AERB approved.

The bidder should provide AERB approved certification for machine and site. Please enclose the necessary certificates

**(18) Principal manufacturer to give undertaking that they will maintain and service the equipment in case Indian agent / supplier fails to provide the service.**

**(19) TURNKEY:**

The bidder will carry out installation on turnkey basis. To assess the turnkey costs the bidder may visit the sites before quoting for the equipment.

# **C-ARM MOBILE IMAGE INTENSIFIER**

## **TECHNICAL SPECIFICATIONS**

### **I X-RAY GENERATOR**

- i. Type: High Frequency minimum 20 Khz.
- ii. Fluoroscopy anode potential: 40 to 110Kvp (1Kvp step)
- iii. Fluoroscopy mA range: Normal mode : 0.5 to 4 mA
- iv. Power: 2.5 Kw

### **II. X-RAY TUBE**

- i.Type: Stationary anode
- ii.Focal spot: 0.6/1.5 mm

### **III. IMAGE INTENSIFIER**

- i.Input field size: 9" (Triple field)
- ii.Grid on the entrance field: circular grid.

### **IV. TV CAMERA SYSTEM**

- ii.Type: CCD with 752x582 pixels.
- iii.Memory: Minimum 25 images non-volatile storage.
- iv.Video Standard: PAL/NTSC

### **V TV monitor**

- i. Two 15" or more medical grade LCD/TFT/CRT monitors.
- ii. One for LIH and one for memory display

### **VI C-ARM CART**

- i. SID: 880mm
- ii. Orbital Travel: 115° (90/25)
- iii. C-arm Pivotal rotation:  $\pm 180^\circ$
- iv. Horizontal Travel: 200mm
- v. Vertical Travel: 400mm
- vi. Panning Movement:  $\pm 12.5^\circ$
- vii. Depth of C-arm: 600mm

### **VII OTHER SPECIFICATONS**

The unit should have the following facilities

- i. Automatic KV and mA technique selection and manual mode.
- ii. Cumulative exposure timer for fluoroscopy.
- iii. Audible and Visual indication for X-ray emission.
- iv. 360 ° rotations of images should be possible for LIH image after fluoroscopy and should not be CRT based rotation.
- v. Image vertical and horizontal reversal should be possible on the LIH image after fluoroscopy.

- vi. Should not be a PC based system.
- vii. Iris collimation or two pairs of parallel shutters which can be controlled independently and can be rotated.
- viii. Should have at least 20cm distance between the focal spot and skin for radiation safety.
- ix. Should have a single double steering wheel with 180° rotation.
- x. Two sets of sterile drape for the X-ray tube assembly, Image intensifier and C-arm and clips to hold the drape on the c-arm should be provided.
- xi. Cassette holder should be supplied.
- xii. Five lead aprons with thyroid guards
- xiii. The quoted model and tube should be AERB type approved for usage up to 4mA. Relevant copies of the certificate should be attached with the bid.

#### **VIII POWER REQUIREMENTS**

- i. Single phase, 230 Vac, 50Hz
- ii. Suitable stabilizer should be provided along with the unit.

#### **IX SPECIFICATION OF LEAD APRON.**

- i. Should be AERB approved.
- ii. Should be light weight 0.5mm lead equivalent.
- iii. Should be hook and loop type (Velcro).
- iv. Should be supplied along with thyroid guard

## **Portable Ultrasound Machine**

- Portable ultrasound machine with biopsy attachment.
- Scanning mode: linear and convex
- Display mode: b mode, m mode and b+m
- Monitor screen should be 10” high resolution latest technology
- Display depth:25 cm max
- Magnification: 8 steps
- Facility for image zoom, image reverse, image freeze and image magnification, reduction
- Alphanumeric keyboard for character entry and character display
- Gynaecology measurement facility for uterus, cervix, special probe for gynecology to be provided
- 6-13 MHZ linear Array probe.
- 3.5 mhz convex probe.
- Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.

### **Accessories:**

Ultrasound gel, biopsy attachment and black and white video printer with at least 10 no.Rolls,mobile stand.

## **Specification for Portable Color Doppler USG Machine**

- The system should be a multipurpose, high performance color imaging system designed for abdominal, vascular, obstetrics, gynecology, cardiology,
- Gray Scale, 256 Channel system, Scanning depth 25 cms or more should be possible. Scanning Methods should be - Electronic Sector, Electronic Convex and Electronic Linear. Tissue Harmonic imaging in Abdomen and cardiac.
- Dynamic range should be up to 120 db, should have 2 or more active probe ports,
- Should have 15" or more high resolution monitor with tilt and swivel facility
- Transducer Types should be - Electronic- Convex Array, Sector phases array, Micro convex array and linear array probe
- Operating Modes- B mode, M-Mode, Color M-Mode, power Angio Mode, Directional power Angio, pulse Wave Doppler with preferably CFM Mode, Steerable Continuous wave Doppler Mode, Flexibility of M-Mode curve across any plane
- Alphanumeric keyboard with hard keys and time gain compensation controls, Multiple focus selection – minimum 4 focal point,
- Hard Disk Image Storage (up to 30 GB Memory) and flexibility to review the stored images in clipboard format.
- Should be able to store single frames and Cine loops and post processing of all stored images & Doppler spectrum
- Integrated CD Read/ Write drive to store and review images on CD with option of storing images in different formats including AVI
- Complete measurement and analysis package with Real Time Doppler Calculations Fetal trend graph etc.
- Trapezoidal imaging (Convex like image on Linear probe) to view complete large organs
- The system should have weight less than 10 kgs and can be hand carried
- **Probes** -All probes should be wide band frequency probes for Convex, TV, Linear, Tissue Harmonic Imaging should be available in all the above probes, and Biopsy should be available with Convex & TV / TR Probe.
- All transducers should have broad band width technology for extremely high resolution imaging. All transducers should have multifrequency selection.
  - Broad band convex array transducer with frequency range 2-5MHz
  - Broad band linear array probe with frequency range 3-6 MHz
  - Broad band transvaginal with frequency range 5-9 MHz
  - Broad band transrectal probe with frequency range 5-9 MHz

### **Accessories :**

Black & White Thermal Printer, Color Inkjet Printer, UPS with 30 Minutes Battery BACK UP.

## 1 Description of Function

- |     |  |
|-----|--|
| 1.1 | Autoclaves are required for sterilizing an object in high temperature and high pressure steam. |
|-----|--|

## 2 Operational Requirements

- |     |  |
|-----|--|
| 2.1 | High pressure horizontal cylindrical steam sterilizers pressure type designed to operate on steam from steam generator |
|-----|--|

## 3 Technical Specifications

- |      |  |
|------|--|
| 3.1  | Rectangular, horizontal, single door, high pressure, high vacuum fully automatic, microprocessor based autoclave for sterilizing hospital materials  |
| 3.2  | Electrically operated 18 KW in built electric steam generator with the unit.   |
| 3.3  | Double jacketed, Outer jacket shall be SS 304; and inner chamber shall be SS 316 heavy duty, with full argon welding.  |
| 3.4  | Door should be (SS 316) and designed with several automatic mechanical and control features that provide for safety. Door must be lined with silicone rubber gaskets (expansion type).   |
| 3.5  | Temperature adjustable from 121o to 135o C   |
| 3.6  | Working pressure range from 15 to 32 psi   |
| 3.7  | Sterilization cycles:<br>a. The autoclave residence time should not be less than 60 minutes if the autoclave operates at the working temperature (inner chamber) of 121oC at a pressure of 15 pounds per square inch (psi).<br>b. If the temperature of inner chamber is 135o C at a pressure of 31 pound per square inch (psi) then the autoclave residence time should be not be less than 45 minutes. |
| 3.8  | Size: 1000-1100 mm (W) x 1100-1400 mm (H) x 2000-2200 mm (L) approx.   |
| 3.9  | Autoclave should be properly equipped with door safety locks, steam traps, pressure gauges and safety valves for chamber and jacket.   |
| 3.10 | Autoclave should have insulation jacket with glass wool, covered with aluminium foil.  |
| 3.11 | The unit should have a timer from 1 to 250 minutes that may be used for process or to record turn- on and shutdown timings.  |
| 3.12 | The unit should have integral alarms that ring, flash, or otherwise display information when temperature set-points are exceeded or fallen below.  |
| 3.13 | Pressure safety valve, over-temperature limiter, anti-scorch limiter, door (lid) interlock, overpressure limiter, current fuse   |
| 3.14 | The unit includes a data logger or chart recorder for monitoring operational history   |
| 3.15 | Integral controls, keypad, and/or display on the panel of the unit. Controls mounted over the door are not acceptable. The control panel must document all cycle information. including key transition points in the cycle. alarms and deviations that   |

	may jeopardize the sterilization process, resulting in inadequate sterilization.
3.16	The sterilizer should be supported on a steel stand, appropriately coated for corrosion protection.
3.17	Gauge glass with safety valves and dram temperature gauge
3.18	Pressure switches for economic power consumption and prevention of accidental pressure build up.
3.19	The steam generator shall be electrically heated by immersion heaters having mineral filled sheathed heating elements. Heater plate: Brass/SS
3.20	Air shall be removed from the sterilization chamber after loading the sterilizer by one or more pressure/vaccum pulsing for proper steam penetration either manually or automatically through the automatic control system
3.21	Vacuum breaker for jacket
3.22	Low water protection with audio-visual indicator

#### 4 System Configuration Accessories, spares and consumables

4.1	System as specified-
4.2	It should be supplied with the following accessories(which should be quoted separately individually):- a. Carriages b. Baskets c. Trays d. Trolleys e. Heating element f. Gasket g. Valves
4.3	All consumables required for installation and standardization of system to be given free of cost.
4.4	All infrastructural work including plumbing and electrical work will have to be done by the supplier. The hospital will only provide space and water outlet

#### 5 Environmental factors

5.1	The unit shall be capable of being stored continuously in ambient temperature of 0 - 50deg C and relative humidity of 15-90%
5.2	The unit shall be capable of operating in ambient temperature of 20-30 deg C and relative humidity of less than 70%

#### 6 Power Supply

6.1	Power input to be 440VAC, 3-phase, 50Hz fitted with Indian plug
6.2	The immersion heaters shall be wired for operation on 3 Phase 4 wire 400/440 V 50 Hz AC supply and electric load for the unit shall be 18 KW. The control will be of independent switch

## 7 Standards & Safety

7.1	Should be FDA , CE, UL or BIS approved product (as per IS:3829 (Part2)/1978 (reaffirmed 2001) with amendment Nos. 1 to 4)
7.2	Manufacturer should have ISO certification for quality standards.
7.3	Autoclave should conform to IS 3829 part 1, 1999 or later standards.
7.4	The immersion heater used for electric heating shall have mineral filled sheathed heating elements and shall be BIS marked conforming to IS:4159/2002 and complying with safety requirements as per IS:302-2-201(1992).
7.5	Electrical safety conforms to standards for electrical safety IEC-60601 / IS-13450
7.6	Should be compliant to ISO 13485: Quality systems – Medical devices - Particular requirements for the application of ISO 9001 applicable to manufacturers and service providers that perform their own design activities.

## 8 Training

8.1	Comprehensive training for staff of user department and support services till familiarity with the system.
-----	--

## 9 Warranty & Service

9.1	After sales service must be provided in the city of installation. In situations requiring service/repair of the unit outside the city of installation, the expenditure on account of this will have to be borne by the supplier
-----	---

## 10 Documentation

10.1	Product Literature in original along with that of accessories and indigenous components, if any. Photocopies/computer generated copies are not acceptable
10.2	Statement of compliance with tender specifications with clear and unambiguous links to relevant portions of product literature/authentic document, which should be highlighted. Alternatives provided for noncompliant specifications with justification must be described in detail with supporting literature.
10.3	Certificate of compliance with standards and approvals stated above
10.4	Certificate of manufacturer/principal regarding authorisation of service facility provided by the supplier
10.5	List of Equipment available in the Service Centre for providing calibration and routine Preventive Maintenance Support as per manufacturer documentation in service/technical manual.
10.6	List of important spare parts and accessories, which are required for maintenance and repair, with their part number and costing.
10.7	Terms and conditions of warranty and CMC including schedules of visit by service personnel with check list of services to be carried out

10.8	Commitment for supply of log book with check list for daily, weekly, monthly and quarterly maintenance checklist along with contact details of service personnel along with the equipment. The job description of the hospital technician and company service engineer should be clearly spelt out in the log book.
10.9	List of users of quoted model with performance certificate from major institutions

## **ECHOCARDIOGRAPHY**

### **TECHNICAL SPECIFICATIONS**

1. Should be a stand alone system integrated on a light weight mobile cart.
2. The system should be a color Doppler Echocardiography all digital beam former system to study the anatomical abnormalities and blood flow in the heart and associated vessels. Should be a stand alone system integrated on a light weight mobile cart.
3. Should be a latest generation Electronic Phased array Color Doppler system with minimum 512 Electronic independent channels.
4. Should have 256 gray shades for sharp contrast resolutions.
5. Should be supplied with adult and pediatric cardiac and vascular probes of wide band transducers without frequency selection for higher sensitivity of response over a broad frequency range of operation.
6. Should have 2D, M-mode, Anatomical M-mode, Color M-mode, PW and CW doppler, Steerable CW doppler.
7. The system should have a very high dynamic range of at least 200dB to pick up subtle echoes.
8. Should have three active ports.
9. Should have 2-4 Mhz broadband phased array sector probe for adult cardiac imaging.
10. Should have 3-8 Mhz broadband phased array sector probe for paediatric and neonatal cardiac imaging.
11. Should have 3-12 Mhz broadband Linear Array probe for vascular imaging.
12. Should have multi frequency convex array probe 3-10 MHz for paediatric imaging
13. Pencil probe (optional)
14. Should have advanced tissue Harmonic Imaging.
15. Should have color flow imaging.
16. Should have color Tissue Doppler Imaging.
17. Should have gain control in Axial Plane.
18. Should have triple imaging possibility on the system.
19. Should have PW/CW Doppler facility in all imaging phased array sector probes.
20. Should have 15" or more high resolution TFT monitor with tilt and swivel facility and should be able to view in all angles and all light conditions.
21. Should have greater than 5000 images in the system hard disk drive
22. Should have in built CD/DVD writer.
23. Should have patient reporting page with embedded images.
24. Should have full functional measurement facility and calculation should be possible.
25. Should be supplied with thermal printer and 6 packs of thermal paper and the unit should have option to connect external printer.
26. Unit should function with 200-240Vac, 50/60 Hz input power supply.
27. Unit should be supplied with suitable UPS with a minimum 30 minutes back-up time.
28. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.

## Mortuary Chamber

- Capacity :Six Bodies
- Temperature Range :-2 degree to 5 degree C
- Controller :Microprocessor Based
- Interior Panel :Made of Stainless Steel
- Outer Panel :Made of Stainless Steel
- Door :Hinged Insulated Doors with magnetic gaskets
- Insulation :Polyurethane form with thickness of 80-100mm
- Body Trays :Stainless Steel with Telescopic Mechanism
- Compressor :Heavy Duty low noise & minimal vibration
- Condenser :Efficient with Auto-condensate evaporating system
- Air Circulation :Forced Air Circulation
- Alarms :Audio-Visual Alarms to warn high or low Power temperatures.
- supplies :220 V/50 Hz
- Accessories :Temperature recorder, data logger, voltage safety system, alarm system for various parameters battery backup rechargeable.
- Should be CE/BIS approved product

# **Equipment Specifications for 32 CHANNEL DIGITAL EEG SYSTEM FOR NEUROLOGY**

## **A. Description of Function**

An electroencephalograph uses electrodes placed on a patient's scalp to measure, amplify, display in graphic form, and record the weak electrical signals generated by the brain.

Electroencephalography is useful in observing and diagnosing a variety of neurologic conditions, including epilepsy, related convulsive disorders, and brain death. It can also be used to evaluate psychiatric disorders and differentiate among various psychiatric and neurologic conditions. In addition, electroencephalographic studies can assist in localizing tumors or lesions on or near the surface of the brain.

## **B. Operational Requirements**

EEG System complete with software for acquisition and review and the compatible computer with necessary interface and printer is required.

## **C. Technical Specifications**

- Hardware:
  1. Should be PC based with minimum following PC specifications: core i5, 1 GB DDR RAM, 160 GB HDD, CD/DVD RW, 17-25" LCD TFT Display, Key Board, Mouse and UPS.
  2. Number of EEG Channels should be 32 with coding, including eight channels for Polygraphy. Also any two channels can be configured as Bipolar, AC or DC through software.
  3. Facility for simultaneous sampling of all EEG channels and multiple sampling rates.
  4. Photic Stimulator with software programmable for manual or automatic sequences.
  5. Networking facility
  6. DICOM compatible.
- Technical Specifications:
  1. CMRR should be > 110 dB or better.
  2. Noise < 2uV peak to peak.

3. Input Impedance > 100 Mohm.
  4. 16 bit ADC resolution voltage of 0.153 uV.
- Acquisition Software:
    1. Facility to combine all user defined settings into templates or protocol, for use in different applications.
    2. Facility for Individual Channel Control, Customization of Montages, along with Remontage Capabilities.
    3. Should display a graphical view of the current montage during the EEG recording.
    4. Facility to define New Sensors should be possible as standard i.e assign to amplifier inputs, define traces in a mintage, define calculated channels (Average, Source/Laplacian), or define trends.
    5. Facility to click any point to display corresponding traces & Slide pointer to change displayed 146 duration of the Overview.
    6. Facility for sortable list of all events placed in the recording, both automatically and manually.
    7. Facility to review and add events to recorded traces.
    8. Facility for automatic time counters and fevent insertion during Hyperventilation.
    9. Facility to controlled display Sensitivity for User defined value.
    10. Facility to choose Low & High Cut Filters along with facility to enter any user defined value.
    11. Facility to file zip.
    12. Facility of configurable Time Base.
    13. Spike & Seizure software
    14. Trend Analysis software.
  - Review Software:
    1. Paging facility as Automatic Paging, Mouse controlled Paging and/ or Keyboard Paging.
    2. Playback of EEG for one or more channels.
    3. Facility for Zoom/ Magnify EEG trace,
    4. Facility for Copy & Paste of EEG or Trends to reports and presentations

5. Facility for Automatic generation of reports.

6. Facility for viewing several recordings in tiled or cascading windows.

- Patient Administration Software:

1. Network supported patient and test management software, archive to CD or DVD, powerful search, patient folder, workspaces.

2. Should have an option of upgrading the digital EEG to Video EEG with day/night camera using MPEG-2 3rd generation technology.

- System Configuration Accessories, spares and consumables:-

System as specified

Compatible Laser Printer with 600 DPI Resolution and A4 Size printing facility.-01

Standard accessories to include the patient cable and connectors with electrodes and Papers for at least 1000 EEG Exams and all the necessary power cables and other interfaces.

#### OPTIONAL REQUIREMENTS COMPONENTS FOR VIDEO EEG UPGRADATION.

#### **Environmental factors**

The unit shall be capable of being stored continuously in ambient temperature of 0 -50 deg C and relative humidity of 15-90%

The unit shall be capable of operating in ambient temperature of 20-30 deg C and relative humidity of less than 70%

#### **Power Supply**

Power input to be 220-240VAC, 50Hz fitted with Indian plug147

Resettable overcurrent breaker shall be fitted for protection

Voltage corrector/stabilizer of appropriate ratings meeting ISI Specifications.( Input 160-260 V and output 220-240 V and 50 Hz)

Suitable UPS with maintenance free batteries for minimum one-hour back-up should be supplied with the system.

#### **Standards, Safety and Training**

Manufactures/Supplier should have ISO certificate to Quality Standard.

Should be compliant with IEC 61010-1:(or any international equivalent eg EN/UL 61010)  
covering safety requirements for electrical equipment for measurement control and laboratory use

Should be FDA, CE, UL or BIS approved product

Comprehensive training for lab staff and support services till familiarity with the system.

Shall be certified to be meeting the safety standards IEC- 60601-2-26 PART 2: Particular requirements for safety of EEG Systems.

### **Documentation**

User/Technical/Maintenance manuals to be supplied in English.

Certificate of calibration and inspection.

List of Equipments available for providing calibration and routine maintenance support as per manufacturer documentation in service / technical manual.

List of important spare parts and accessories with their part number and costing

Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out.

Compliance Report to be submitted in a tabulated and point wise manner clearly mentioning the page/para number of original catalogue/data sheet. Any point ,if not substantiated with authenticated catalogue/manual, will not be considered. Equipment

## **Specifications for EMG / EP SYSTEM WITH ACCESSORIES (PC BASED TWO CHANNEL)**

### **Description of Function**

Electromyographs detect, process, and record the electrical activity of the skeletal muscles .EP graphic recorders measure and document the brain's electrical response to visual, auditory, or somatosensory stimuli. Electromyographs test the functional ability of peripheral nerves by using integral stimulators to measure nerve conduction velocity (NCV), the rate at which a nerve can carry a signal from the point of stimulus by an electrode to the muscle that it innervates.

### **Operational Requirements**

EMG System complete with EP recorders and all software and hardware is required.

### **Technical Specifications**

1. Standard programmes for recording motor nerve conduction velocity, sensory nerve conduction velocity, repetitive nerve stimulation, F response, H reflex and blink reflex.
  2. Standard programme for routine electromyogram (EMG) recording, motor unit potential (MUP) analysis, interference pattern analysis, single fiber EMG, jitter analysis
  3. Standard programme for recording brain stem auditory evoked response, middle latency response and slow vertex response
  4. Standard programme for recording pattern reversal visual evoked potential (VEP), LED VEP.
  5. Standard programme for recording P300
  6. Standard programme for recording somatosensory evoked potentials (upper limb & lower limb) and short latency evoked potentials
- OPTIONAL Specifications ( features can be added or deleted depending upon user department)
7. Standard programme for recording sympathetic skin response (OPTIONAL)
  8. Standard programme for recording electroretinogram (ERG) and electrooculogram (EOG)(Optional)
  9. Facilities for checking electrode -skin impedance( optional required alongwith Standard programme for recording sympathetic skin response)

#### Amplifiers:

- i. Input impedance: 100 mega ohms or more
- ii. Sensitivity: 2 microvolt – 10 millivolts per division
- iii. Time base: 0.1 millisecond – 0.5 seconds per division in variable steps
- iv. Filters: Standard low cut, high cut filters for all recordings

PC workstation with Core 2 Duo CPU with inkjet printer(colour),17" LCD/TFT Monitor, 120 GB HDD, DVD Read/Write, 1GB RAM.4 USB Port.

### **System Configuration Accessories, spares and consumables**

1. Surface stimulating and recording electrodes – 10
2. Concentric needle electrodes (30 mm long with connecting cable) – 4
3. Minimum number of Single fiber EMG electrode – 2

4. Ground electrode – 2
  5. Headphones and child ear tips with cables – 2
  6. VEP monitor and LED goggles – 1
  7. Flash stimulator – 1
  8. Electrode gel – 10
  9. Recording paper – 3
  10. Power cable – 2
  11. Ground lead – 2
- Optional specifications:
12. ERG contact lens electrode – 2 (optional : to be asked for in the case of the ERG provision in the programme)

### **Environmental factors**

Shall meet IEC-60601-1-2 :2001(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility.or should comply with 89/366/EEC; EMC-directive.

The unit shall be capable of operating continuously in ambient temperature of 20-30 deg C and relative humidity of 15-90%

The unit shall be capable of being stored continuously in ambient temperature of 0-50deg C and relative humidity of 15-90%

### **Power Supply**

Power input to be 220-240VAC, 50Hz fitted with Indian plug

UPS of suitable rating with voltage regulation,spike protection and maintenance free batteries for 60 minutes back up

### **Standards, Safety and Training**

Should be FDA, CE, UL or BIS approved product

Comprehensive training for lab staff and support services till familiarity with the system.

Manufacturer should have ISO certification for quality standards.

Should have local service facility .The service provider should have the necessary equipments recommended by the manufacturer to carry out preventive maintenance test as per guidelines provided in the service/maintenance manual.

Shall meet IEC 60601-2-040 Safety requirements - Part 2-040: Particular requirements for Electromyographs and Evoked Response Equipments

### **Documentation**

User/Technical/Maintenance manuals to be supplied in English.

List of important spare parts and accessories with their part number and costing.

Certificate of calibration and inspection.

List of Equipments available for providing calibration and routine Preventive Maintenance Support. as per manufacturer documentation in service/technical manual.

Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist.

The job description of the hospital technician and company service engineer should be clearly spelt out.

Compliance Report to be submitted in a tabulated and point wise manner clearly mentioning the page/para number of original catalogue/data sheet. Any point ,if not substantiated with authenticated catalogue/manual, will not be considered.

## Evoked Potential Machine (digital)

### **Technical Specifications for EMG/EP system 4 channels: for recording of the evoked response & EMG with accessories**

Essential Requirements:

- 1) Minimum four channel
- 2) With facility for up gradation
- 3) Standard programs for recording motor nerve conduction velocity, sensory nerve conduction velocity, repetitive nerve stimulation, F response, H reflex and blink reflex.
- 4) Standard program for routine electromyogram (EMG) recording motor unit potential (MUP) analysis, interference pattern analysis, automatic computation with display .
- 5) Standard program for recording sympathetic skin response
- 6) Standard program for recording brain stem auditory evoked response, middle latency response, VEMP and slow vertex response.
- 7) Standard program for recording pattern reversal visual evoked potential (VEP) and LED VEP
- 8) Standard program for recording P300 with audiovisual paradigms
- 9) Standard program for recording somato-sensory evoked potentials (upper limb & lower limb) and short latency evoked potentials
- 10) Facilities for checking electrode-skin impedance
- 11) PC requirements: Intel® Core™ i5-760 processor (2.80GHz, 1333MHz FSB, 8MB Cache) Genuine Windows® 7 Professional, 64bit (English) or higher ; 21.5” Full HD Widescreen Flat Panel Monitor ; 6 GB DDR3 SDRAM, 500GB SATA Hard Drive ; Single Drive: Blu-ray Disc Combo (DVD+/-RW + BD-ROM). Facility for internet connectivity, with facility of up gradation
- 12) Color laser printer & UPS with 20 minutes back up for whole system along with computer
- 13) Patient Data management software & archiving facility
- 14) MS Word based report generation facility
- 15) Amplifier
  - i. Input impedance: 100 mega ohms or more
  - ii. Sensitivity: 2 microvolt – 10 millivolts per division
  - iii. Time base: 0.1 millisecond – 0.5 seconds per division in variable steps
  - iv. Filters: Standard low cut filter (0.2 to 2KHz), high cut filters for all recordings

### **Standard accessories to be provided [items-quantity in numbers]**

1. Surface stimulating (reusable) – 4 [Four]
2. Surface Recording electrodes (reusable) – 12 [Twelve]
3. Concentric needle electrodes (adult size, disposable, with adequate length of the connecting cable) – 50 [fifty]
4. Concentric needle electrodes (pediatric size, disposable, with adequate length of the connecting cable) – 50 [fifty]
5. Needle holder for disposable needles – 3 [three]
6. Single fiber EMG electrode with needle holder – 1 [one]
7. Ground electrode – 4 [four]
8. Headphones and child ear tips with cables – 2 [two]
9. VEP monitor and LED goggles – 1 [one]
10. Headphones for auditory evoked potentials – 1 [one]
11. Flash stimulator – 1 [one]
12. Electrode gel – 10 [ten]
13. EMG conductive paste (200 gms or more) – 10 [ten]
14. Recording paper – 3 [three]

15. Power cable – 2 [two]
16. Ground lead – 2 [two]
17. Power requirements: 220 V AC, 50 Hz

1. All essential spare parts and service manuals should be available with the local service center and all steps should be taken for immediate servicing to prevent down time

Annual Maintenance Contract:

2. A copy of service manual should be available with local service center

**Installations, Commissioning, Testing, Maintenance and After Sales Service:**

1. The equivalent and all accessories should be installed, tested and commissioned at the Department where the instrument is installed, free of cost.
2. One engineer should be posted for a week for imparting training at site of installation
3. All spare parts and consumables should be available with the supplier or principals for a period of 10 years.

**SECTION VI: SAMPLE FORMS**

## Notes to Bidders on the Preparation of Sample Forms

---

The Purchaser has prepared the forms in this section of the Bidding Documents to suit the specific requirements of the procurement. In its bid, the Bidder **MUST** use these forms (or forms that present in the same sequence substantially the same information). If the Bidder has a question regarding the meaning or appropriateness of the contents or format of the forms and/or the instructions contained in them, these questions should be brought to the Purchaser's attention as soon as possible during the bid clarification process, by addressing them to the Purchaser in writing pursuant to ITB Clause 7.

The Purchaser has provided explanatory text and instructions to help the Bidder prepare the forms accurately and completely. The instructions that appear directly on the forms themselves are indicated by use of typographical aides such as italicized text within square brackets.

In preparing its bid, the Bidder **MUST** ensure all such information is provided and that the typographical aides are removed.

# **1. Bid Form**

Date: 10<sup>th</sup> Oct 2013

[ insert: *date of bid* ]

[ Purchaser specify: “IFB No.: BMSICL/2013-14/MC-005” ]

[ insert: **Procurement and Rate Contracting of Medical equipment for Government Medical Colleges in Bihar** ]

To:

*Managing Director,  
Bihar Medical Services and Medical Services Corporation,  
Gandhi Maidan, Patna.*

Dear Sir or Madam:

Having examined the Bidding Documents, including Addenda Nos. [ insert **numbers** ], the receipt of which is hereby acknowledged, we, the undersigned, offer to supply and deliver the Goods under the above-named Contract in full conformity with the said Bidding Documents for the sum of Rs. 10,000/- (hereinafter called “the Total Bid Price”) or such other sums as may be determined in accordance with the terms and conditions of the Contract. The above amounts are in accordance with the Price Schedules attached herewith and are made part of this bid.

We undertake, if our bid is accepted, to deliver the Goods in accordance with the delivery schedule specified in the Schedule of Requirements.

If our bid is accepted, we undertake to provide an advance payment security and a performance security in the form, in the amounts, and within the times specified in the Bidding Documents.

We agree to abide by this bid, for the Bid Validity Period specified in Clause 18 of the ITB and it shall remain binding upon us and may be accepted by you at any time before the expiration of that period.

Until the formal final Contract is prepared and executed between us, this bid, together with your written acceptance of the bid and your notification of award, shall constitute a binding Contract between us. We understand that you are not bound to accept the lowest or any bid you may receive.

We undertake that, in competing for (and, if the award is made to us, in executing) the above contract, we will strictly observe the laws against fraud and corruption in force in India namely “Prevention of Corruption Act 1988”.

We confirm that we comply with the eligibility requirements as per ITB Clause 3 of the bidding documents.

We understand that you are not bound to accept the lowest or any bid you may receive.

Dated this [ insert: **number** ] day of [ insert: **month** ], [ insert: **year** ].

Signed: \_\_\_\_\_

Date: \_\_\_\_\_

In the capacity of [ insert: **title or position** ]

Duly authorized to sign this bid for and on behalf of [ insert: **name of Bidder** ]

<b>PRICE SCHEDULE</b>													
1	2	3	4	5							6	7	8
Sch No	Item Description	Country of origin	Quantity	Ex-factory Ex-warehouse Showroom off-shelf (A)	Excise duty if any (B)	Packing & Forwarding (C)	Inland transport, Insurance & incidental costs to Incidental delivery (D)	Incidental services as listed in GCC (E)	Customs Duty (F)	Unit Price A+B+C+D+E+F	Total Price per schedule for delivery at final destination (4X6)	Sales & Other taxes payable if contract is awarded	

Unit Price (6) ( Rs. In words)

AMC Charges (Labour only)

Equipment name	AMC CHARGES						
	4 <sup>TH</sup> YEAR	5 <sup>TH</sup> YEAR	6 <sup>TH</sup> YEAR	7 <sup>TH</sup> YEAR	8 <sup>TH</sup> YEAR	9 <sup>TH</sup> YEAR	10 <sup>TH</sup> YEAR
<b>TOTAL</b>							

CMC CHARGES

Equipment name	AMC CHARGES						
	4 <sup>TH</sup> YEAR	5 <sup>TH</sup> YEAR	6 <sup>TH</sup> YEAR	7 <sup>TH</sup> YEAR	8 <sup>TH</sup> YEAR	9 <sup>TH</sup> YEAR	10 <sup>TH</sup> YEAR
<b>TOTAL</b>							

Note:

- i. In case id discrepancy between unit price & total price Unit price shall prevail.
- ii. This price schedule should be placed in separate envelope sealed 'Cover B'

**Place  
Date**

**Signature of Bidder**.....  
**Name** .....  
**Address** .....

### **3. Earnest Money Deposit (EMD) Form**

**This has been deleted, EMD in the form of DD will be accepted**

To:  
Managing Director,  
Bihar Medical Services And Infrastructure Corporation Limited,  
Patna

WHEREAS [ *insert: name of Bidder* ] (hereinafter called “the Bidder”) has submitted its bid dated [ *insert: date of bid* ] for the performance of the above-named Contract (hereinafter called “the Bid”)

KNOW ALL PERSONS by these present that WE [ *insert: name of bank* ] of [ *insert: address of bank* ] (hereinafter called “the Bank”) are bound unto [ *insert: name of Purchaser* ] (hereinafter called “the Purchaser”) in the sum of: [ *insert: amount* ], for which payment well and truly to be made to the said Purchaser, the Bank binds itself, its successors and assigns by these presents.

Sealed with the Common Seal of the said Bank this [ *insert: number* ] day of [ *insert: month* ], [ *insert: year* ].

THE CONDITIONS of this obligation are the following:

1. If, after the bid submission deadline, the Bidder
  - (a) withdraws its bid during the period of bid validity specified by the Bidder in the Bid Form, or
  - (b) does not accept the Purchaser’s corrections of arithmetic errors in accordance with the Instructions to Bidders; or
2. If the Bidder, having been notified of the acceptance of its bid by the Purchaser during the period of bid validity
  - (a) fails or refuses to sign the Contract Agreement when required; or
  - (b) fails or refuses to issue the performance security in accordance with the Instructions to Bidders.

We undertake to pay to the Purchaser up to the above amount upon receipt of its first written demand, without the Purchaser having to substantiate its demand, provided that in its demand the Purchaser will note that the amount claimed by it is due it, owing to the occurrence of any one of the two above-named CONDITIONS, and specifying the occurred condition or conditions.

This guarantee will remain in full force up to and including [ *insert: the date that is 30 days after the period of bid validity* ], and any demand in respect thereof must reach the Bank not later than the above date.

For and on behalf of the Bank

Signed: \_\_\_\_\_

Date: \_\_\_\_\_

in the capacity of: [ *insert: title or other appropriate designation* ]

Common Seal of the Bank

## Form – 4 Form of Contract Agreement

THIS CONTRACT AGREEMENT is made the \_\_\_\_\_ day of \_\_\_\_\_ [month and year purchase] and between the Bihar Medical Services And Infrastructure Corporation Limited, Patna [Name of Purchaser] on behalf of Governor of Bihar (hereinafter referred to as the ‘Purchaser’) and \_\_\_\_\_ [ Name of Supplier ], having its principal place of business at \_\_\_\_\_ [ address of Supplier ] (hereinafter referred to as the “Supplier) on the other part.

WHEREAS the Purchaser invited bids for certain goods and ancillary services, viz., [insert: **brief description of goods and services**] and has accepted a bid by the Supplier for the supply of those goods and services in the sum of [ insert: **contract price in words and figures** ] (hereinafter called “the Contract Price”).

NOW THIS AGREEMENT WITNESSETH AS FOLLOWS:

1. In this Agreement words and expressions shall have the same meanings as are respectively assigned to them in the Conditions of Contract referred to.
2. The following documents shall constitute the Contract between the Purchaser and the Supplier, and each shall be read and construed as an integral part of the Contract:
  - (a) This Contract Agreement
  - (b) General Conditions of Contract
  - (c) Special Conditions of Contract
  - (d) Technical Requirements (including Functional Requirements and Implementation Schedule)
  - (e) The Supplier’s original Techno-commercial and Price bid
  - (f) The Schedule of Requirements
  - (g) The Purchaser’s Notification of Award
  - (h) [Add here: **any other documents**]
3. In consideration of the payments to be made by the Purchaser to the Supplier as hereinafter mentioned, the Supplier hereby covenants with the Purchaser to provide the Goods and Services and to remedy defects therein in conformity in all respects with the provisions of the Contract.
4. The Purchaser hereby covenants to pay the Supplier in consideration of the provision of the Goods and Services and the remedying of defects therein, the Contract Price or such other sum as may become payable under the provisions of the Contract at the times and in the manner prescribed by the Contract.

Brief particulars of the goods and services which shall be supplied/provided by the Supplier are as under:

SL. NO.	BRIEF DESCRIPTION OF GOODS	QUANTITY TO BE SUPPLIED	UNIT PRICE	TOTAL PRICE	DELIVERY TERMS
---------	----------------------------	-------------------------	------------	-------------	----------------

**TOTAL VALUE:**

**5. Payment:** That the 95% payment will be made by BMSICL within 30 days against supply, installation and commissioning of the equipment at the respective Institution and upon submission of satisfactory completion of installation certificate signed by the hospital authorities. The balance 5% payment will be released on submission of Bank Guarantee for the said amount with a validity to cover up the warranty period. In case, Bank Guarantee is not submitted by the “Supplier”, the said amount will be released after expiry of warranty period of three years.

**6. Post-Sales Service Centre & Contract persons of the supplier:**

Name & Address Detail \_\_\_\_\_

Contact persons Detail \_\_\_\_\_

**7. Liquidated Damages:**

If the successful tenderer fails to deliver any or all the goods or fails to perform the services within the time frame(s) prescribed in the contract, the Tender inviting the authority/User Institution under the contract, deduct from the contract price, as liquidated damages as per clause 20 under General conditions of contract.

**Delivery Schedule:**

For and on behalf of the Purchaser

Signed: \_\_\_\_\_  
in the capacity of [ *insert: title or other appropriate designation* ]

in the presence of \_\_\_\_\_

For and on behalf of the Supplier

Signed: \_\_\_\_\_  
in the capacity of [ *insert: title or other appropriate designation* ]

in the presence of \_\_\_\_\_

**CONTRACT AGREEMENT**

dated the [ *insert: number* ] day of [ *insert: month* ], [ *insert: year* ]

**BETWEEN**

Bihar Medical Services And Infrastructure Corporation Limited, “the Purchaser”

and

[ *insert: name of Supplier* ], “the Supplier”

## **5. Performance Security Bank Guarantee**

(Unconditional)

Date: *[insert: date]*

IFB: *[insert: name or number of IFB]*

Contract: *[insert: name or number of Contract]*

To:  
Managing Director,  
Bihar Medical Services And Infrastructure Corporation Limited,  
Patna

Dear Sir or Madam:

We refer to the Contract Agreement (“the Contract”) signed on *[insert: date]* between you and *[insert: name of Supplier]* (“the Supplier”) concerning the supply and delivery of *[insert: a brief description of the Goods]*. By this letter we, the undersigned, *[insert: name of bank]*, a bank (or company) organized under the laws of *insert: country of bank* and having its registered/principal office at *[insert: address of bank]*, (hereinafter, “the Bank”) do hereby jointly and severally with the Supplier irrevocably guarantee payment owed to you by the Supplier, pursuant to the Contract, up to the sum of *[insert: amount in numbers and words]*. This guarantee shall be reduced or expire as provided for by GCC Sub-Clause 5.4.

We undertake to make payment under this Letter of Guarantee upon receipt by us of your first written demand signed by your duly authorized officer declaring the Supplier to be in default under the Contract and without cavil or argument any sum or sums within the above-named limits, without your need to prove or show grounds or reasons for your demand and without the right of the Supplier to dispute or question such demand. Our liability under this Letter of Guarantee shall be to pay to you whichever is the lesser of the sum so requested or the amount then guaranteed under this Letter in respect of any demand duly made under this Letter prior to expiry of this Letter of Guarantee, without being entitled to inquire whether or not this payment is lawfully demanded.

This Letter of Guarantee shall be valid from the date of issue until the date of expiration of the guarantee, as governed by the Contract. Except for the documents herein specified, no other documents or other action shall be required, notwithstanding any applicable law or regulation. Our liability under this Letter of Guarantee shall become null and void immediately upon its expiry, whether it is returned or not, and no claim may be made under this Letter after such expiry or after the aggregate of the sums paid by us to you shall equal the sums guaranteed under this Letter, whichever is the earlier. All notices to be given under this Letter shall be given by registered (airmail) post to the addressee at the address herein set out or as otherwise advised by and between the parties hereto.

We hereby agree that any part of the Contract may be amended, renewed, extended, modified, compromised, released, or discharged by mutual agreement between you and the Supplier, and this security may be exchanged or surrendered without in any way impairing or affecting our liabilities hereunder without notice to us and without the necessity for any additional endorsement, consent, or guarantee by us, provided, however, that the sum guaranteed shall not be increased or decreased.

No action, event, or condition that by any applicable law should operate to discharge us from liability hereunder shall have any effect, and we hereby waive any right we may have to apply such law, so that in all respects our liability hereunder shall be irrevocable and, except as stated herein, unconditional in all respects.

For and on behalf of the Bank

Signed: \_\_\_\_\_

Date: \_\_\_\_\_

in the capacity of: [ insert: *title or other appropriate designation* ]

Common Seal of the Bank

## **6. Manufacturer's Authorization Form**

(Manufacturer's or Producer's letterhead)

To:

Managing Director,

Bihar Medical Services And Infrastructure Corporation Limited,

Patna

WHEREAS [ *name of the manufacturer or producer* ] (hereinafter, "we" or "us") who are established and reputable manufacturers or producers of [ *name and/or description of the Goods requiring this authorization* ] (hereinafter, "Goods") having production facilities at [ *insert: address of factory* ] do hereby authorize [ *name and address of Bidder* ] (hereinafter, the "Bidder") to submit a bid, and sign the Contract with you against IFB [ *title and reference number of the Invitation for Bids* ] including the above Goods produced by us.

We hereby extend our full guarantee and warranty for the above specified Goods against these Bidding Documents.

For and on behalf of the Manufacturer or Producer

Signed: \_\_\_\_\_

Date: \_\_\_\_\_

In the capacity of [ *title, position, or other appropriate designation* ] and duly authorize to sign this Authorization on behalf of [ *name of manufacturer or producer* ]

Note: This letter of authority should be on the letter head of the manufacturers and should be signed by a person competent and having the power of attorney to legally bind the manufacturer. This should be included by the bidder in his/her bid.

## **7. Proforma for performance statement**

(For a period of last three years)

Bid No: \_\_\_\_\_ Date of Opening: \_\_\_\_\_ Time : \_\_\_\_\_ Hours

Name of the Firm : \_\_\_\_\_

---

<u>Order Placed By</u> <u>(Full address of</u> <u>Purchaser)</u>	<u>Order No.</u> <u>and Date</u>	<u>Description and quantity</u> <u>of ordered Goods</u>	<u>Value of order</u>	<u>Date of completion of delivery</u>		<u>Remarks indicating</u> <u>reasons for late</u> <u>delivery, if any</u>	<u>Was the supply of Goods</u> <u>satisfactory ?</u> <u>(Attach a certificate from</u> <u>the Purchaser/Consignee)</u>
				<u>As per contract</u>	<u>Actual</u>		

---

Signature and seal of the Bidder \_\_\_\_\_

## 8. LETTER OF AUTHORISATION FOR ATTENDING BID OPENING

(To reach the Purchaser before date of bid opening )

To

Managing Director,  
Bihar Medical Services And Infrastructure Corporation Limited,  
Patna

Subject : Authorisation for attending bid opening on \_\_\_\_\_(date) in the Tender of  
\_\_\_\_\_.

Following persons are hereby authorised to attend the bid opening for the tender mentioned above on behalf of  
\_\_\_\_\_ (Bidder) in order of preference given below.

Order of Preference	Name	Specimen Signatures
---------------------	------	---------------------

I.

II.

Alternate  
Representative

Signatures of bidder

Or

Officer authorized to sign the bid  
Documents on behalf of the bidder.

Note : 1. Maximum of two representatives will be permitted to attend bid opening. In cases where it is restricted to one, first preference will be allowed. Alternate representative will be permitted when regular representatives are not able to attend.

2. Permission for entry to the hall where bids are opened, may be refused in case authorization as prescribed above is not recovered.

.....

## 9. CONSIGNEE RECEIPT CERTIFICATE/ Installation Report

(To be given by consignee and the user of the item)

The following equipments has / have been received in good condition:

Name of item supplied	
Name of the Supplier / Manufacturer	
Quantity supplied	
Purchase Order reference no.	
Serial Nos of equipment supplied	
Place of destination	
Name and Address of the Consignee along with tel. no. and fax no.	
Date of receipt by the Consignee	
Date of Installation	
Installation Location at Hospital.	
Accessories supplied and the serial numbers of Accessories	
Training satisfactorily completed Yes/No	
Name and Designation of Personnel trained.	
Date of commencement of warranty	
Date of expiry of warranty	
Stock Book page no. where the items have been entered	
Signature of Authorized Representative of Consignee with date	
Name and designation of the authorized representative	
Seal of the consignee	

Note: In case of Hospital the Incharge of the hospital concerned would be treated as consignee. In case of office (other than hospital), the office incharge of the office would be treated as consignee.

(Hospital / Office Incharge)

(User Department)



**FORMAT OF GENERAL GUARANTEE FOR WARRANTY**

(To be submitted on Firms Letterhead)

**Warranty Certificate**

Date:

We the Undersigned .....hereby guarantee satisfactory operation of ..... supplied by us to you against your purchase order No.....for a period of..... *calendar months* from the date of commissioning and shall be responsible for failure of the equipment to conform to the standard of performance, proficiency, production and / or out-turn stipulated or implicit in the order and for any defects that may develop under proper use arising from the use of faulty materials, design or workmanship in the supply made and shall remedy such defects at our cost.

for.....

Station : (Signature with Name and Designation)

Date :

Company Seal

## **SECTION VII CHECK LIST**

## CHECK LIST

Name of Tender, Name of Manufacturer

Sr. No.	Activity	Yes/No/N A	Page No. in the TE document	Remark
1a.	Have You Enclosed EMD of requirement amount for the quoted Schedules.			
b.	Have you enclosed tender document fee of requirement amount in case the tender document has been downloaded from websites? In case the tender document is purchased from BMSICL office. Have you enclosed the receipt?			
c	In case EMD is furnished in the form of Bank Guarantee, has been furnished as per Section VI.			
d.	In case Bank guarantee is furnished have you kept its validity of 195 days from Tender opening date as per clause 17.3 Of ITB?			
2a.	Have you enclosed dully filled Tender Form as per format in section VI?			
2b.	Have you enclosed Power of Attorney in favour of the signatory?			
3	Are you a SSI unit, if yes have you enclosed certificate of registration issued by Directorate of Industries /NSIC.			
4a	Have you enclosed clause-by-clause technical compliance statement for the quoted goods vis-a-vis the Technical specifications?			
b	In case of Technical deviations in the compliance statement, have you identified and marked the deviations?			
5a.	Have you submitted satisfactory performance certificate as per the proforma for performance statement in /sec. VI of TE document in respect of all orders?			
b.	Have you submitted copy of the order(s) and end user certificate to determine 80% supply?			
6	Have you submitted manufacturer's authorization as per Section VI?			
7	Have you submitted prices of goods turnkey (if any), CMC etc. in the Price Schedule as per Section VI?			
8	Have you kept validity of 150 days from the Techno Commercial Tender Opening date as per the TE document?			
9a.	In case of Indian Tenderer, have you furnished Income Tax Account No. as allotted by Income tax Department of Government of India?			
b	In case of Foreign Tenderer, have you furnished Income Tax Account No. of your Indian Agent as allotted by the Income Tax Department of			

	Government of India?			
10.	Have you intimated the name and full address of your Banker (s) along with your Account Number			
11.	Have you fully accepted payment terms as per TE document?			
12.	Have you fully accepted delivery period as per TE document?			
13.	Have you submitted the certificate of Incorporation?			
14.	Have you accepted the warranty as per TE document?			
15.	Have you accepted terms and conditions of TE document?			
16.	Have you furnished documents establishing your eligibility & qualification criteria as per TE documents?			
17.	Have you furnished Annual Report (Balance Sheet and Profit & Loss Account) for last three years prior to the date of Tender Opening?			

- 1 All pages of the Tender should be page numbered and indexed.
2. The Tenderer may go through the checklist and ensure that all the documents/ confirmations listed above are enclosed in the tender and no column is left blank. If any column is not applicable, it may be filled up as N.A.
3. It is the responsibility of tenderer to go through the TE document to ensure furnishing all required documents in addition to above, if any.

.....  
Signature with date)

(Full name, designation & address of the person duly authorized sign on behalf of the Tenderer)

For and on behalf of

.....  
(Name address and stamp of the tendering firm)