



**Bihar Medical Services & Infrastructure Corporation
Limited**

**2nd & 3rd Floor, Swasthya Bhawan, Behind IGIMS,
Sheikhpura, Adjacent to State Health Society, Bihar,
Patna- 800014**

Phone/Fax: +919471009074

CORRIGENDUM-I

Tender for rate contract and supply of Medical Devices/Consumables for different healthcare facilities of state of Bihar

Notice Inviting Tender Ref No.: - BMSIC/MEDICAL DEVICE/CONSUMABLES/23-05

Dated: - 17/04/2023

(Only through E- Tender on website: -www.eproc2.bihar.gov.in)

Bihar Medical Services and Infrastructure Corporation Limited (BMSICL) has invited E-Bids from the interested parties for “**Tender for rate contract and supply of Medical Devices/Consumables for different healthcare facilities of state of Bihar**”, vide Notice Inviting Tender No.- **BMSIC/MEDICAL DEVICE/CONSUMABLES/23-05**. Detailed tender document containing eligibility criteria, selection mechanism, other terms and conditions is available on the website www.eproc2.bihar.gov.in.

After considering the suggestion/queries received from the prospective bidders & after due deliberation on all aspects, certain amendments have been made in the product specification as per Annexure A & Annexure B from the tendered product list as mentioned in this Corrigendum-I.

In order to ensure wider participation of the bidders the tender schedule is also being revised as follows: -

Revised Tender Schedule

Tender Reference No.	BMSIC/MEDICAL DEVICE/CONSUMABLES/23-05
Last date and time of submission of online bids	Upto 03rd May 2023 by 18:00 Hrs.
Last date and time for submission of EMD and Tender Fees	Upto 04th May 2023 till 14:00 Hrs.
Date, Time and Place of opening of Technical Bid	Upto 04th May 2023 (at 15:00 Hrs.) on the website of www.eproc2.bihar.gov.in in the office of BMSICL
Date and time of opening of Financial Bids	To be announced later on www.bmsicl.gov.in and www.eproc2.bihar.gov.in
Validity of Tender	180 Days
Cost of the tender document	Rs. 11,800/- (Eleven Thousand Eight Hundred only) Non-refundable.

**Sd/-
GM (Procurement)
BMSICL**

Annexure-A
Specification of Bed Side Filter Set for Thalassemia Patients

S. N.	Specifications
1.	The pre deposit storage leukodepletion filter for the leukodepletion of packed red cells.
2.	Filtration of red cells must be completed for >95% of bags within 45 minutes from time at which flow of blood in to the filter is opened
3.	The filter need to be able to reduce the final count of leukocytes in the product to $<5 \times 10^4$ per bag
4.	Recovery: The Filtration process need to not reduce red cell to less than 90% of the initial Red cell mass. Percentage of haemolysis $<1\%$
5.	Usable with blood of core temperatures in the range $4^{\circ}\text{C} - 30^{\circ}\text{C}$
6.	Need to be a hard filter
7.	Filter Material need to be surface modified polyester material to Ensure quality of red cell during filtration.
8.	Filter Housing: Material need to be polycarbonate with hold up Volume of max 25 ml
9.	Sterilization method need to be Ethylene oxide gas
10.	Air vent: Need to have sterile air venting Elimination system in the filter housing. The Device need to be with a by-pass line and one-way valve to remove air inside the Bag.
11.	Transfer bag need to be attached and have ports
12.	Mechanism of Action; selective adsorption or adhesion of Leukocytes
13.	Need to have a sampling pouch for air removal and QC sampling
13.	CE marked/USFDA approved

Annexure-B
Specification of Leukodepleted Triple Blood Bag with SAGM

S.N.	Specifications
1.	Log 4 Leukodepleted Triple Blood bag with SAGM and without any automatic components extractor.
2.	Easy to use during processing.
3.	Log 4: 99.99 % Leukodepletion.
4.	Optimal recovery: $\geq 90\%$
5.	Short filtration time: 20-30 minutes
6.	Residual WBC: $\leq 5 \times 10^6$
7.	100% close system with complete sterility
8.	<u>Each triple bag must consists of:</u> a) Pre-donation Sampling Bag b) Primary Collection Bag c) MIRC Leukocyte removal filter for RBC d) Leukodepleted RBC bag with SAGM e) Platelet Bag with TOTM sheet f) Plasma Bag/Transfer Bag with SAGM

Sd/-
GM (Procurement)
BMSICL

-END-