



**Bihar Medical Services & Infrastructure Corporation
Limited**

**2nd & 3rd Floor, Swasthya Bhawan, Behind IGIMS,
Sheikhpura, Adjacent to State Health Society, Bihar,
Patna- 800014**

Phone/Fax: +919006405255, + 919471009193

CORRIGENDUM-I

**Tender for Rate Contract & Supply of Glucometer Machine, Glucose Test Strips and Auto
Disable Lancets for the State of Bihar**

Notice Inviting Tender Ref No.: -BMSIC/MEDICAL DEVICES/CONSUMABLES/24-04

Dated: - 21-06-2024

(Only through E- Tender on website: -www.eproc2.bihar.gov.in)

Bihar Medical Services and Infrastructure Corporation Limited (BMSICL) invites E-Bids from the interested parties for “**Tender for Rate Contract & Supply of Glucometer Machine, Glucose Test Strips and Auto Disable Lancets for the State of Bihar**”, vide Notice Inviting Tender No.- **BMSIC/MEDICAL DEVICES/CONSUMABLES/24-04**. Detailed tender document containing eligibility criteria, selection mechanism, other terms and conditions are available on the website www.eproc2.bihar.gov.in.

After considering the suggestion/queries received from the prospective bidders & after due deliberation on all aspects, certain amendment has been made in the Technical Specification in Annexure VII of the bid document of this Corrigendum-I. Rest of the terms and condition shall remain unchanged.

Revised Tender Schedule

Tender Reference No.	BMSIC/MEDICAL DEVICES/CONSUMABLES/24-04
Last date and time of submission of online bids	08th July 2024 by 17:00 Hrs.
Last date and time for submission of EMD and Tender Fees	10th July 2024 till 14:00 Hrs.
Date, Time and Place of opening of Technical Bid	10th July 2024 (at 15:00 Hrs.) on the website of www.eproc2.bihar.gov.in in the office of BMSICL
Date and time of opening of Financial Bids	To be announced later on www.bmsicl.gov.in and www.eproc2.bihar.gov.in
Validity of Tender	180 Days
Cost of the tender document	Rs. 11,800/- (Eleven Thousand Eight Hundred only) Non-refundable.

**Sd/-
GM (Procurement)
BMSICL**

ANNEXURE-VII
TECHNICAL SPECIFICATIONS

GLUCOMETER WITH STRIPS

Product Eligibility Criteria:

- Product should be CE as per IVD or USFDA Certified.
- Manufacturer should be ISO 13485 certified for quality standards.
- The device and test strips should be certified by the National Institutes of Biologicals (NIB), Govt. of India.

Technical Specifications

1. Small, portable and user friendly device is required. Blood should not go into the glucometer while measurement.
2. It should be able to measure whole blood in capillary mode.
3. Measurement range: 30 to 600 in mg/dl.
4. Accuracy should be as per International Standard ISO 15197: 2013 (Requirements for Blood-glucose monitoring systems for self-testing in managing diabetes mellitus). Supporting certificate or test reports from the National Institutes of Biologicals (NIB) must be furnished of last 2years with the technical bid.
5. Reproducibility/Precision: $\pm 5\%$
6. Display should be 40mm ± 5 mm or better measured diagonally.
7. It should be battery operated electronic system and the battery life should be for at least 500 tests.
8. Shelf life of strips: Minimum 18 months at the time of delivery to consignee.
9. Packing of strips should not be more than 100 strips in a pack.
10. Strips should work till expiry after opening of strips pack.
11. Operating temperature for both device and test strip should be 10⁰C to 50⁰C.
12. Control solution for checking reliability of strips will be supplied free of cost as & when required.
13. Ready availability of reagent test strips, battery & other consumables across BIHAR for at least 5 years.
14. A complete user operational guide shall have to be supplied along with each machine, printed in Hindi and English language.
15. A lancet applicator/ lancet holder shall have to be supplied along with each machine.

Scope of supply:

- a) Glucometer: 1no.
- b) Standard batteries: 1Set
- c) Carrying case: 1no.
- d) Control solution/Control Strips
- e) Glucose test strips: As per Order
- f) Auto disables lancets: As per order

Sd/-
GM (Procurement)
BMSICL