



**Bihar Medical Services & Infrastructure Corporation  
Limited**  
4<sup>th</sup> floor State Building Construction Corporation Limited.  
Hospital Road, Shastri Nagar, Patna 800023  
Phone/Fax: +917008050665, + 919471009193

### **CORRIGENDUM-II**

## **Tender for rate contract and supply of Medical Devices & Consumables for different healthcare facilities of state of Bihar for year 2020-2022**

**Notice Inviting Tender Ref No.:- BMSIC/MEDICAL DEVICES/CONSUMABLES/20-02**

**Dated:- 10-10-2020**

(Only through E- Tender on website: -[www.eproc.bihar.gov.in](http://www.eproc.bihar.gov.in))

Bihar Medical Services and Infrastructure Corporation Limited (BMSICL) has invited e-Bids from the interested parties for “**Tender for rate contract and supply of Medical Devices & Consumables for different healthcare facilities of state of Bihar for year 2020-2022**”, vide **Notice Inviting Tender No.- BMSIC/MEDICAL DEVICES/CONSUMABLES/20-02**. Detailed tender document containing eligibility criteria, selection mechanism, other terms and conditions are available on the website [www.eproc.bihar.gov.in](http://www.eproc.bihar.gov.in).

After considering the suggestions/ queries received from the prospective bidders during and after Pre-Bid meeting the amendments with regard to the product list have been made as per Annexure-A (Revised Tendered Product List). All other technical specifications and terms conditions remain unchanged.

**Since, any change text in the price BOQ can't be changed in the e-proc portal without its replacement, the amendments mentioned in the specifications are to be treated as amendments in the price bid/ BOQ also.**

In order to ensure wider participation, the tender schedule is being revised as follows: -

Tender Reference No.	<b>BMSIC/MEDICAL DEVICES/CONSUMABLES/20-02</b>
Date and time for downloading of bid document	<b>Upto 27<sup>th</sup> October 2020 till 1500 Hrs.</b>
Last date and time of submission of online bids	<b>Upto 30<sup>th</sup> October 2020 till 1500 Hrs.</b>
Last date and time for submission of original bid documents with EMD and Tender Fees	<b>02<sup>nd</sup> November 2020 by 1400 Hrs.</b>
Date, Time and Place of opening of Technical Bid	<b>02<sup>nd</sup> November 2020 till 1500 Hrs.</b>
Date and time of opening of Financial Bids	<b>To be announced later on <a href="http://www.bmsicl.gov.in">www.bmsicl.gov.in</a> and <a href="http://www.eproc.bihar.gov.in">www.eproc.bihar.gov.in</a></b>
Validity of Tender	<b>180 Days</b>
Cost of the tender document	<b>Rs. 10,000/- (Ten Thousand only) Non-refundable.</b>
Bid Processing Fee	<b>Rs 1180/- (One thousand one hundred eighty only) Non-refundable.</b>

**Sd/-  
GM (Procurement)  
BMSICL**

<b>ANNEXURE- A (Revised Tendered Product List)</b>					
<b>BMSIC/MEDICAL DEVICES/CONSUMABLE/20-02</b>					
<b>S. N.</b>	<b>Items</b>	<b>Floated Specification</b>	<b>Pack Size</b>	<b>Estimated Tendered Quantity For Year 2020-2022</b>	<b>Amended Specification</b>
1.	Absorbent gauze	Absorbent Gauze made of 100% absorbent cotton, mainly used as the raw material of basic gauze dressings, such as gauze sponges and lap sponges etc. Available in zig-zag fold (May be Sterile/Non Sterile) 20 mtr x 90 cms (Wt.540gms plus or minus 90gms) Than packed in poly bag with label (As per Schedule: F)	Per Piece	600000	<b>Absorbent Gauze Size: - 20 mtr x 90 cms (Must comply as per Schedule: F II of Drugs &amp; Cosmetics Rules)</b>
2.	Adhesive Tape	Adhesive Tape (such as paper, which tears easily) or tear-resistant (often fabric or plastic) may be either waterproof or breathable or both water-resistant and partly-porous, allowing some limited passage of air and moisture. should have appropriate adhesive power, breathable and at the same time impermeable, easy to remove without irritating the skin, resistant to extreme temperatures, aging and solar radiation. Roll Size 5 cm X 10 Metre	Per Piece	321000	<b>No Change</b>
3.	Adhesive Tape	Adhesive Tape, 7.5cm x 10 metre Adhesive Tape (such as paper, which tears easily) or tear-resistant (often fabric or plastic) may be either waterproof or breathable or both water-resistant and partly-porous, allowing some limited passage of air and moisture. should have appropriate adhesive power, breathable and at the same time impermeable, easy to remove without irritating the skin, resistant to extreme temperatures, aging and solar radiation. Roll Size 7.5 cm X 10 Metre	Per Piece	287000	<b>No Change</b>
4.	Bandage Cloth	20 mtr x 90cms (Wt. 1026 gms plus or minus 90gms) than packed in poly bag with label (As per Schedule: F)	Per Piece	327000	<b>Bandage Cloth Size: - 20 mtr x 90cms (Must comply as per Schedule: F II of Drugs &amp; Cosmetics Rules)</b>

5.	Disposable Syringe 2 ml	Disposable Syringe 2 ml, With colour coded (as per BIS) needle Sterilised, Luer Mount, Non - toxic CGS 2CC as per Drugs and Cosmetics Act 1940, IS No. 10258-2102 with CE certification	Per Piece	2100000	<b>Disposable Syringe- 2 ml, with colour coded (as per BIS) needle Sterilised, Luer Mount, Non - toxic as per Drugs and Cosmetics Act 1940, IS No. 10258-2002 with BIS/ CE/ USFDA certification</b>
6.	Disposable Syringe 5 ml	Disposable Syringe 5 ml, With colour coded (as per BIS) needle Sterilised, Luer Mount, Non - toxic CGS 2CC as per Drugs and Cosmetics Act 1940, IS No. 10258-2102 with CE certification	Per Piece	2710000	<b>Disposable Syringe- 5 ml, with colour coded (as per BIS) needle Sterilised, Luer Mount, Non - toxic as per Drugs and Cosmetics Act 1940, IS No. 10258-2002 with BIS/ CE/ USFDA certification</b>
7.	Disposable Syringe 10 ml	Disposable Syringe 10 ml, With colour coded (as per BIS) needle Sterilised, Luer Mount, Non - toxic CGS 2CC as per Drugs and Cosmetics Act 1940, IS No. 10258-2102 with CE certification.	Per Piece	1050000	<b>Disposable Syringe- 10 ml, with colour coded (as per BIS) needle Sterilised, Luer Mount, Non - toxic as per Drugs and Cosmetics Act 1940, IS No. 10258-2002 with BIS/ CE/ USFDA certification</b>
8.	Elastic Adhesive Bandage	Elastic Adhesive Bandage- 10 cm x 2.70 metre Elastic Adhesive Bandage B.P. consists of a woven fabric, elastic in warp, which has been spread evenly with an adhesive mass containing Zinc Oxide, which does not offset when the bandage is unrolled. The warp threads consist of two-fold cotton threads crepe twisted and welt threads are of cotton/viscose /both mixed. The warp threads are arranged two threads S twisted and two threads Z twist repeated. It is one continuous length without joints and has fast edges.	Per Piece	95000	<b>No Change</b>
9.	Rolled Bandage	Rolled Bandage – Size: - 5mt x 5cm, Bleached, purified textile, plain weave Size Rolled Bandage made of 100% Cotton Thread count: Warp: 12 threads/cm Weft: 8 threads/cm Weight: +/- 27.5g/m <sup>2</sup> Components: selected: Bandage, gauze roll.	Per Piece	305000	<b>Rolled Bandage (Bleached) Size: - 5mt x 5cm (Must comply as per Schedule: F II of Drugs &amp; Cosmetics Rules)</b>

		Width: 5cm. Length: 5m. Single use Non-sterile.			
10	Standard Wire Gauze	As per License (Per Piece)	Per Piece	40000	No Change
11.	Adhesive Paper Tape	Adhesive Paper Tape 1.25 Adhesive Paper tape manufacture by coating uniformly on one side of paper with a water soluble adhesive composition. Comply as per IS 4185:1989	Per Piece	252000	Adhesive Paper tape manufacture by coating uniformly on one side of paper with a water-soluble adhesive composition. Size: - 1.25 cm x 9 metre
12.	Hydrogen Peroxide Solution	Hydrogen Peroxide Solution 6% (100ml Bottle)	100ml Bottle	80000	No Change
13.	Intracatheter 18	Size 18 Sterile, Disposable, G 18 with wings & port	Per Piece	294000	Deleted
14.	Intracatheter 20	Size 20 Sterile, Disposable, G 20 with wings & port	Per Piece	630000	
15.	Intracatheter 24	Size 24 Sterile, Disposable, G 24 with wings & port	Per Piece	525000	
16.	Water Seal bag	(Water Seal bag of 1litre Capacity) Under water seal drainage system for collection of drainage fluid from thoracic cavity. Super smooth kink resistance tube provided with universal tapered connector. PVC drainage bag with 1000 ml capacity suitable for most thoracic applications. Specially designed moulded handle for easy carrying and hanging of the bag. Transparent sheeting allows visual inspection. Available with DEHP Free Material. Sterile / Disposable / Individually Packed	Per Piece	7700	No Change
17.	EC Lotion	(Chlorinated Lime Water) 30 % Chlorinated Lime Water (1000 ml Bottle)	1000ml Bottle	104600	No Change
18.	A.D Syringe 50 cc	(1. Clear transparent chamber, Prominent graduation 2. Inert material gasket at the piston to minimize friction during movement & prevent leakage and back flow. The words "DESTROY AFTER SINGLE USE" or equivalent should be written on Unit Container. Shall conform to IS 12050 Use-50 ml)	Per Piece	165000	(1. Clear transparent chamber, Prominent graduation 2. Inert material gasket at the piston to minimize friction during movement & prevent leakage and back flow. The words "DESTROY AFTER SINGLE USE" or equivalent

					should be written on Unit Container. Shall conform to <b>ISO:7886-4</b> Use-50 ml)
19.	A.D Syringe 20 cc	(1. Clear transparent chamber, Prominent graduation 2. Inert material gasket at the piston to minimize friction during movement & prevent leakage and back flow the words "DESTROY AFTER SINGLE USE" or equivalent should be written on Unit Container. Shall conform to IS 12050 Use-20 ml)	Per Piece	635500	(1. Clear transparent chamber, Prominent graduation 2. Inert material gasket at the piston to minimize friction during movement & prevent leakage and back flow the words "DESTROY AFTER SINGLE USE" or equivalent should be written on Unit Container. Shall conform to <b>ISO:7886-4</b> Use-20 ml)
20.	T-Tube 12	(Tracheostomy Tube, made from medical grade PVC, Plain, Sterile, Single Use - Size 12 Soft flexible flange at for easy fixation 15 mm connector at terminal end which can be rotated in 360-degree direction Non-irritant, Radio-opaque line throughout the length. Size: 20 X 60cm)	Per Piece	19200	<b>No Change</b>
21.	T-Tube 24	(Tracheostomy Tube, made from medical grade PVC, Plain, Sterile, Single Use - Size 24 Soft flexible flange at for easy fixation 15 mm connector at terminal end which can be rotated in 360-degree direction Non-irritant, Radio-opaque line throughout the length. Size: 20 X60 mm)	Per Piece	18100	<b>No Change</b>
22.	Skeletal Traction Kit	(Complete kit with H. Sq. Traction pulley, Foam padding, Spreader, Cord, Crepe Bandage and optional water/sand weight bag).	Per Piece	7600	<b>No Change</b>
23.	Central Venous Catheter Central Line-Triluminal, Adult Use	<b>Central Venous Catheter, Triluminal, Adult Use</b> , Radio-opaque polyurethane catheter, with soft tip, length markings, junction hub for stitch fixation and fully transparent extension lines, with color-coded hub and white slide clamps. Components: Kink-proof guide wire with "J" tip and straight tip on dispenser, movable fixation wing with clamp, introducer needle, scalpel, dilator, valve, 5 ml	Per Piece	400	<b>No Change</b>

		luer lock syringe (depending on set configuration)			
24.	Dressing Gauze (Sterile)	7.5 cm X 7.5 cm, 100% cotton, Bleached, purified textile, plain weave. Absorbent gauze, 100% cotton. Woven. Type of gauze: 17 threads/cm <sup>2</sup> (23g/m <sup>2</sup> ). sterile compresses, sterile peel pack, single-use	Per Piece	598200	Dressing Gauze (Sterile) Size: - 7.5 cm X 7.5 cm <b>(Must comply as per Schedule: F II of Drugs &amp; Cosmetics Rules)</b>
25.	Plaster of Paris	Plaster of Paris (Powder) Pack Size: Per Kg	Per Kg	183800	<b>No Change</b>

**Note:-I- Product standards for medical device:—**

(1) The medical device shall conform to the standards laid down by the Bureau of Indian Standards established under section 3 of the Bureau of Indian Standards Act, 1985 (63 of 1985) or as may be notified by the Ministry of Health and Family Welfare in the Central Government, from time to time.

(2) Where no relevant Standard of any medical device has been laid down under sub-rule (1), such device shall conform to the standard laid down by the International Organisation for Standardisation (ISO) or the International Electro Technical Commission (IEC), or by any other pharmacopoeial standards.

(3) In case of the standards which have not been specified under sub-rule (1) and sub-rule (2), the device shall conform to the validated manufacturer's standards.

**Note:-II- The Regulations of Medical Device Rule 2017 is applicable accordingly.**