

FLEXIBLE NASO-PHARYNGO LARYNGOSCOPE

TECHNICAL SPECIFICATIONS

1. Should have a field of view of at least 75°.
2. Should have a depth of field from 5 to 50 mm
3. The insertion tube should have maximum 3.5mm diameter or less.
4. Should have at least 130° upwards and 130° downwards angulations.
5. Should have a working length of at least 300 mm.
6. Should have a light guide illuminating system.
7. Should provide suitable light source.
8. Should be supplied with all standard accessories including storage box and list of Standard accessories should be specified in the technical bid.

LIGHT SOURCE:

1. Should be a halogen light source with minimum 150W light output with 5 spares.
2. Should have manual light intensity control.
3. Should have cooling system.
4. Should work with input 200 to 240Vac 50 Hz supply.
5. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.