

**MINUTES OF TECHNICAL EVALUATION COMMITTEE MEETING FOR TENDER NO.-
BMSIC/MEDICAL DEVICES/CONSUMABLES/21-01 (TENDER FOR RATE CONTRACT AND
SUPPLY OF MEDIAL DEVICES/ CONSUMABLES FOR DIFFERENT HEALTHCARE FACILITIES
OF STATE OF BIHAR).**

Date: 24th May 2021

Venue: BMSICL Conference Hall.

The Technical Evaluation Committee Meeting for Tender No. BMSIC/MEDICAL DEVICES/CONSUMABLES/21-01 was held on 24th May 2021 under the Chairmanship of Dr. Navin Chandra Prasad, Director in Chief, Department of Health, GoB. The other committee members present in the meeting were:

1. Sh. Rajani Kant, CGM (Supply Chain), BMSICL.
2. Sh. Sudhir Kumar, GM (Procurement), BMSICL.
3. Sh. Suman Prasad Sah, Deputy Secretary-cum-I/C Procurement, SHSB.
4. Sh. Ramendra Kumar, OSD (Drug & Equipment), BMSICL.
5. Sh. Navindra Kumar, I/c DGM (Drugs), BMSICL.

2. The Details of the invited tender are as follows:

1. Tender Reference No. BMSIC/MEDICAL DEVICES/CONSUMABLES/21-01
2. Tender floated date- 19/02/2021.
3. Last date of submission of Online bids- 19/03/2021.

3. The Tender has been floated on e-proc Portal for 30 (Thirty) MEDICAL DEVICES/CONSUMABLES.

4. Total 22 bidders were found to have successfully submitted their bids on the e-proc online system. The submission status of hard copy of the technical bids including the tender fee and Earnest Money Deposit (EMD), till the stipulated time was found as following: -

| Sl No. | Name of the Firm | Online submission status of bids on E-Proc | Requisite Tender fee of Rs 10,000/- receipt Status | Requisite Amount of EMD receipt Status | Remarks |
|--------|--------------------------------------|--|--|--|-----------|
| 1 | Zenith Infratech | Qualified | Received | Received | Qualified |
| 2 | Unisur Lifecare Pvt. Ltd. | Qualified | Received | Received | Qualified |
| 3 | Futura Surgicare Pvt. Ltd. | Qualified | Received | Received | Qualified |
| 4 | Healthium Medtech Pvt. Ltd. | Qualified | Received | Received | Qualified |
| 5 | Hi-Tech Medicare Devices Pvt. Ltd. | Qualified | Received | Received | Qualified |
| 6 | Lotus Surgicals Pvt. Ltd. | Qualified | Received | Received | Qualified |
| 7 | Mais India Medical Devices Pvt. Ltd. | Qualified | Received | Received | Qualified |
| 8 | MCo Hospital Aids Pvt. Ltd. | Qualified | Received | Received | Qualified |
| 9 | Mediplus (India) Limited | Qualified | Received | Received | Qualified |
| 10 | Mehr Enterprises | Qualified | Received | Received | Qualified |
| 11 | Meril Endo Surgery Aids Pvt. Ltd. | Qualified | Received | Received | Qualified |
| 12 | MRK Healthcare Pvt. Ltd. | Qualified | Received | Received | Qualified |
| 13 | Om Surgical Industries | Qualified | Received | Received | Qualified |

| | | | | | |
|----|-----------------------------------|-----------|--------------|--------------|--------------|
| 14 | Peters Surgical India Pvt. Ltd. | Qualified | Received | Received | Qualified |
| 15 | Plasti Surge Industries Pvt. Ltd. | Qualified | Received | Received | Qualified |
| 16 | Anondita Healthcare | Qualified | Received | Received | Qualified |
| 17 | Paramount Surgimed Ltd. | Qualified | Received | Received | Qualified |
| 18 | Romsons Group Pvt. Ltd. | Qualified | Received | Received | Qualified |
| 19 | Sanmati Packaging | Qualified | Received | Received | Qualified |
| 20 | Stoe Exim | Qualified | Received | Received | Qualified |
| 21 | BL Lifesciences Pvt. Ltd. | Qualified | Received | Received | Qualified |
| 22 | Tradition Globe Trade India | Qualified | Received | Received | Qualified |
| 23 | Poly Medicure Ltd. | Null | Not Received | Not Received | Disqualified |

5. The firm M/s Poly Medicure Ltd. neither submitted the hardcopy nor considered by the online e-proc system. Accordingly, it was unanimously decided by the Technical Committee in its meeting dated 23.03.2021 to recommend for considering only those above mentioned 22 bidders (Sl. No. 1 to 22) who have successfully submitted their bids online as well as Earnest Money Deposit (EMD) & Tender fee, for further technical evaluation process.

6. The MEDICAL DEVICES/CONSUMABLES wise bid received status was found as follows:

| S.N. | Name of the Product with Specification | Participating Firms | No. of Bids |
|------|---|--|-------------|
| 1 | Black Braided Silk 1/0 (Sterile, Length 2 X 75 cm (Pre-Cut & Pre-Sterilized) with CE certification). | Unisur Lifecare Private Limited Futura Surgicare Pvt. Ltd. Healthium Medtech Pvt. Ltd. Lotus Surgicals Pvt. Ltd. MCo Hospital Aids Pvt. Ltd. Meril Endo Surgery Private Limited | 6 |
| 2 | Black Braided Silk 1 (Sterile, Length 2 X 75 cm (Pre-Cut & Pre-Sterilized) with CE certification). | Unisur Lifecare Private Limited Futura Surgicare Pvt. Ltd. Healthium Medtech Pvt. Ltd. Lotus Surgicals Pvt. Ltd. MCo Hospital Aids Pvt. Ltd. Meril Endo Surgery Private Limited | 6 |
| 3 | Cap (Surgical Cap, Disposable Should be manufactured from non-woven fabric Blue/Green/ Colour Round upon wearing, with elastic Air ermeable/ Breathable should retain skin and hair particles). | MEHR Enterprises Om Surgical Industries Plasti Surge Industries Pvt. Ltd. Sanmati Packaging Tradition Globe Trade India | 5 |
| 4 | Catgut Plane with Needle Size "0" (Absorbable needled suture 152 cm ,1/2 circle round body with needle 30 mm). | Unisur Lifecare Private Limited Futura Surgicare Pvt. Ltd. Lotus Surgicals Pvt. Ltd. | 5 |

| | | | |
|---|--|------------------------------------|---|
| | | MCo Hospital Aids Pvt. Ltd. | |
| | | Meril Endo Surgery Private Limited | |
| 5 | Catgut Chromic with needle 1-0 (Absorbable Length 70-90 cm, 1/2 circle round body with 34-40 mm needle). | Unisur Lifecare Private Limited | 7 |
| | | Futura Surgicare Pvt. Ltd. | |
| | | Healthium Medtech Pvt. Ltd. | |
| | | Lotus Surgicals Pvt. Ltd. | |
| | | MCo Hospital Aids Pvt. Ltd. | |
| | | Meril Endo Surgery Private Limited | |
| | | Peters Surgical India Pvt. Ltd. | |
| 6 | Catgut Chromic with needle 2-0 (Absorbable Length 70-90 cm, 1/2 circle round body with 30 mm needle). | Unisur Lifecare Private Limited | 7 |
| | | Futura Surgicare Pvt. Ltd. | |
| | | Healthium Medtech Pvt. Ltd. | |
| | | Lotus Surgicals Pvt. Ltd. | |
| | | MCo Hospital Aids Pvt. Ltd. | |
| | | Meril Endo Surgery Private Limited | |
| | | Peters Surgical India Pvt. Ltd. | |
| 7 | Catgut Chromic with needle 3-0 (Absorbable Length 70-90 cm, 1/2 circle round body with 20 mm needle) | Unisur Lifecare Private Limited | 7 |
| | | Futura Surgicare Pvt. Ltd. | |
| | | Healthium Medtech Pvt. Ltd. | |
| | | Lotus Surgicals Pvt. Ltd. | |
| | | MCo Hospital Aids Pvt. Ltd. | |
| | | Meril Endo Surgery Private Limited | |
| | | Peters Surgical India Pvt. Ltd. | |
| 8 | Catgut Chromic with needle 1 (Absorbable Length 70-90 cm, 1/2 circular round body with 34-40 mm needle). | Unisur Lifecare Private Limited | 7 |
| | | Futura Surgicare Pvt. Ltd. | |
| | | Healthium Medtech Pvt. Ltd. | |
| | | Lotus Surgicals Pvt. Ltd. | |
| | | MCo Hospital Aids Pvt. Ltd. | |
| | | Meril Endo Surgery Private Limited | |
| | | Peters Surgical India Pvt. Ltd. | |

| | | | |
|----|--|--------------------------------------|---|
| 9 | <p>Central Venous Cathetar Central Line-Monoluminal, Adult use Central Venous Catheter, (monoluminal, Adult use, Radio-opaque polyurethane catheter, with soft tip, length markings, junction hub for stitch fixation and fully transparent extension lines, with colorcoded hub and white slide clamps. Components: Kink-proof guide wire with “J” tip and straight tip on dispenser, movable fixation wing with clamp, introducer needle,scalpel, dilator, valve, 5 ml luer lock syringe) (depending on set configuration).</p> | BL Lifesciences Private limited | 1 |
| 10 | <p>Central Venous Cathetar Central Line-Biluminal,Adult use (Central Venous Catheter biluminal, Adult use, Radio-opaque polyurethane catheter, with soft tip, length markings, junction hub for stitch fixation and fully transparent extension lines, with color-coded hub and white slide clamps. Components: Kink-proof guide wire with “J” tip and straight tip on dispenser, movable fixation wing with clamp, introducer needle, scalpel, dilator, valve, 5 ml luer lock syringe) (depending on set configuration).</p> | Romsons Group Private Limited | 2 |
| | | BL Lifesciences Private limited | |
| 11 | <p>Central Venous Cathetar Central Line-Triluminal,Adult Use (Triluminal, Adult Use, Radio-opaque polyurethane catheter, with soft tip, length markings, junction hub for stitch fixation and fully transparent extension lines, with color-coded hub and white slide clamps. Components: Kink-proof guide wire with “J” tip and straight tip on dispenser, movable fixation wing with clamp, introducer needle,scalpel, dilator, valve, 5 ml luer lock syringe) (depending on set configuration)</p> | Romsons Group Private Limited | 2 |
| | | BL Lifesciences Private limited | |
| 12 | <p>Central Venous Cathetar Central Line-Biluminal,Paediatric (Central Venous Catheter biluminal, Paediatric Use, Radio-opaque polyurethane catheter, with soft tip, length markings, junction hub for stitch fixation and fully transparent extension lines, with colorcoded hub and white slide clamps.Components: Kink-proof guide wire with “J” tip and straight tip on dispenser, movable fixation wing with clamp, introducer needle, scalpel, dilator, valve, 5 ml luer lock syringe) (depending on set configuration).</p> | BL Lifesciences Private limited | 1 |
| 13 | <p>IV Canula 24 (Sterile Disposable (single Use) Teflon / PTFE IV Cannula with injection with wing port. Size 24 . Should be pack in Transparent, single blister pack. Should conform to IS 10555 Standard).</p> | Hi-Tech Medicare Devices Pvt. Ltd. | 4 |
| | | Mais India Medical Devices Pvt. Ltd. | |
| | | Mediplus (India) Limited | |

| | | | |
|----|--|--|---|
| | | Romsons Group Private Limited | |
| 14 | <p align="center">IV Canula 26</p> <p>(Sterile Disposable (single Use) Teflon / PTFE IV Cannula without injection port. Size 26. Should be pack in Transparent, single blister pack. Should conform to IS 10555 Standard).</p> | Hi-Tech Medicare Devices Pvt. Ltd. Mais India Medical Devices Pvt. Ltd. Mediplus (India) Limited Romsons Group Private Limited | 4 |
| 15 | <p align="center">Mask</p> <p>(Fabric & Construction: - It must have a 3-ply type construction consisting of: a. First layer – 30 GSM – SS b. Second layer – 25 GSM – Melt blown c. Third layer – 25 GSM – SMS • UV – C Sanitized • Sealing: - The mask should be heat sealed and no stitching to keep three layers together. • Size: - Not less than 17cm x 09cm of mask and having comfortable elastic ear loop. The length of elastic loop should be minimum of 24cm in normal condition i.e., without stretching. • Nose Clip: - Rust proof adjustable nose clip. • Ear loop:- Must be with ultrasonic welding (No manual welding or stitching is permitted). • Packaging: - 25 nos. in individual poly packaging with adequate labelling/ noting • Must be ISI marked or European CE/USFDA Approved. • Breathability: - Differential pressure (Delta P) As per IS 16289 Standard. • Splash Resistance: - As per ASTM F1862 Standard. • Particulate Filtration efficiency at 0.3-microns: - As per ASTM F2299. • Bacterial Filtration Efficiency: - As per ASTM F2101. • Flammability: - As per 16 CFR. • All batches to be supplied need to be accompanied with certificate of analysis from laboratories designated for this purpose by the CDSCO/Ministry of Textiles/ Health, Government of India).</p> | Zenith Infratech Lotus Surgicals Pvt. Ltd. MEHR Enterprises Om Surgical Industries Plasti Surge Industries Pvt. Ltd. Paramount Surgimed Ltd. Romsons Group Private Limited Sanmati Packaging Tradition Globe Trade India | 9 |
| 16 | <p align="center">Spinal Needle 18</p> <p>(Specially designed administer lumbar/ subarachnoid anaesthesia. Transparent hub provides rapid detection of Cerebro-spinal fluid (CSF) flashback for confirming accurate placement. Sharp bevel design for low puncture force ensures minimal puncture trauma. Fine gauze needle design greatly reduces the risk of PDPH (Post Dura Puncture Headache). Colour coded style enables rapid size identification. Sterile, individually packed in hard blister pack/Peelable Pouche Gauze :1.2 mm and colour Pink).</p> | Romsons Group Private Limited | 1 |

| | | | |
|----|--|-------------------------------|---|
| 17 | <p>Spinal Needle 20 (Specially designed to administer lumbar/subarachnoid anaesthesia. Transparent hub provides rapid detection of Cerebro-spinal fluid (CSF) flashback for confirming accurate placement. Sharp bevel design for low puncture force ensures minimal puncture trauma. Fine gauze needle design greatly reduces the risk of PDPH (Post Dura Puncture Headache). Colour coded style enables rapid size identification. Sterile, individually packed in hard blister pack/Peelable Pouche) (Gauze : 0.9 mm and colour Yellow).</p> | Romsons Group Private Limited | 1 |
| 18 | <p>Spinal Needle 22 (Specially designed to administer lumbar/subarachnoid anaesthesia. Transparent hub provides rapid detection of Cerebro-spinal fluid (CSF) flashback for confirming accurate placement. Sharp bevel design for low puncture force ensures minimal puncture trauma. Fine gauze needle design greatly reduces the risk of PDPH (Post Dura Puncture Headache). Color coded style enables rapid size identification. Sterile, individually packed in hard blister pack/Peelable Pouche. Gauze : 0.7 mm and colour Black).</p> | Romsons Group Private Limited | 1 |
| 19 | <p>Spinal Needle 23 (the risk of PDPH (Post Dura Puncture Headache). Color coded stylet enables rapid size identification. Sterile, individually packed in hard blister pack/Peelable Pouche. Gauze : 0.6 mm and colour Deep Blue Specially designed to administer lumbar/ subarachnoid anaesthesia. Transparent hub provides rapid detection of Cerebro-spinal fluid (CSF) flashback for confirming accurate placement. Sharp bevel design for low puncture force ensures minimal puncture trauma. Fine gauze needle design greatly reduces).</p> | Romsons Group Private Limited | 1 |
| 20 | <p>Surgical Gloves Sterile 6 (Disposable Sterile Surgical Rubber Gloves Size 6 Inches Made of natural rubber Latex, powdered, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non-allergenic should Conform to IS 13422 ISI marked / CE certified / FDA Approved).</p> | Healthium Medtech Pvt. Ltd. | 4 |
| | | MRK Healthcare Pvt. Ltd. | |
| | | Anondita Healthcare | |
| | | Stoe Exim | |
| 21 | <p>Surgical Gloves Sterile 6.5 (Disposable Sterile Surgical Rubber Gloves Size 6 ½ Inches Made of natural rubber Latex, powdered, without tear, properly folded in a paper Free of holes, with</p> | Healthium Medtech Pvt. Ltd. | 4 |
| | | MRK Healthcare Pvt. Ltd. | |
| | | Anondita Healthcare | |

| | | | |
|----|---|---|---|
| | Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non-allergenic should Conform to IS 13422 ISI marked / CE certified / FDA approved). | Stoe Exim | |
| 22 | Surgical Gloves Sterile 7 (Disposable Sterile Surgical Rubber Gloves Size 7 Inches Made of natural rubber Latex, powdered, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non-allergenic should Conform to IS 13422 ISI marked / CE certified / FDA approved) | Healthium Medtech Pvt. Ltd. MRK Healthcare Pvt. Ltd. Anondita Healthcare Stoe Exim | 4 |
| 23 | Surgical Gloves Sterile 7.5 (Disposable Sterile Surgical Rubber Gloves Size 7 ½ Inches Made of natural rubber Latex, powdered, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non-allergenic should Conform to IS 13422 ISI marked / CE certified / FDA approved) | Healthium Medtech Pvt. Ltd. MRK Healthcare Pvt. Ltd. Anondita Healthcare Stoe Exim | 4 |
| 24 | Surgical Gloves Sterile 8 (Disposable Sterile Surgical Rubber Gloves Size 8 Inches Made of natural rubber Latex, powdered, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non-allergenic should Conform to IS 13422 ISI marked/CE certified/FDA approved) | Healthium Medtech Pvt. Ltd. MRK Healthcare Pvt. Ltd. Anondita Healthcare Stoe Exim | 4 |
| 25 | Surgical Gloves Non Sterile 6 (Disposable Non-Sterile Surgical Rubber Gloves Size 6 Inches Made of natural rubber Latex, powdered free, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 should Conform to IS 4148 : 1989 ISI marked / CE certified / FDA approved) | MRK Healthcare Pvt. Ltd. Anondita Healthcare Stoe Exim | 3 |
| 26 | Surgical Gloves Non-Sterile 6.5 (Disposable Non-Sterile Surgical Rubber Gloves Size 6 ½ Inches Made of natural rubber Latex, powdered free, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 should Conform to IS 4148 : 1989 ISI marked /CE certified /FDA approved) | MRK Healthcare Pvt. Ltd. Anondita Healthcare Stoe Exim | 3 |
| 27 | Surgical Gloves Non-Sterile 7 (Disposable Non-Sterile Surgical Rubber Gloves Size 7 Inches Made of natural rubber Latex, powdered free, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level | MRK Healthcare Pvt. Ltd. Anondita Healthcare | 3 |

| | | | |
|----|--|--|---|
| | (AQL) of 1.5 or less Tensile strength as per EN 455-2 should Conform to IS 4148 : 1989 ISI marked / CE certified / FDA approved) | Stoe Exim | |
| 28 | Surgical Gloves Non-Sterile 7.5 Disposable Non-Sterile Surgical Rubber Gloves Size 7 ½ Inches Made of natural rubber Latex, powdered free, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 should Conform to IS 4148 : 1989 (ISI marked /CE certified / FDA approved) | MRK Healthcare Pvt. Ltd. Anondita Healthcare Stoe Exim | 3 |
| 29 | Surgical Gloves Non-Sterile 8 Disposable Non-Sterile Surgical Rubber Gloves Size 8 Inches Made of natural rubber Latex, powdered free, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 should Conform to IS 4148 : 1989 (ISI marked /CE certified /FDA approved) | MRK Healthcare Pvt. Ltd. Anondita Healthcare Stoe Exim | 3 |
| 30 | Urine Collection Bag 2000 ml (Should have suitability for both male and female patients Should be provided with adhesive for fixation and good grip with minimal risk of allergy and injury Capacity 2000 ml and having a very low dead). | Romsons Group Private Limited | 1 |

7. The summary of Bid received status was found as follows:

1. Single Bids received for **07 MEDICAL DEVICES/CONSUMABLES**
2. Multiple Bids received for **23 MEDICAL DEVICES/CONSUMABLES**

8. It was therefore unanimously decided by the Technical Committee in its meeting dated 23.03.2021 to recommend for Re-tender of MEDICAL DEVICES/CONSUMABLES at NIT SN- 9, 12, 16, 17, 18, 19 & 30 (7 MEDICAL DEVICES/CUNSUMABLES) due to single bid condition and to evaluate the bids for **those 23 MEDICAL DEVICES/CONSUMABLES** (NIT SN- 1, 2, 3, 4, 5, 6, 7, 8, 10, 11, 13, 14, 15, 20, 21, 22, 23, 24, 25, 26, 27, 28 & 29) for which multiple bids has been received.

9. The Preliminary Evaluation of technical bids of the above-mentioned bidding firms was done by the assigned team including Drug Inspectors deputed at BMSICL. Evaluation sheets were published on the website of BMSICL and Claims/objections were invited vide notice no. 605 dated 24.04.2021.

10. After comparison of the clarification received from the bidders in the light of that notice and close examination of the documents submitted by the bidding firms, the drug wise technical evaluation findings by the TEC in today's meeting were recorded as follows:

| S.N | NIT S.N. | Name of the Product with Specification | Participating Firms | Technical Qualification Status (Qualified/ Disqualified) | Remarks |
|-----|----------|---|---|--|----------|
| | 1 | Black Braided Silk 1/0 (Sterile, Length 2 X | Unisur Lifecare Private Limited Futura Surgicare Pvt. Ltd. | Qualified Qualified | OK OK |

| | | | | | |
|---|---|---|------------------------------------|---------------------|--|
| 1 | | 75 cm (Pre-Cut & Pre-Sterilized) with CE certification). | Healthium Medtech Pvt. Ltd. | Qualified | OK |
| | | | Lotus Surgicals Pvt. Ltd. | Qualified | OK |
| | | | MCo Hospital Aids Pvt. Ltd. | Qualified | OK |
| | | | Meril Endo Surgery Private Limited | Disqualified | Valid pollution control clearance certificate as per clause 3 (f), Valid WHO-GMP/GMP Certificate as per Clause 3 (i), Maximum batch production capacity certificate as per Clause 3 (j) not submitted. |
| 2 | 2 | Black Braided Silk 1 (Sterile, Length 2 X 75 cm (Pre-Cut & Pre-Sterilized) with CE certification). | Unisur Lifecare Private Limited | Qualified | OK |
| | | | Futura Surgicare Pvt. Ltd. | Qualified | OK |
| | | | Healthium Medtech Pvt. Ltd. | Qualified | OK |
| | | | Lotus Surgicals Pvt. Ltd. | Qualified | OK |
| | | | MCo Hospital Aids Pvt. Ltd. | Qualified | OK |
| | | | Meril Endo Surgery Private Limited | Disqualified | Valid pollution control clearance certificate as per clause 3 (f), Valid WHO-GMP/GMP Certificate as per Clause 3 (i), Maximum batch production capacity certificate as per Clause 3 (j) not submitted. |
| 3 | 3 | Cap (Surgical Cap, Disposable Should be manufactured from non-woven fabric Blue/Green/ Colour Round upon wearing, with elastic Air ermeable/ Breathable should retain skin and hair particles). | MEHR Enterprises | Disqualified | Valid pollution control clearance certificate as per clause 3 (f), Market Standing Certificate for minimum three years as per Clause 3 (g), Non conviction certificate as per clause 3 (h), Valid WHO-GMP/GMP Certificate as per Clause 3 (i), Maximum batch production capacity certificate as per |

| | | | | | |
|--|--|--|-----------------------------------|--------------|--|
| | | | | | Clause 3 (j), Annual average turnover for Rs 5 Crores as per clause 3 (k) not submitted. |
| | | | Om Surgical Industries | Disqualified | Market Standing Certificate for minimum three years as per Clause 3 (g), Non conviction certificate as per clause 3 (h) & Maximum batch production capacity certificate as per Clause 3 (j) not submitted. |
| | | | Plasti Surge Industries Pvt. Ltd. | Disqualified | Market Standing Certificate for minimum three years as per Clause 3 (g), Non conviction certificate as per clause 3 (h), Maximum batch production capacity certificate as per Clause 3 (j) not submitted. |
| | | | Sanmati Packaging | Disqualified | Manufacturing license as per Clause 3 (f), Market Standing Certificate for minimum three years as per Clause 3 (g), Non conviction certificate as per clause 3 (h), Maximum batch production capacity certificate as per Clause 3 (j) not submitted. |
| | | | Tradition Globe Trade India | Disqualified | Valid pollution control clearance certificate as per clause 3 (f), |

| | | | | | |
|---|---|--|------------------------------------|---------------------|---|
| | | | | | Market Standing Certificate for minimum three years as per Clause 3 (g), Non conviction certificate as per clause 3 (h), Maximum batch production capacity certificate as per Clause 3 (j), Valid WHO-GMP/GMP Certificate as per Clause 3 (i) not submitted. |
| 4 | 4 | Catgut Plane with Needle Size "0" (Absorbable needled suture 152 cm ,1/2 circle round body with needle 30 mm). | Unisur Lifecare Private Limited | Qualified | OK |
| | | | Futura Surgicare Pvt. Ltd. | Qualified | OK |
| | | | Lotus Surgicals Pvt. Ltd. | Qualified | OK |
| | | | MCo Hospital Aids Pvt. Ltd. | Qualified | OK |
| | | | Meril Endo Surgery Private Limited | Disqualified | Valid pollution control clearance certificate as per clause 3 (f), Valid WHO-GMP/GMP Certificate as per Clause 3 (i), Maximum batch production capacity certificate as per Clause 3 (j) not submitted. |
| 5 | 5 | Catgut Chromic with needle 1-0 (Absorbable Length 70-90 cm, 1/2 circle round body with 34-40 mm needle). | Unisur Lifecare Private Limited | Qualified | OK |
| | | | Futura Surgicare Pvt. Ltd. | Qualified | OK |
| | | | Healthium Medtech Pvt. Ltd. | Qualified | OK |
| | | | Lotus Surgicals Pvt. Ltd. | Qualified | OK |
| | | | MCo Hospital Aids Pvt. Ltd. | Qualified | OK |
| | | | Meril Endo Surgery Private Limited | Disqualified | Valid pollution control clearance certificate as per clause 3 (f), Valid WHO-GMP/GMP Certificate as per Clause 3 (i), Maximum batch production capacity certificate as per Clause 3 (j) not submitted. |

| | | | | | |
|---|---|--|------------------------------------|--------------|--|
| | | | Peters Surgical India Pvt. Ltd. | Qualified | OK |
| 6 | 6 | Catgut Chromic with needle 2-0 (Absorbable Length 70-90 cm, 1/2 circle round body with 30 mm needle). | Unisur Lifecare Private Limited | Qualified | OK |
| | | | Futura Surgicare Pvt. Ltd. | Qualified | OK |
| | | | Healthium Medtech Pvt. Ltd. | Qualified | OK |
| | | | Lotus Surgicals Pvt. Ltd. | Qualified | OK |
| | | | MCo Hospital Aids Pvt. Ltd. | Qualified | OK |
| | | | Meril Endo Surgery Private Limited | Disqualified | Valid pollution control clearance certificate as per clause 3 (f), Valid WHO-GMP/GMP Certificate as per Clause 3 (i), Maximum batch production capacity certificate as per Clause 3 (j) not submitted. |
| | | | Peters Surgical India Pvt. Ltd. | Qualified | OK |
| 7 | 7 | Catgut Chromic with needle 3-0 (Absorbable Length 70-90 cm, 1/2 circle round body with 20 mm needle) | Unisur Lifecare Private Limited | Qualified | OK |
| | | | Futura Surgicare Pvt. Ltd. | Qualified | OK |
| | | | Healthium Medtech Pvt. Ltd. | Qualified | OK |
| | | | Lotus Surgicals Pvt. Ltd. | Qualified | OK |
| | | | MCo Hospital Aids Pvt. Ltd. | Qualified | OK |
| | | | Meril Endo Surgery Private Limited | Disqualified | Valid pollution control clearance certificate as per clause 3 (f), Valid WHO-GMP/GMP Certificate as per Clause 3 (i), Maximum batch production capacity certificate as per Clause 3 (j) not submitted. |
| | | | Peters Surgical India Pvt. Ltd. | Qualified | OK |
| 8 | 8 | Catgut Chromic with needle 1 (Absorbable Length 70-90 cm, 1/2 circular round body with 34-40 mm needle). | Unisur Lifecare Private Limited | Qualified | OK |
| | | | Futura Surgicare Pvt. Ltd. | Qualified | OK |
| | | | Healthium Medtech Pvt. Ltd. | Qualified | OK |
| | | | Lotus Surgicals Pvt. Ltd. | Qualified | OK |
| | | | MCo Hospital Aids Pvt. Ltd. | Qualified | OK |
| | | | Meril Endo Surgery Private Limited | Disqualified | Valid pollution control clearance certificate as per clause 3 (f), Valid |

| | | | | | |
|----|----|--|---------------------------------|---------------------|---|
| | | | | | WHO-GMP/GMP Certificate as per Clause 3 (i), Maximum batch production capacity certificate as per Clause 3 (j) not submitted. |
| | | | Peters Surgical India Pvt. Ltd. | Qualified | OK |
| 9 | 10 | Central Venous Catheter Central Line-Biluminal,Adult use (Central Venous Catheter biluminal, Adult use, Radio-opaque polyurethane catheter, with soft tip, length markings, junction hub for stitch fixation and fully transparent extension lines, with color-coded hub and white slide clamps. Components: Kink-proof guide wire with “J” tip and straight tip on dispenser, movable fixation wing with clamp, introducer needle, scalpel, dilator, valve, 5 ml luer lock syringe) (depending on set configuration). | Romsons Group Private Limited | Disqualified | Valid pollution control clearance certificate as per clause 3 (f) & MSC as per clause 3(g) not submitted. |
| | | | BL Lifesciences Private limited | Disqualified | Maximum batch production capacity certificate as per Clause 3 (j) not submitted. |
| 10 | 11 | Central Venous Catheter Central Line-Triluminal,Adult Use (Triluminal, Adult Use, Radio-opaque polyurethane catheter, with soft tip, length markings, junction hub for stitch fixation and fully transparent extension lines, with | Romsons Group Private Limited | Disqualified | Valid pollution control clearance certificate as per clause 3 (f) & Market Standing Certificate for minimum three years as per Clause 3 (g) not submitted. |
| | | | BL Lifesciences Private limited | Disqualified | |

| | | | | | |
|----|----|---|--------------------------------------|---------------------|--|
| | | color-coded hub and white slide clamps. Components: Kink-proof guide wire with “J” tip and straight tip on dispenser, movable fixation wing with clamp, introducer needle,scalpel, dilator, valve, 5 ml luer lock syringe) (depending on set configuration) | | | Maximum batch production capacity certificate as per Clause 3 (j) not submitted. |
| 11 | 13 | IV Canula 24 (Sterile Disposable (single Use) Teflon / PTFE IV Cannula with injection with wing port. Size 24 . Should be pack in Transparent, single blister pack. Should conform to IS 10555 Standard). | Hi-Tech Medicare Devices Pvt. Ltd. | Disqualified | Non Conviction Certificate as per clause 3 (h) & Maximum batch production capacity certificate as per Clause 3 (j) not submitted. |
| | | | Mais India Medical Devices Pvt. Ltd. | Disqualified | Valid WHO-GMP/GMP Certificate as per Clause 3 (i) not submitted. |
| | | | Mediplus (India) Limited | Disqualified | Valid pollution control clearance certificate as per clause 3 (f) not submitted. |
| | | | Romsons Group Private Limited | Disqualified | Valid pollution control clearance certificate as per clause 3 (f) not submitted. |
| 12 | 14 | IV Canula 26 (Sterile Disposable (single Use) Teflon / PTFE IV Cannula without injection port. Size 26. Should be pack in Transparent, single blister pack. Should conform to IS 10555 Standard). | Hi-Tech Medicare Devices Pvt. Ltd. | Disqualified | Non Conviction Certificate as per clause 3 (h) & Maximum batch production capacity certificate as per Clause 3 (j) not submitted. |
| | | | Mais India Medical Devices Pvt. Ltd. | Disqualified | Valid WHO-GMP/GMP Certificate as per Clause 3 (i) not submitted. |
| | | | Mediplus (India) Limited | Disqualified | Valid pollution control clearance |

| | | | | | |
|----|----|--|-------------------------------|--------------|---|
| | | | | | certificate as per clause 3 (f) not submitted. |
| | | | Romsons Group Private Limited | Disqualified | Valid pollution control clearance certificate as per clause 3 (f) not submitted. |
| 13 | 15 | Mask (Fabric & Construction: - It must have a 3-ply type construction consisting of: a. First layer – 30 GSM – SS b. Second layer – 25 GSM – Melt blown c. Third layer – 25 GSM – SMS • UV – C Sanitized • Sealing: - The mask should be heat sealed and no stitching to keep three layers together. • Size: - Not less than 17cm x 09cm of mask and having comfortable elastic ear loop. The length of elastic loop should be minimum of 24cm in normal condition i.e., without stretching. • Nose Clip: - Rust proof adjustable nose clip. • Ear loop:- Must be with ultrasonic welding (No manual welding or stitching is permitted). • Packaging: - 25 nos. in individual poly packaging with adequate labelling/ noting • Must be ISI marked or European CE/USFDA Approved. • | Zenith Infratech | Disqualified | Market Standing Certificate for minimum three years as per Clause 3 (g), Non conviction certificate as per clause 3 (h), Valid WHO-GMP/GMP/BIS Certificate as per Clause 3 (i) & Maximum batch production capacity certificate as per Clause 3 (j) not submitted. |
| | | | Lotus Surgical Pvt. Ltd. | Disqualified | Valid Manufacturing License as per clause 3 (f), Market Standing Certificate for minimum three years as per Clause 3 (f), Maximum batch production capacity certificate as per Clause 3 (j) & BIS Certificate not submitted |
| | | | MEHR Enterprises | Disqualified | Valid pollution control clearance certificate as per clause 3 (f), Market Standing Certificate for minimum three years as per Clause 3 (g), Non conviction certificate as per clause 3 (h), Valid |

| | | | | |
|--|--|-----------------------------------|---------------------|---|
| | <p>Breathability: - Differential pressure (Delta P) As per IS 16289 Standard. • Splash Resistance: - As per ASTM F1862 Standard. • Particulate Filtration efficiency at 0.3-microns: - As per ASTM F2299. • Bacterial Filtration Efficiency: - As per ASTM F2101. • Flammability: - As per 16 CFR. • All batches to be supplied need to be accompanied with certificate of analysis from laboratories designated for this purpose by the CDSCO/Ministry of Textiles/ Health, Government of India).</p> | | | WHO-GMP/GMP Certificate as per Clause 3 (i), Maximum batch production capacity certificate as per Clause 3 (j), Annual average turnover for Rs 5 Crores as per clause 3 (k), BIS Certificate not submitted. |
| | | Om Surgical Industries | Disqualified | Market Standing Certificate for minimum three years as per Clause 3 (g), Non conviction certificate as per clause 3 (h) & Maximum batch production capacity certificate as per Clause 3 (j) not submitted. |
| | | Plasti Surge Industries Pvt. Ltd. | Disqualified | Market Standing Certificate for minimum three years as per Clause 3 (f), Non Conviction Certificate as per clause 3 (h) & Maximum batch production capacity certificate as per Clause 3 (j) not submitted. |
| | | Paramount Surgimed Ltd. | Disqualified | Market Standing Certificate for minimum three years as per Clause 3 (f) & Non Conviction Certificate as per clause 3 (h) not submitted. |
| | | Romsons Group Private Limited | Disqualified | Manufacturing license & Valid pollution control clearance |

| | | | | | |
|----|----|---|-----------------------------|---------------------|---|
| | | | | | certificate as per clause 3 (f) & BIS Certificate not submitted. |
| | | | Sanmati Packaging | Disqualified | Manufacturing license as per Clause 3 (f), Market Standing Certificate for minimum three years as per Clause 3 (g), Non conviction certificate as per clause 3 (h), Maximum batch production capacity certificate as per Clause 3 (j) not submitted. |
| | | | Tradition Globe Trade India | Disqualified | Valid pollution control clearance certificate as per clause 3 (f), Market Standing Certificate for minimum three years as per Clause 3 (g), Non conviction certificate as per clause 3 (h), Valid WHO-GMP/GMP/BIS Certificate as per Clause 3 (i) & Maximum batch production capacity certificate as per Clause 3 (j) not submitted. |
| 14 | 20 | Surgical Gloves Sterile 6 (Disposable Sterile Surgical Rubber Gloves Size 6 Inches Made of natural rubber Latex, powdered, without tear, properly folded in a paper Free of holes, with | Healthium Medtech Pvt. Ltd. | Qualified | OK |

| | | | | | |
|----|----|--|-----------------------------|-----------|----|
| | | Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non-allergenic should Conform to IS 13422 ISI marked / CE certified / FDA Approved). | MRK Healthcare Pvt. Ltd. | Qualified | OK |
| | | | Anondita Healthcare | Qualified | OK |
| | | | Stoe Exim | Qualified | OK |
| | | | | | |
| 15 | 21 | Surgical Gloves Sterile 6.5 (Disposable Sterile Surgical Rubber Gloves Size 6 ½ Inches Made of natural rubber Latex, powdered, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non-allergenic should Conform to IS 13422 ISI marked / CE certified / FDA approved). | Healthium Medtech Pvt. Ltd. | Qualified | OK |
| | | | MRK Healthcare Pvt. Ltd. | Qualified | OK |
| | | | Anondita Healthcare | Qualified | OK |
| | | | Stoe Exim | Qualified | OK |
| 16 | 22 | Surgical Gloves Sterile 7 (Disposable Sterile Surgical Rubber Gloves Size 7 Inches Made of natural rubber Latex, powdered, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non-allergenic should Conform to IS | Healthium Medtech Pvt. Ltd. | Qualified | OK |
| | | | MRK Healthcare Pvt. Ltd. | Qualified | OK |
| | | | Anondita Healthcare | Qualified | OK |
| | | | Stoe Exim | Qualified | OK |

| | | | | | |
|----|----|---|-----------------------------|-----------|----|
| | | 13422 ISI marked / CE certified / FDA approved) | | | |
| 17 | 23 | Surgical Gloves Sterile 7.5 (Disposable Sterile Surgical Rubber Gloves Size 7 ½ Inches Made of natural rubber Latex, powdered, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non- allergenic should Conform to IS 13422 ISI marked / CE certified / FDA approved) | Healthium Medtech Pvt. Ltd. | Qualified | OK |
| | | | MRK Healthcare Pvt. Ltd. | Qualified | OK |
| | | | Anondita Healthcare | Qualified | OK |
| | | | Stoe Exim | Qualified | OK |
| 18 | 24 | Surgical Gloves Sterile 8 (Disposable Sterile Surgical Rubber Gloves Size 8 Inches Made of natural rubber Latex, powdered, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non- allergenic should Conform to IS 13422 ISI marked/CE certified/FDA approved) | Healthium Medtech Pvt. Ltd. | Qualified | OK |
| | | | MRK Healthcare Pvt. Ltd. | Qualified | OK |
| | | | Anondita Healthcare | Qualified | OK |
| | | | Stoe Exim | Qualified | OK |
| | 25 | Surgical Gloves Non Sterile 6 (Disposable Non- | MRK Healthcare Pvt. Ltd. | Qualified | OK |
| | | | Anondita Healthcare | Qualified | OK |

| | | | | | |
|----|----|---|--------------------------|-----------|----|
| 19 | | Sterile Surgical Rubber Gloves Size 6 Inches Made of natural rubber Latex, powdered free,without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 should Conform to IS 4148 : 1989 ISI marked / CE certified / FDA approved) | Stoe Exim | Qualified | OK |
| 20 | 26 | Surgical Gloves Non-Sterile 6.5 (Disposable Non-Sterile Surgical Rubber Gloves Size 6 ½ Inches Made of natural rubber Latex, powdered free,without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 should Conform to IS 4148 : 1989 ISI marked /CE certified /FDA approved) | MRK Healthcare Pvt. Ltd. | Qualified | OK |
| | | | Anondita Healthcare | Qualified | OK |
| | | | Stoe Exim | Qualified | OK |
| 21 | 27 | Surgical Gloves Non-Sterile 7 (Disposable Non-Sterile Surgical Rubber Gloves Size 7 Inches Made of natural rubber Latex, powdered free,without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as | MRK Healthcare Pvt. Ltd. | Qualified | OK |
| | | | Anondita Healthcare | Qualified | OK |
| | | | Stoe Exim | Qualified | OK |

| | | | | | |
|----|----|---|--------------------------|-----------|----|
| | | per EN 455-2 should Conform to IS 4148 : 1989 ISI marked / CE certified / FDA approved) | | | |
| 22 | 28 | Surgical Gloves Non-Sterile 7.5 Disposable Non-Sterile Surgical Rubber Gloves Size 7 ½ Inches Made of natural rubber Latex, powdered free, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 should Conform to IS 4148 : 1989 (ISI marked /CE certified / FDA approved) | MRK Healthcare Pvt. Ltd. | Qualified | OK |
| | | | Anondita Healthcare | Qualified | OK |
| | | | Stoe Exim | Qualified | OK |
| 23 | 29 | Surgical Gloves Non-Sterile 8 Disposable Non-Sterile Surgical Rubber Gloves Size 8 Inches Made of natural rubber Latex, powdered free, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 should Conform to IS 4148 : 1989 (ISI marked /CE certified /FDA approved) | MRK Healthcare Pvt. Ltd. | Qualified | OK |
| | | | Anondita Healthcare | Qualified | OK |
| | | | Stoe Exim | Qualified | OK |

11. As per the above comparative chart 9 bidders namely 1. Unisur Lifecare Private Limited 2. Futura Surgicare Pvt. Ltd. 3. Healthium Medtech Pvt. Ltd. 4. Lotus Surgicals Pvt. Ltd. 5. MCo Hospital Aids Pvt. Ltd 6. Peters Surgical India Pvt. Ltd. 7. MRK Healthcare Pvt. Ltd. 8. Anondita Healthcare 9. Stoe Exim were found technically qualified. The firm M/s Unisur Lifecare Private Limited, M/s Futura Surgicare Pvt. Ltd., M/s MCo Hospital Aids Pvt. Ltd., M/s Anondita Healthcare & M/s MRK Healthcare Pvt. Ltd. are having valid rate contract with BMSICL and M/s Lotus Surgicals Pvt. Ltd. has successfully executed the rate contract with

BMSICL. The firm M/s Stoe Exim is importer firm. Hence factory inspection of these firms were not recommended. And in the prevailing second phase of Covid-19 pandemic factory inspection of the remaining technically qualified firms was considered to be ill-advised. It was therefore unanimously recommended to carry out the physical verification of manufacturing unit of the remaining 2 (Two) newly participated technically qualified firms i.e., Healthium Medtech Pvt. Ltd. & Peters Surgical India Pvt. Ltd. later on provided they are declared as L1, L2 or L3 as per Clause 15 (a) of the bid document **“Inspections of the production and related facilities of bidders/ suppliers will be at the discretion of the Tender Inviting Authority. Such inspection may be at any stage before or after acceptance of the Bid or Award of Contract”**. It was therefore unanimously decided to recommend for opening of the financial bids of the technically qualified bidders on e-proc Portal.

12. The meeting thus concluded with a vote of thanks.

Sd/-
(Navindra Kumar)
DGM (Drugs), BMSICL, Patna.

Sd/-
(Ramendra Kumar)
OSD (D & E), BMSICL.

Sd/-
(Suman Prasad Sah)
Deputy Secretary-cum-I/C Procurement, SHSB.

Sd/-
(Sudhir Kumar)
GM (Procurement), BMSICL, Patna

Sd/-
(Rajani Kant)
CGM (Supply Chain), BMSICL, Patna.

Sd/-
(Dr. Navin Chandra Prasad)
Director in Chief, Health Dept.
Govt. of Bihar.