

**BMSICL**

www.bmsicl.gov.in
md-bmsicl-bih@nic.in
5th Floor, Biscomaun Bhavan, Gandhi maidan,
Patna 800 001 Phone: +91 612 2219634

Bihar Medical Services & Infrastructure Corp Ltd.
(A Govt. of Bihar undertaking)

Corrigendum-1

Bihar Medical Services and Infrastructure Corporation Limited (BMSICL) had invited E-Bids from the interested parties for Tender for the supply of Medical Equipment for different Govt. Medical Colleges and Hospitals of Bihar vide Notice Inviting Tender No.-BMSICL/2016-17/ME-054. Detailed tender document containing eligibility criteria, selection mechanism, other terms and conditions are available on the website www.eprocbihar.gov.in and www.bmsicl.gov.in. In order to facilitate the downloading of amendments as per Pre-bid meeting held on 17/03/2017, the tender schedule is being revised as follows:-

Tender Reference No.	BMSICL/2016-17/ME-054
Date and time for downloading of bid document	up to 11th April 2017 till 15:00 Hrs.
Last date and time of submission of online bids	21th April 2017 upto 17:00 Hrs.
Last date and time for submission of original documents of EMD and Document Fee	24st April 2017 until 14:00 Hrs.
Date, Time and Place of opening of Technical Bid	24st April 2017(at 15:00 Hrs.) on the website of www.eproc.bihar.gov.in in the office of BMSICL
Date and time of opening of financial Bids	To be announced later on www.eproc.bihar.gov.in

Note:- Please refer to **Annexure- I (pages: 1)** of this corrigendum before submission of bids

Sd/-
GM (Procurement)
BMSICL

Annexure-1

Name of Equipment - OT LED Light		
Sl no.	Technical specification Before Amendment	Technical specification After Amendment
3 (i)	LUX intensity 1,40,000 Lux & Satellite 1,40,000 Lux or above.	LUX intensity 1,60,000 Lux or above & Satellite 1,40,000 Lux or above.
3 (ii)	Light Field diameter shall be above 24 cm or better.	Light Field diameter shall be above 20 cm or more

Name of Equipment - Computerized Virtual Dissection Table		
Sl no.	Technical specification Before Amendment	Technical specification After Amendment
	Full size Male and female virtual cadavers, head to toe in real life 1:1 size.	Full size Male and female virtual cadaveric data
	It should provided 3D Scans for visualizing beating hearts and respiration with full interactivity and should allow users cut into and manipulate the scans while retaining motion capabilities.	optional
	Screen Dimension 216cm 957 cm x207cm and Resolution 3840 x1080 3840 x1080.	Screen dimensions should be atleast 55 inch and resolution 1920x1080

Name of Equipment - OT Table Electrohydraulic		
Sl no.	Technical specification Before Amendment	Technical specification After Amendment
2 (i)	Up / Down 680-1000mm	Up / Down 680-1070mm
2(v)	Top Slide 300mm	Top Slide 300mm-350mm
2 (vi)	Braking (by hand switch)	Braking (by hand switch/Foot Pedal)
4 (ii)	Back Plate (Sitting Position) -40 to +80	Optional
7	The base cover, lifting column cover and side rails should be made of stainless steel grade SS grade 304	Optional
10	The weight bearing capacity of the table shall be 175kg.	The weight bearing capacity of the table shall be 175kg or above