



Bihar Medical Services & Infrastructure Corporation Limited, 2nd & 3rd Floor, Swasthya Bhawan, Behind IGIMS, Sheikhpura, Adjacent to State Health Society, Patna 800023, Phone/Fax: +91612 2283287,+ 91612 2283288

Corrigendum-III

Bihar Medical Services and Infrastructure Corporation Limited (BMSICL) had invited E-Bids from the interested parties for the procurement, rate contract and the supply of medical equipment for different Govt. Institutions of Bihar vide Notice Inviting Tender No.- BMSICL/2023-24/ME-329. During and after Pre-bid meeting various suggestions were received from different prospective bidders regarding amendment in technical specification of equipment which were discussed and deliberated on by the experts, who after due deliberation recommended certain amendments in the technical specification of the equipment, which are annexed as Annexure-I of this corrigendum. In order to facilitate maximum participation of bidders the tender schedule is being revised as follows:-

Tender Reference No.	BMSICL/2023-24/ME-329
Last date and time of submission of online bids	07th February 2024 till 17:00 Hrs.
Last date and time of submission of original documents of EMD, Tender Fee and Document	08th February 2024 till 14:00 Hrs.
Date, Time and Place of opening of Technical Bid	08th February 2024 (at 15:00 Hrs.) on the website of https://eproc2.bihar.gov.in in the office of BMSICL
Date and time of opening of financial Bids	To be announced later on https://eproc2.bihar.gov.in

Note:-

- 1. Bidders are advised to refer to the Annexure-I of this corrigendum before submission of bid.**
- 2. Those who have submitted their bids are requested to re-submit their bids in accordance with this corrigendum.**

Annexed:- as above

**Sd/-
GM (Procurement)
BMSICL**

Annexure-I			
Name of Equipment: -Automated Blood Cell Counter 5 Part			
Sl. No	Technical Specification as per tender	Amendment as per Corrigendum-1	Final Amendment as on 18.01.2024
	Technical Characteristics (specific to this type of device)		
7. STANDARDS AND SAFETY			
1	Certificates (pre-market, sanitary); Performance and safety standards (specific to the device type); Local and/or international	1. Should be USFDA & CE (with notified body) approved product. 2. Manufacturer and Supplier should have ISO 13485 certification. 3. Shall meet internationally recognized for Electromagnetic Compatibility (EMC) for electro-medical equipment: 61326-1 OR Certified to be compliant with IEC 61010-1, IEC 61010-2-281, 61010-2-101 for safety.	Should be USFDA/CE (with notified body) /CEIVD approved product. 2. Manufacturer and Supplier should have ISO 13485 certification.

Name of Equipment: - Flow cytometer			
Sl. No	Technical Specification as per tender	Amendment as per Corrigendum-1	Final Amendment as on 18.01.2024
1	Should have minimum capability of at least 10 fluorescent colours and 12 parameters. For each parameter the flow cytometer should be capable of measuring area, height and width.	1. System Should have minimum capability of at least 10 fluorescent colours and 12 parameters. 2. System should have a provision of future upgradable for more colors. (optional) 3. For each parameter the flow cytometer should be capable of measuring area, height and width simultaneously.	1. System Should have minimum capability of at least 10 fluorescent colours and 12 parameters. 2. System should have a provision of future upgradable for more colors. (optional) 3. For each parameter the flow cytometer should be capable of measuring area, height and width.