

Minutes of Technical Committee Meeting for Evaluation of Technical Bids for Tender No. BMSICL/2023-24/ME-320 held on 07th November 2023 at Conference hall of BMSICL, Patna.

1. Attendance- As per the attendance register.
2. Tender for Procurement, Rate contract and Supply of Medical Equipment was floated vide Tender No.- BMSICL/2023-24/ME-320. In the TEC meeting held on 05/10/2023 for online opening of technical bids on the website <https://eproc2.bihar.gov.in>, equipment wise successfully submitted bids were found as following:-

Sl. N o.	Name of Equipment	Name of Bidders	Remarks
1	Laproscopy System with Hand Instruments	M/s Vikas Medical Devices	Three Bids (03)
		M/s Indigenous	
		M/s Heager India Pvt Ltd	
2	Hystoroscopy System with Camera System	M/s Vikas Medical Devices	Three Bids (03)
		M/s Decent Medical Equipment Systems Pvt Ltd	
		M/s Heager India Pvt Ltd	

3. In the said meeting it was decided unanimously to proceed for the preliminary technical evaluation of the bids received for the above said equipment.
4. The equipment/firm wise evaluation sheets of technical bids for the above said equipment, were uploaded on the website of BMSICL vide letter no- BMSIC/45025/133-2023/5889, dated:-12.10.2023 for clarification/objection, if any.
5. The clarification /objection received from the bidders were examined by the committee in the TEC meeting held today, in which the equipment wise/firm wise evaluation sheets of post clarification documents were examined by the committee and the observations were found as following.

Sl. No.	Name of Equipment	Name of Bidder	Observation found before clarification	Observation found after clarification	Remarks
1.	Laproscopy System with Hand Instruments	1.M/s Vikas Medical Devices	<p>1.Hand Instrument Make& Model/Cat No. not submitted Hence not conforming to checklist point B1 .</p> <p>2.Manufacturer's Authorization of Shalya not submitted, Hence not conforming to checklist point B11.</p> <p>3. Two (2) end user certificate submitted instead of three, Hence not conforming to checklist point B15.</p> <p>4.Quality certificate of Light Source and Hand Instrument not submitted, Hence not conforming to checklist point B19.</p>	<p>1.Hand Instrument Make-Karl Storz and Mindray & Model/Cat No. submitted, Hence conforming to checklist point B1.</p> <p>2.Manufacturer's Authorization of Shalya submitted, Hence conforming to checklist point B11.</p> <p>3. End user Certificate dt 03.10.2023,Issued by Tata Steel Limited Jamshedpur Jharkhand submitted, Hence conforming to checklist point B15.</p> <p>4. Valid quality certificate of Light Source and Hand Instrument submitted, Hence conforming to checklist point B19.</p>	Technically Qualified
		2. Indigeno us	<p>1. Turnover Certificate not submitted, Hence not conforming to checklist point B6.</p> <p>2. Manufacturer's Authorization of</p>	<p>1. Turnover Certificate submitted, Hence conforming to checklist point B6.</p> <p>2. Manufacturer's Authorization of</p>	Technically Dis-qualified

			<p>RZ medizintechnik submitted but not as per Annexure V and Manufacturer's Authorization of LG Electronics not submitted, Hence not conforming to checklist point B 11.</p> <p>3. Bidder has submitted one no. of valid purchase order of quoted model instead of three, Hence not conforming to checklist point B 13.</p> <p>4. End user certificate not submitted, Hence not conforming to checklist point B 15.</p> <p>5. Quality certificate of Suction Irrigation, CO2 insufflator and 4K Camera not submitted, Quality certificate and ISO 13485 of 4k Medical Grade Monitor not submitted, ISO 13485 of Hand Instrument not submitted, Hence not conforming to</p>	<p>LG Electronics submitted. Manufacturer's Authorization of RZ medizintechnik submitted but not as per Annexure V, Hence not conforming to checklist point B 11.</p> <p>3. Bidder has submitted new purchase order which is not mention in the submitted performance statement as per annexure VI, Hence not conforming to checklist point B13.</p> <p>4. End user certificate not submitted, Hence not conforming to checklist point B 15.</p> <p>5. Valid Quality certificate and ISO 13485 of 4k Medical Grade Monitor submitted, ISO 13485 of Hand Instrument submitted but Quality certificate of Suction Irrigation not submitted, Hence not conforming to</p>	
--	--	--	--	---	--

			checklist point B 19.	checklist point B 19.	
		3.Heager India Pvt.Ltd.	<p>1. Article of Association not submitted, Hence not conforming to checklist point B3.</p> <p>2.End user certificate Dt- 30.01.2023,Issued by District Health Department Zila Parishad, Satara, Maharashtra in which Date of issuing of performance certificate is older than 6 months from the date of publication of the tender, hence not conforming to checklist point B15.</p> <p>3. ISO 13485 of 4k Medical Grade Monitor-not submitted, Hence not conforming to checklist point B19.</p>	<p>1. Article of Association submitted, Hence conforming to checklist point B3.</p> <p>2. End user certificate Dated13.04.2023,is sued by District Health Department Zila Parishad, Satara, Maharashtra submitted, hence conforming to checklist point B15.</p> <p>3. ISO 13485 of 4k Medical Grade Monitor-submitted, Hence conforming to checklist point B19.</p>	Technically Qualified
2.	Hysteroscopy System with Camera System	1.M/s Vikas Medical Devices	1.Instrument Make& Model/Cat No. not submitted and Hysteroscopy pump for intra uterine fluid	1. 1. Instrument Make-Karl Storz & Model/Cat No. submitted and Hysteroscopy pump for intra	Technically Qualified

			<p>management Make& Model/Cat No. not submitted, Hence not conforming to checklist point B1.</p> <p>2. Manufacturer's Authorization of Shalya not submitted, Hence not conforming to checklist point B11.</p> <p>3. Quality certificate and ISO 13485 of IMAGE/ VIDEO RECORDING System, Instrument and Hysteroscopy pump for intra uterine fluid management not submitted, Hence not conforming to checklist point B19.</p>	<p>uterine fluid management Make-Shalya India & Model/Cat No.- Aquaflow submitted, Hence conforming to checklist point B1.</p> <p>2. Manufacturer's Authorization of Shalya submitted, Hence, conforming to checklist point B11.</p> <p>3. Valid Quality certificate and ISO 13485 of IMAGE/ VIDEO RECORDING System, Instrument and Hysteroscopy pump for intra uterine fluid management submitted, Hence conforming to checklist point B19.</p>	
		2. Heager India Private Limited	<p>1. Hysteroscopy pump for intra uterine fluid management Make& Model/Cat No. not submitted, Hence not conforming to checklist point B1.</p> <p>2. Article of Association not submitted, hence</p>	<p>1. Hysteroscopy pump for intra uterine fluid management Make-Shalya India & Model/Cat No.- Aquaflow submitted, Hence conforming to checklist point B1.</p> <p>2. Article of Association submitted, hence conforming to checklist point B3.</p>	Technically Qualified

			<p>not conforming to checklist point B3.</p> <p>3. End user certificate not submitted for the offered model equipment. Hence not conforming to checklist point B15.</p> <p>4. Quality certificate & ISO 13485 of Hysteroscopy pump for intra uterine fluid management not submitted, hence not conforming to checklist point B19.</p>	<p>3. End user certificate submitted for 03 nos. of offered model equipment. Hence conforming to checklist point B-15.</p> <p>4. Valid Quality certificate & ISO 13485 of Hysteroscopy pump for intra uterine fluid management submitted, hence conforming to checklist point B19</p>	
		3. Decent Medical Equipment Systems Pvt. Ltd.	<p>1. Make & Model/Cat No.:- Not Submitted of below items</p> <p>1. Camera control unit</p> <p>2. 4K medical grade Monitor</p> <p>3. Light source</p> <p>4. IMAGE/ VIDEO RECORDING System</p> <p>5. Instrument</p> <p>6. Hysteroscopy pump for intra uterine fluid Management, hence not conforming to checklist point B1.</p>	<p>Make & Model/Cat No.:- Submitted</p> <p>1.(a) Camera control unit Make-Tontarra Germany Model-camera control unit Endoor cam 4K</p> <p>(b). 4K medical grade Monitor Make-Tontarra Germany Model-4K medical grade Monitor Tontarra 32"4K</p> <p>(c). Light source Make-Tontarra Germany Model-LED Light source ENDOOR LIGHT 300 LED Touch screen.</p> <p>(d) IMAGE/</p>	Technically Dis-qualified

			<p>VIDEO RECORDING System Make- Tontarra Germany Model- Image/video recording and data Archiving system Endo Cap Touch. (e). Instrument Make-Tontarra Germany Model- Hysteroscopy Instrument on pg no 78. 6.Hysteroscopy pump for intra uterine fluid Make- Tontarra Germany Model-Suction irrigation pump,Lap Pump Touch screen, hence conforming to checklist point B1.</p>	
		<p>2. Article of Association & Memorandum of Association not submitted, hence not conforming to checklist point B3 .</p>	<p>2. Article of Association & Memorandum of Association submitted, hence conforming to checklist point B3</p>	
		<p>3. Copy of certificate of Audited Report for any three of last four consecutive Assessment years not submitted, hence not</p>	<p>3. Copy of certificate of Audited Report for FY 2019-20,2020- 21 and 2021-22 submitted, hence conforming to checklist point B5.</p>	

			<p>conforming to checklist point B5.</p> <p>4. Purchase order No.-24/273/2017-Rishi(Admin)/1571 ,Dated 02.06.2020, issued by AIIMS Rishikesh of (02) units of quoted model of equipment is older than 3 Financial Year from the BID Publishing, hence not conforming to checklist point B 13.</p> <p>5.End User Certificate submitted issued by AIIMS, Risikesh dated 13.03.2023 where the date of issuing of performance certificate is older than 6 months from the date of publication, hence not conforming to checklist point B15.</p> <p>6.Quality certificate and ISO 13485 certificate of quoted model not submitted, hence not conforming to checklist point B19.</p>	<p>4 Copies of new purchase order submitted which is not mention in the submitted performance statement as per annexure VI, hence not conforming to checklist point B13.</p> <p>5.End User Certificate submitted issued by AIIMS, Risikesh dated 13.03.2023 where the date of issuing of performance certificate is older than 6 months from the date of publication, hence not conforming to checklist point B15</p> <p>6.Valid Quality certificate and ISO 13485 certificate of quoted model submitted, hence conforming to checklist point B19</p>	
--	--	--	--	---	--

6.It was therefore decided for the demonstration of the quoted model of equipment by the following technically qualified bidders:-

Sl. No.	Name of Equipment	Technically Qualified Bidders
1	Laproscopy System with Hand Instruments	M/s Vikas Medical Devices
		M/s Heager India Pvt Ltd
2	Hysteroscopy System with Camera System	M/s Vikas Medical Devices
		M/s Heager India Pvt Ltd

The meeting thus concluded with vote of thanks

Sd/-
(Sh. Navindra Kumar)
DGM (Equipment),
BMSICL

Sd/-
(Sh. Mirtunjay Kumar)
A.O–Cum-Incharge, Procurement,
SHSB

Sd/-
(Sh. Ravindra Kumar)
GM, Procurement
BMSICL

Sd/-
(Sh. Najar Hussan)
CGM, Supply Chain
BMSICL

Sd/-
(Sh. Sudhir Kumar)
Joint secretary, Health Deptt.,
G.O.B

Sd/-
(Dr. Sunil Kumar Jha)
Director in Chief Health Deptt.,
G.O.B