

Tender No.: BMSICL/201-16/ME-043

Minutes of Pre-Bid Meeting for tender No.: BMSICL/2016-17/ME-043 held on 15.09.2016 at conference Hall of BMSICL 5th floor, Biscomoun Bhawan, Patna.

Background:

Notice Inviting tender (NIT) was published (Advertisement ID: 6077, Letter no.1730) in various newspapers.

The tender notice was also uploaded on the website of BMSICL i.e. www.bmsicl.gov.in & www.eproc.bihar.gov.in for information of all concern.

Pre-Bid Meeting:-

The Pre-Bid Meeting of the prospective bidders for above mentioned tender was held on 03.09.2016 at 3.00PM at conference Hall of BMSICL 5th floor, Biscomoun Bhawan, Patna in presence of following members:


1. Dr. Harsh Kumar (Sr. Resident) NMCH, Patna
2. Dr. Ashutosh Kumar Sinha (Assistant Professor), NMCH, Patna.
3. Dr. D.K. Singh (Prof. Dental PDCH, Patna)
4. Dr. Arvind Kumar (Jt. Director) IGIC, Patna.
5. Dr. Anil Kumar Singh (Asst. Director) IGIC, Patna.
6. Sri Sunil Kumar Singh, GM (Procurement), BMSICL, Patna.

For all prospective bidders or their authorized representatives participated in the pre-bid meeting. Attendance of participants was taken prior to the commencement of the meeting and records is maintained in the attendance register.

As an outcome of discussion held in the meeting, some necessary amendments in technical specification of equipment's were considered to be incorporated in bidding document. A detail of amendments is attached is annexure- (I) & (II)

03.09.16

CGM (Supply Chain)



166

Amendment sheet of Tender Ref. no.BMSICL/2016-17/ME-043		
Annexure-I		
Heart Lung Machine		
SR.NO. As per Tender Document	EXISTING TECHNICAL SPECIFICATION	TECHNICAL SPECIFICATION AFTER AMENDMENTS
A-14	Instrument tray position able with long monitoring arm.	Optional
A-15	Light weight surface table; writing surface.	CE/FDA endorsement for light weight, surface table, writing surface is optional
D-(i)	Safety monitor should have capability for computer interface to retrieve perfusion data	Optional
E-3	Instrument tray with mounting	Optional
E-4	Thermal blanket for adult & Paediatric patients.	Item stands Deleted from scope of supply
E-5	On- line measurement of PH, PCO2 & Hb for neonatal cardiac surgery.	Provision of the item is optional
F-(i)	Machine cover.	Provision of the item is optional
F-(ii)	System should be provided with appropriate furniture like adjustable revolving chair for the per fusionist to operate the system.	2 number branded make adjustable revolving chairs for perfusionist to be supplied

A
26/19
eh
24/09/16

