



**Bihar Medical Services & Infrastructure Corporation Limited**  
**2<sup>nd</sup> & 3<sup>rd</sup> Floor, Swasthya Bhawan, Behind IGIMS, Sheikhpura,**  
**Adjacent to State Health Society, Patna- 800014, Bihar,**  
**Phone/Fax: +91612 2283287, 91612 2283288**

---

#### **Corrigendum-IV**

Bihar Medical Services and Infrastructure Corporation Limited (BMSICL) had invited E-Bids from the interested parties for Procurement, Rate contract, Supply, Installation of Medical Equipment vide Tender No.-BMSICL/2022-23/ME-302. During and after Pre-bid meeting various suggestions were received from different prospective bidders regarding amendment in technical specification of equipment which were discussed and deliberated on by the experts. On the basis of their recommendations certain amendments in the technical specification of the equipment have been made which are annexed as **Annexure-I** of this corrigendum.

**S/d**  
**GM (Procurement)**  
**BMSICL**

### **Annexure-1**

<b>Name of Equipment - Ice Lined Refrigerator (ILR)</b>		
<b>Sl. No</b>	<b>Technical Specification as per tender</b>	<b>Final Amendment</b>
1	Capacity: 200 – 300 Liters	No Change
2	Vertical Ice Lined Refrigerator to maintain temperature of +2° C to +8° C at ambient temperature or +5° C to +45 degree C with intermittent or continuous electricity supply for at least 8 hrs. in a 24 hrs cycle.	Horizontal/Vertical Ice Lined Refrigerator to maintain temperature of +2° C to +8° C at ambient temperature or +5° C to +45 degree C with intermittent or continuous electricity supply for at least 8 hrs. in a 24 hrs cycle.
3	The temperature difference between any two points in the cabinet should not be more than +2 degree C once stabilized.	No Change
4	Holdover Time (maintain the temperature +2°C to +8° C): 20 hrs or more in a continuous ambient temperature of 43° C	No Change
5	Solid /Double Glazed Glass door with lock and handle	No Change
6	Type: Compressor cycled. CFC free (both for refrigeration and insulation). All system tubing (suction tube. Freezer tube and condensing tube) SHOULD BE OF COPPER.	No Change
7	A microprocessor based control unit should be provided for setting of temperature and display of the following features: Cabinet temperature Power on LED/LCD indicator	No Change
8	Auto visual alarm against violation of temperature range (less than + 2 degree Celsius and more than 8 degree Celsius). Door open alarm	No Change
9	Adjustable storage wire shelves (minimum three) allowing free circulation of air	No Change
10	digital thermometer (temp range - 30° C to + 50°C)	No Change
11	Product should be BIS or USFDA or European CE approved	Product should be BIS or USFDA or European CE (Issued by notified Body) approved
12	Power supply :220VAC/50Hz	No Change
13	Suitable voltage stabilizer should be supplied with each unit	No Change
14	The supplier will provide the initial 03 years warranty and 07 years comprehensive Annual Maintenance Contract (CAMC) including all spare parts. Compressor & repairs.	No Change