



Bihar Medical Services & Infrastructure Corporation Limited, 2nd & 3rd Floor, Swasthya Bhawan, Behind IGIMS, Sheikhpura, Adjacent to State Health Society, Patna 800023, Phone/Fax: +91612 2283287,+ 91612 2283288

Corrigendum-II

Bihar Medical Services and Infrastructure Corporation Limited (BMSICL) had invited E-Bids from the interested parties for the procurement, rate contract and the supply of medical equipment for different Govt. Institutions of Bihar vide Notice Inviting Tender No.- BMSICL/2023-24/ME-309. During and after Pre-bid meeting various suggestions were received from different prospective bidders regarding amendment in technical specification of equipment which were discussed and deliberated on by the experts, who after due deliberation recommended certain amendments in the technical specification of the equipment, which are annexed as Annexure-I of this corrigendum. Rest of the terms & conditions of the NIT & Corrigendum-I shall remain unchanged:-

Tender Reference No.	BMSICL/2023-24/ME-309
Last date and time of submission of online bids	26th June 2023 till 17:00 Hrs.
Last date and time of submission of original documents of EMD, Tender Fee and Document	27th June 2023 till 14:00 Hrs.
Date, Time and Place of opening of Technical Bid	27th June 2023 (at 15:00 Hrs.) on the website of https://eproc2.bihar.gov.in in the office of BMSICL
Date and time of opening of financial Bids	To be announced later on https://eproc2.bihar.gov.in

Note:-

- 1. Bidders are advised to refer to the Annexure-I of this corrigendum before submission of bid.**
- 2. Those who have submitted their bids are requested to re-submit their bids in accordance with this corrigendum.**

Annexed:- as above

**Sd/-
GM (Procurement)
BMSICL**

Annexure-I		
Name of Equipment: - MORTUARY CHAMBER (TWO BODY)		
Sl. No	Technical Specification as per tender	Final Amendment
1	Storage temperature: 2-5° C	No Change
2	Solid state digital controller, vapour proof light with switch.	No Change
3	Electronic Safety system which operates whenever there is fluctuation/ deviation from set temperature operates even in case of power failure by audio/ visual alarm.	No Change
4	Temperature variation adjustment along with safety thermostat present Construction.	No Change
5	All stainless steel 304 which include inner, front door and carriage. Carriage frame to be have smooth and noiseless movement.	No Change
6	Base should be 2- 2.5 feet from the ground level & 304 stainless steel.	No Change
7	Hermetically sealed compressor serves refrigeration unit, to maintain temperature range of 2° C to 5° C Unit is fitted with Digital Temperature Indicator-cum-controller.	No Change
8	Automatic locking facility of carriage for storage when returned to close position.	No Change
9	Air circulation Forced Refrigerant.	No Change
10	PUF insulation on all sides.	No Change
11	Maximum permitted load per shelf – 155 kg and of Body height 6.5 Feet.	No Change
12	US FDA / European CE /BIS/CE Approved model should be offered.	US FDA/European CE/BIS/CE/ISO Approved model should be offered, The ISO certifications must be issued by bodies registered with NABCB under the Medical Devices Quality Management System.
13	Voltage – 220VAC ±10%,	No Change
14	Suitable stablizer for smooth functioning of equipment (Automated Type).	No Change