

Bihar Medical Services & Infrastructure Corporation Limited, 4th floor, Bihar State Building Construction Corporation Limited. Hospital Road, Shastri Nagar, Patna 800023, Phone/Fax: +91612 2283287,+ 91612 2283288

<u>Corrigendum-III</u>

Bihar Medical Services and Infrastructure Corporation Limited (BMSICL) had invited E-Bids from the interested parties for the procurement, rate contract and the supply of BOD Incubator, Deep Freezer (-20 Degree C) & Binocular Microscope for different Govt. Institutions of Bihar vide Notice Inviting Tender No.- BMSICL/2021-22/ME-273. During and after Pre-bid meeting various suggestions were received from different prospective bidders regarding amendment in technical specification of equipment which were discussed and deliberated on by the experts, who after due deliberation recommended certain amendments in the technical specification of the equipment, which are annexed as Annexure-I of this corrigendum. In order to facilitate maximum participation of bidders the tender schedule is being revised as follows:-

Tender Reference No.	BMSICL/2021-22/ME-273
Date and time for downloading of bid document	Up to 04 th May 2022 till 17:00 Hrs.
Last date and time of submission of online bids	05 th May 2022 till 17:00 Hrs.
Last date and time of submission of original documents of EMD, Tender Fee and Document	06 th May 2022 till 14:00 Hrs.
Date, Time and Place of opening of Technical Bid	06 th May 2022 (at 15:00 Hrs.) on the website of <u>www.eproc.bihar.gov.in</u> in the office of BMSICL
Date and time of opening of financial Bids	To be announced later on www.eproc.bihar.gov.in

Note:- Bidders are advised to refer to the Annexure-I of this corrigendum before submission of bid.

Annexed:- as above

Sd/-GM (Procurement) BMSICL

	Annexure-I	
SI. No	Technical Specification as per tender	Amendment Proposed
	1. BOD Incubator	
1	The temperature should be controlled by the microprocessor based digital temperature controller with LED display along with provision for manual thermometer recording.	No Change
2	Inner chamber capacity: 6 cubic feet.	No Change
3	Temperature range : +5 degree C to 80 degree C.	Temperature range: ± 15 degree C (± 05) to 60 degree C (± 20).
4	Interior chamber: Stainless steel for easy cleaning and decontamination, rust free.	No Change
5	The gas between inner and outer chamber should be filled with high grade glass wool/PUF insulation to avoid thermal loss.	No Change
6	Digital display of temperature and time.	No Change
7	Timer: 1 minute to 100 hours and hold position.	1 Minute to 100 Hours and Continuous
8	Heating and natural convection for homogenous temperature distribution.	No Change
9	Temp. Accuracy +/-1 degree C.	No Change
10	Inner chamber should have transparent, glass/fiber door for the observation.	No Change
11	Minimum two adjustable shelves.	No Change
12	Power 230+/- 10V 50Hz.	Power 230+/- 10V 50Hz with stabilizer 160V to 280V
13	European CE (issued by notified body)/US (FDA) approved model should be offered.	European CE (issued by notified body)/US (FDA) /BIS approved model should be offered.
	2. Binocular Microsco	
1	Viewing Head: Sliding Binocular Head inclined at 30 degree to 45degree, 360 degree Rotatable.	No Change
2	Eyepiece: Wide Field Eyepiece WF 10x/15x/18x.	No Change
3	Objective: Achromatic Para focal objective 4x, 10x, 40x, 100x	No Change
4	Nosepiece: Quadrupole nosepiece.	No Change
5	Stage; Double layer Mechanical stage 135 to 140x130-140mm.	No Change
6	Condenser: Abbe NA 1.25 with tris diaphragm & Filter.	No Change
7	Illumination: Koehler illumination, Built in Illumination, LED Lamp.	No Change
8	Filter; Blue Filter.	No Change
9	Must provide following Accessories:	No Change
	• LED Lamp	No Change
	• Vinyl Cover.	No Change

	Cleaning brush.	No Change		
10	European CE (issued by notified body)/US (FDA) approved model should be offered.	European CE (issued by notified body)/US (FDA) /BIS approved model should be offered.		
	3. Deep Freezer (-20 degree C)			
1	Vertical type with twin door and adjustable shelves.	No Change		
2	Capacity: 380 to 450 liters.	No Change		
3	Temperature: -20 degree C	No Change		
4	Accuracy +/- 1 degree C.	No Change		
5	Must be able to perform at an ambient temperature up to 40 degree C. user friendly, non CFC refrigeration. Body is mounted on a sturdy angle iron frame and is mounted on castor wheels.	Must be able to perform at an ambient temperature up to 45 degree C. user friendly, non CFC refrigeration. Body is mounted on a sturdy angle iron (power coated) frame and is mounted on castor wheels.		
6	Temperature recording range: For recording temperature tolerance limit 8 degree C to -20 degree C.	Temperature recording facility for 8 degree C to -20 degree C Graph paper for 1 Yr.		
7	Power supply 230+/- 10 volts :50 Hz.	No Change		
8	Voltage stabilizer: Voltage regulator of appropriate rating to be included to cope with 160-260V.	No Change		
9	Microprocessor control, digital display with battery back-up and alarm should be there.	No Change		
10	European CE (issued by notified body)/US (FDA) approved model should be offered.	European CE (issued by notified body)/US (FDA) /BIS approved model should be offered.		