



**Bihar Medical Services & Infrastructure Corporation
Limited 2nd & 3rd Floor, Swasthya Bhawan, Behind IGIMS,
Sheikhpura, Adjacent to State Health Society, Patna-
800014, Bihar, Phone/Fax: +91612 2283287, 91612 2283288**

Corrigendum-III

Bihar Medical Services and Infrastructure Corporation Limited (BMSICL) had invited E-Bids from the interested parties for Procurement, Rate contract, Supply, Installation of Medical Equipment vide Tender No.-BMSICL/2023-24/ME-314. After publishing corrigendum-II for this tender representation of certain prospective bidders were received regarding amended technical specification. After due deliberations on the representation submitted by the prospective bidders, following amendment has been suggested which is annexed as annexure-I of this corrigendum. Rest of the term and conditions remains unchanged. In order to facilitate maximum participation of bidders tender schedule is being revised as following:-

Tender Reference No.	BMSICL/2023-24/ME-314
Last date and time of submission of online bids	27th July 2023 till 17:00 Hrs.
Last date and time of submission of original documents of EMD, Tender Fee and Document.	28th July 2023 till 14:00 Hrs.
Date, Time and Place of opening of Technical Bid	28th July 2023 (at 15:00 Hrs.) on the website of https://eproc2.bihar.gov in the office of BMSICL
Date and time of opening of financial Bids	To be announced later on https://eproc2.bihar.gov

**SD/-
GM (Procurement)
BMSICL**

Annexure-I

Name of Equipment: - Apheresis Machine		
Sl. No	Technical Specification after Amendment	Technical Specification after Re-Amendment
1	Continuous/Intermittent flow Blood Cell separator	No Change
2	Dual/single Needle operation.	No Change
3	Built in automated protocols for majority (4 of 6) of the below procedures.	No Change
a	Leukoreduced Plasma Collection	No Change
		b. Therapeutic Plasma Exchange.
		c. Single or double RBC collection and/or RBC Exchange
		d. Peripheral Blood Stem Cell Collections
		e. Granulocyte Collection.
f	Leukoreduced platelet collection or plateleapheresis	No Change
4	Automatic Pump Loading & Priming of disposables sets.	No Change
5	Automated Self-test to ensure maximum Donor Safety.	No Change
6	Automatic Leukoreduction validation of platelets and plasma at the end of procedure.	No Change
7	Adjustable product concentration	No Change
8	Separate Anticoagulation pump with custom programming adjustability	No Change
9	Safety-check to prevent Platelets count dropping below safety level for Donor safety.	No Change
10	Configurable maximum volume depletion levels either by weight or percentage of Total Blood Volume	No Change
11	Extracorporeal volume less than 250 ml.	No Change
12	Built in Access & Return Pressure sensor.	No Change
13	Built in air detectors to prevent air embolism	No Change
14	Built in ACD Detector.	No Change

15	Built in contamination monitor for monitoring & preventing RBC contaminations in platelet collection and plasma exchange.	No Change
16	Audio visual alarms.	No Change
17	Periodic Instrument Calibration certificate for the various parameters and QC of the products should be provided/maintained by the vendor.	No Change
18	USFDA 510(K) / European CE (issued by notified body) certified approved Model should be offered.	No Change
19	Power Supply 220VAC +/- 10 %, 50Hz.	No Change
20	Online UPS of suitable rating with voltage regulation, spike protection and maintenance free batteries for 60 minutes back up.	No Change
21	Cost of 500 (Five hundred) EACH test of Platelet kit & Plasma Exchange kit along with the price of startup and shut down consumption of reagents and other Reagents (Cleaner/Washer/Diluent /Kits /Calibrator, QCs (Positive control and Negative control required daily) /Tips required /Any other accessory required for the above parameters according to the mentioned number of tests must be quoted in BOQ / Financial Bid and the rate will be frozen for next 5 years from the rate contract done. L-1 will be decided on considering unit price of the equipment and Cost of 500 (Five hundred) for each test as listed above along with their all consumables to perform those tests.	No Change
22	As a startup kit, 50 no's (fifty) test of Platelet kit & 10 no's (Ten) of Plasma Exchange kit along with all consumables of reagents and other Reagents (Cleaner/Washer/Diluent /Kits/ ACD Solution/Calibrator, QCs (Positive control and Negative control required daily) /Tips required Any other accessory required for the above parameters must be supplied complementary at the time of installation and commissioning.	As a startup kit, 50 no's (fifty) test of Platelet kit & 05 no's (Five) of Plasma Exchange kit along with all consumables of reagents and other Reagents (Cleaner/Washer/Diluent /Kits/ ACD Solution/Calibrator, QCs (Positive control and Negative control required daily) /Tips required Any other accessory required for the above parameters must be supplied complementary at the time of installation and commissioning.