

ANNEXURE-I

Bidder's amendment request against existing specification				
Microwave diathermy and traction with Couch				
S. No.	Point No.	Existing Specification	Amendment Requested	Technical Specification After Amendment
1	Traction- Point No. 2	Unit should provide belts, straps and flexion stool for lumbar and cervical traction, should with 60 programs for extension i.e. 59 fixed and 01 variable program.	1. Unit should provide belts, Straps, flexion stool for lumbar traction and belt and strap for cervical traction. The total cost of the machine should be inclusive of each mentioned respective assessorly and one extra each for each unit.	1. Unit should provide belts, Straps, flexion stool for lumbar traction and belt and strap for cervical traction. The total cost of the machine should be inclusive of each mentioned respective assessorly and one extra each for each unit.
2	Optional	Belts, straps and adaptor for Hip Joint Traction, hand Traction, foot traction, Thoracal traction	Bidders have requested to Delete the 'optional' clause under technical specification as it may increase the overall cost of the equipment. The same request is also made by the participating technical experts from NMCH, PMCH and Bihar physiotherapy college stating the non requirement of these assessorly with the existing equipment.	Amendment Approved. The 'Optional' Clause under the technical specification is Deleted.

Note:- 1. Rest of the points under technical specification of the machine remains unchanged.

Assist Prof (Physio)
BCPO
28/4/16
SR, PMR
DEPT
N.M.C.H.
28/4/16
Physiotherapy
N.M.C.H. Patna
Assist Prof
Physiotherapy
N.M.C.H. Patna
28/4/16
Shruti
28/4/16

28/4/16
M.W. Enterprises
N. Delhi

28/4/16

28/4/16
28/4/16

Shruti
28/4/16